




Dante C
Huntsman/DHUN/CC01/INEE
L/US

05/26/2010 02:05 PM

To "Courtemanche, Steven" <Steven.Courtemanche@nrc.gov>

cc

bcc

Subject Re: Additional information for NMED Record 100082 

Thanks for the information Steven, I will get it into the NMED shortly. Have a great evening,

Dante

"Courtemanche, Steven" <Steven.Courtemanche@nrc.gov>



"Courtemanche, Steven"
<Steven.Courtemanche@nrc.
gov>



05/26/2010 12:40 PM

To Dante C Huntsman <Dante.Huntsman@inl.gov>

cc "White, Duane" <Duane.White@nrc.gov>

Subject Additional information for NMED Record 100082

On May 20, 2010, you requested additional information for NMED Record No. 100082. The information follows:

- (1) Who was the manufacturer of the HDR unit? 
- (2) What are the model and serial numbers of the HDR?  and serial number 31503
- (3) Was the HDR source Ir-192? Yes, indeed it was.
- (4) What was the activity of that source? 6.689 Curies as of the time of the incident.

Changes:

Consultant Hired: Yes

Abnormal Occurrence: No

Investigation: Yes

The NRC RI management is going over its enforcement options and whether escalated enforcement is necessary. The NMED record should stay open until NRC management completes its deliberations.