



**UNITED STATES  
NUCLEAR REGULATORY COMMISSION**

REGION III  
2443 WARRENVILLE ROAD, SUITE 210  
LISLE, ILLINOIS 60532-4352

**APR 25 2011**

Hector J. Marchand, M.D.  
Radiation Safety Officer  
NW Indiana Cardiovascular Physicians, P.C.  
2000 Roosevelt Road  
Valparaiso, IN 46383

Dear Dr. Marchand:


Enclosed is Amendment No. 12 to your NRC Material License No. 13-26691-01 in accordance with your request dated March 14, 2011. Please note that the changes made to your license are printed in **bold** font.

Please review the enclosed document carefully and be sure that you understand all conditions. If there are any errors or questions, please notify the U.S. Nuclear Regulatory Commission, Region III office at (630) 829-9887 so that we can provide appropriate corrections and answers.

Please be advised that prior to performing positron emission tomography (PET) imaging, you will need to amend your NRC license to include shielding information of your facility and provide a copy of your shielding calculations that shows compliance 10 CFR Part 20, Radiation levels in unrestricted areas.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the NRC's Agencywide Documents Access and Management System (ADAMS), accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Sincerely,

  
James R. Mullauer, M.H.S.  
Health Physicist  
Materials Licensing Branch

License No. 13-26691-01  
Docket No. 030-34025

Enclosure: Amendment No. 12