



U.S. NUCLEAR REGULATORY COMMISSION
 REGION III
 2443 WARRENVILLE ROAD
 LISLE, ILLINOIS 60532-4351
 (630) 829-9892 FAX: (630) 515-1078

TO: File

COMPANY: N/A, see below

PAGES: N/A TEL. : N/A

FAX #: N/A

CONVERSATION RECORD

		DATES 04/06/2011
NAME OF PERSON(S) CONTACTED Ray A. Carlson	TELEPHONE NO. (734) 395-7361	ORGANIZATION American Diagnostic Medicine
REPRESENTED PERSON or PERSONS Ray A. Carlson, Radiation Safety Officer		ORGANIZATION American Diagnostic Medicine
SUBJECT		
License No.: 12-32597-01		Control No.: 574780

SUMMARY

We reviewed the requesting amendment request and found we were unable to continue this licensing action until we received additional information concerning the items noted below:

1. In your letter dated 03/27/2011, you indicated that a termination was requested. A completed NRC Form 314 is required to complete such a termination. It was not enclosed with the request. Please submit a completed NRC Form 314.
RESPONSE: The NRC Form 314 was signed, submitted via facsimile on April 6, 2011. No further action is required.

2. The licensee referenced above has indicated that it has transferred its remaining sealed sources to License No. 21-26253-01; Radiological Physics Service, 3839 Napier Road, Plymouth, MI 48170. Please indicate the isotope, quantity, and serial number for each source transferred to from American Diagnostic Medicine to any other licensee.
RESPONSE: Licensee submitted a listing of transferred sealed sources in the facsimile referenced above. The list includes the isotope, radioactive quantity, and serial number for each source. No further action is required.

3. Leak test results for all transferred sealed sources are needed.
RESPONSE: Leak test results were included in the facsimile referenced above. No further action is required.

4. Please indicate the radioactive material use history for this facility.
RESPONSE: Licensee indicated, via this phone conversation, that material use history is limited to Technetium-99m. No further action is required.

5. Please indicate the final date on which radioactive material was used at the LaGrange facility, and whether all waste had decayed to background.
RESPONSE: Licensee indicated, via this phone conversation, that all waste had been surveyed and decayed to background. No further action is required.

American Diagnostic Medicine, License No. 12-32597-01, Control No. 574780

- 6. The wipe test results for the LaGrange location all appear to be exactly 200 cpm. Please explain the survey procedures used.
RESPONSE: In this phone conversation, the licensee indicated that 1 single wipe had been taken. Contamination was not identified. We have also noted that the area is less than 300 square feet. No further action is required.

- 7. Please provide the name of the person performing the close-out surveys.
RESPONSE: In this phone conversation, the licensee Radiation Safety Officer indicated that he, Ray Carlson, had personally performed the close-out surveys. No further action is required.

No additional information is required.

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

Sara A.B. Forster

Sara A.B. Forster

04/26/2011

J. Ammons