

SARA A.B. FORSTER
MATERIALS LICENSING BRANCH



NUCLEAR REGULATORY COMMISSION
REGION III
2443 WARRENVILLE ROAD
LISLE, ILLINOIS 60532-4351
(630) 829-9892 FAX: (630) 515-1078

TELECON & FAX TRANSMITTAL

TO: File

COMPANY: N/A, see below

PAGES: N/A TEL.: N/A

FAX #: N/A

CONVERSATION RECORD

	TIME	DATE
	1:00 pm	04/28/2011
NAME OF PERSON(S) CONTACTED	TELEPHONE NO.	ORGANIZATION
Patrick J. Byrne, DABR, CHP, DABSNM	(877) 321-2207	St. Joseph Regional Medical Center
REPRESENTED PERSON or PERSONS		ORGANIZATION
Patrick J. Byrne, Radiation Safety Officer	x343	St. Joseph Regional Medical Center
SUBJECT		
License No.: 13-02650-02		Control No.: 574393

SUMMARY

We have reviewed your requesting license amendment request and find that we are unable to continue this action until we have received information regarding the following:

1. In your letter dated April 4, 2011, the facility diagram indicated an address of 9365 Counselors Row, Suite 300, Indianapolis, IN 46240. However, the original application requested that a 53940 Carmichael Drive location of use be added to your Radioactive Materials License. Please explain the discrepancy.
RESPONSE: The Counselors Row address refers to the architect, not the licensee. Updated diagrams were submitted via facsimile on April 29, 2011, removing references to the architect's place of business. No further action is required at this time.

2. In your letter dated April 4, 2011, the PET analysis is incomplete in presenting the actual shielding used, and/or radiation area surveys clearly indicating compliance with the radiation limits presented in 10 CFR Part 20. Since this is a facility currently using PET, actual data is preferred in presenting the adequacy of the PET shielding used.
RESPONSE: The updated PET analysis, submitted via facsimile on April 29, 2011, was limited to theoretical calculations and proposed shielding, incompletely addressing the request. Refer to phone conversation record dated 04/29/2011, for subsequent request and follow-up.

We have requested that you submit the referenced items –

A facility diagram reflecting activities at 53940 Carmichael Drive and an updated PET shielding analysis.

– via facsimile, to (630) 515-1078. Please reference the Control No. 574393, as listed at the top of this memo. We expected to hear from you on or before April 29, 2011.

For future reference, please always include the name, phone number and fax number of at least one person whom we may contact for additional information when reviewing your licensing correspondence and requests.

Additional information required; Refer to record for phone conversation dated 04/28/2011.

Please direct any questions you have to me at (630) 829-9892 or sara.forster@nrc.gov.

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

Sara A.B. Forster

Sara A.B. Forster 05/03/2011