

### Radiation Background Levels

Survey Meter	
Manufacturer:	LUDLUM
Model:	2360
Serial No.:	145474

Detector	
Manufacturer:	LUD
Model:	43-37
Serial No.:	066273

Date Data Collected: 11/16/09

By: MB

Surface	Location/Room	Surface Abbrev.	1 minute scaler counts				
			MB 1	MB 2	3	4	5
Ceiling Tile	Ref. Area	CT	<del>5/35</del> 3/204	<del>2/59</del> 6/270	8/278	4/311	4/319
Vinyl Flooring	Ref. Area	VF	8/251	5/238	2/256	11/253	3/239
Glazed tile Block	Ref. Area	GTB	26/767	20/753	11/742	26/723	22/765
Carpet	Ref. Area	Carpet	6/258	6/302	11/285	6/310	7/305
Concrete	Ref. Area	Concrete	6/437	6/515	10/410	8/456	4/481
Glass	Ref. Area	Glass	10/572	7/510	8/566	14/571	6/559
Wood	Ref. Area	Wood	11/655	10/659	20/638	11/630	14/673
Ceramic tile	Ref. Area	Tile	17/1006	15/1041	22/1072	16/1025	13/1098
Epoxy	Ref. Area	Epoxy	11/423	10/448	6/368	3/430	6/521
Cinder Block	Ref. Area	Cinder Block	3/674	5/720	6/693	7/706	4/720
Steel Door	Ref. Area	Steel	7/371	12/405	10/372	10/407	8/422

6 | 7  
3/366 | 10/38

Comments: Reference Area - Bldg 501

---



---



---



---



---

### Radiation Background Levels

Survey Meter	
Manufacturer:	WALUM
Model:	2360
Serial No.:	225241

Detector	
Manufacturer:	WALUM
Model:	43-68
Serial No.:	160790

Date Data Collected: 11/16/09

By: \_\_\_\_\_

Surface	Location/Room	Surface Abbrev.	1 minute scaler counts				
			1	2	3	4	5
Ceiling T	Ref Area	CT	1/312	0/290	1/320	2/327	1/353
VINYL Floor	Ref Area	VF	4/208	1/225	2/226	0/220	0/226
locl-G Glaze TILE	Ref Area	G.T.B	9/499	13/494	1/517	10/495	6/503
Carpet	Ref Area	CarPet	0/214	2/188	1/226	2/198	1/211
Concrete	Ref Area	Concrete	1/278	1/278	2/269	1/263	1/264
Glass	Ref Area	Glass	4/230	3/245	1/208	0/227	3/177
Wood	Ref Area	WOOD	1/178	2/198	2/215	2/216	1/209
ceramic TILE	Ref Area	TILE	7/460	10/432	5/401	3/404	5/444
EPOXY	↓	EPOXY	1/150	1/173	1/150	2/147	1/180
Cinder Block		<del>SD</del>	3/205	3/218	3/209	2/204	2/228
Steel Door		SD	2/129	2/130	3/128	1/136	4/119

Comments: Reference Area

---



---



---



---

### Radiation Background Levels

Survey Meter	
Manufacturer:	LUDLUM
Model:	2360
Serial No.:	184906

Detector	
Manufacturer:	LUDLUM
Model:	43-68
Serial No.:	116721

Date Data Collected: 11/16/09

By: AHMA Williams

Surface	Location/Room	Surface Abbrev.	1 minute scaler counts				
			1	2	3	4	5
Ceiling Tile	Reference	CT	3/296	1/310	1/303	0/300	1/324
Vinyl Flooring	Reference	VF	2/219	1/218	2/214	1/226	0/232
Glazed Tile	Reference	GTB	7/471	3/448	5/478	11/489	6/445
Carpet	↓	Carpet	0/214	1/182	1/189	1/183	1/168
Concrete		Concrete	2/248	1/276	2/251	2/238	0/256
glass		Glass	1/260	2/215	1/247	2/239	1/211
Wood		Wood	2/190	4/194	0/180	1/173	0/219
Ceramic tile		Tile	5/388	6/363	8/403	8/384	4/383
Epoxu		Epoxu	1/124	0/139	1/123	0/112	1/121
Cinder block		Cinder block	1/184	3/174	3/148	4/198	1/178
Steel		Steel	0/110	3/129	1/110	3/111	2/113

Comments: Bldg 501 was used as reference area

---



---



---



---