

ALS Laboratory Group

ANALYTICAL CHEMISTRY & TESTING SERVICES



Environmental Division

December 30, 2009

Mr. Mike Barsa
Cabrera Services, Inc.
103 E Mount Royal Ave. Ste. 2B
Baltimore, MD 21202

Re: ALS Workorder: 09-12-056
Project Name: Forest Glen Rad Scoping Survey
Project Number: 08-3800.04

Dear Mr. Barsa:

Two wipe samples were received from Cabrera Services, Inc. on December 7, 2009. The samples were scheduled for the following analysis:

Tritium pages 1-86

The results for this analysis are contained in the enclosed reports.

Thank you for your confidence in ALS Laboratory Group. Should you have any questions, please call.

Sincerely,

ALS Laboratory Group
Lance Steere
Senior Project Manager

LRS/mh
Enclosure (s): Report

ALS Laboratory Group -- FC

Sample Number(s) Cross-Reference Table

Paragon OrderNum: 0912056
Client Name: Cabrera Services Inc.
Client Project Name: Forest Glen Rad Scoping Survey
Client Project Number: 08-3800.04
Client PO Number:

Client Sample Number	Lab Sample Number	COC Number	Matrix	Date Collected	Time Collected
SU03-BIAS-02	0912056-1		WIPE	18-Nov-09	12:50
SU03-BIAS-03	0912056-2		WIPE	18-Nov-09	12:55

Chain-of-Custody

Standard or Retest (Due)

Turnaround

Dispose or Return to Client

Sample ID	Date	Time *	circle method or specify under comments		No. of Containers	Matrix	Lab ID	H-3 Via Liquid Scintillation E906.0																									
SU03-BIAS-02	11/18/09	12:50			1	SM	①	X																									
SU03-BIAS-03	11/18/09	12:55			1	SM	②	X																									
								X																									
								X																									
								X																									
								X																									
								X																									
								X																									
								X																									

Report To: Michael Barsa
 Phone: 410-332-8177
 Fax: 410-332-8183
 Company: Cabrera Services
 Address: 103 E. Mount Royal Ave, Suite 2B
 Baltimore, MD 21202

Project Name / No.: Forest Glen Rad Scoping Survey
 Sampler(s): KK, AC, MB, AW (circle one) Turnaround

Relinquished By: (1)
 Signature: [Signature]
 Printed Name: Michael Barsa
 Date: 12/4/09 Time: 1400
 Company: LABORSA

Received By: Cheryl Grumbel
 Signature: [Signature]
 Printed Name: Cheryl Grumbel
 Date: 12-7-09 Time: 1140
 Company: ALS

Comments: 08-3800.04, Task 2
 WRAMC Forest Glen Annex Radiological Survey
 SM=Smeared
 Additional bias samples collected 11/18, held until necessity of analysis was determined
 Submitted to lab - 12/4/2009

Total number of containers: 2

Relinquished By: (2)
 Signature: _____
 Printed Name: _____
 Date: _____ Time: _____
 Company: _____

Received By: (2)
 Signature: _____
 Printed Name: _____
 Date: _____ Time: _____
 Company: _____

Form 2024.xls (1/3/01)



CONDITION OF SAMPLE UPON RECEIPT FORM

Client: CABRERA

Workorder No: 0912056

Project Manager: LS

Initials: CDT Date: 12-7-09

1. Does this project require any special handling in addition to standard Paragon procedures?		YES	<input checked="" type="radio"/> NO
2. Are custody seals on shipping containers intact?	NONE	<input checked="" type="radio"/> YES	NO
3. Are Custody seals on sample containers intact?	<input checked="" type="radio"/> NONE	YES	NO
4. Is there a COC (Chain-of-Custody) present or other representative documents?		<input checked="" type="radio"/> YES	NO
5. Are the COC and bottle labels complete and legible?		<input checked="" type="radio"/> YES	NO
6. Is the COC in agreement with samples received? (IDs, dates, times, no. of samples, no. of containers, matrix, requested analyses, etc.)		<input checked="" type="radio"/> YES	NO
7. Were airbills / shipping documents present and/or removable?	DROP OFF	<input checked="" type="radio"/> YES	NO
8. Are all aqueous samples requiring preservation preserved correctly? (excluding volatiles)	<input checked="" type="radio"/> N/A	YES	NO
9. Are all aqueous non-preserved samples pH 4-9?	<input checked="" type="radio"/> N/A	YES	NO
10. Is there sufficient sample for the requested analyses?		<input checked="" type="radio"/> YES	NO
11. Were all samples placed in the proper containers for the requested analyses?		<input checked="" type="radio"/> YES	NO
12. Are all samples within holding times for the requested analyses?		<input checked="" type="radio"/> YES	NO
13. Were all sample containers received intact? (not broken or leaking, etc.)		<input checked="" type="radio"/> YES	NO
14. Are all samples requiring no headspace (VOC, GRO, RSK/MEE, Rx CN/S, radon) headspace free? Size of bubble: _____ < green pea _____ > green pea	<input checked="" type="radio"/> N/A	YES	NO
15. Do perchlorate LCMS-MS samples have headspace? (at least 1/3 of container required)	<input checked="" type="radio"/> N/A	YES	NO
16. Were samples checked for and free from the presence of residual chlorine? (Applicable when PM has indicated samples are from a chlorinated water source; note if field preservation with sodium thiosulfate was not observed.)	<input checked="" type="radio"/> N/A	YES	NO
17. Were the samples shipped on ice?		YES	<input checked="" type="radio"/> NO
18. Were cooler temperatures measured at 0.1-6.0°C? IR gun used*: #2 #4		RAD ONLY	YES <input checked="" type="radio"/> NO
Cooler #: <u>1</u>			
Temperature (°C): <u>18.11.2</u>			
No. of custody seals on cooler: <u>1</u>			
External µR/hr reading: <u>12</u>			
Background µR/hr reading: <u>12</u>			
Were external µR/hr readings ≤ two times background and within DOT acceptance criteria? <input checked="" type="radio"/> YES / NO / NA (If no, see Form 008.)			

Additional Information: PROVIDE DETAILS BELOW FOR A NO RESPONSE TO ANY QUESTION ABOVE, EXCEPT #1 AND #16

If applicable, was the client contacted? YES / NO / NA Contact: _____ Date/Time: _____

Project Manager Signature / Date: [Signature] 12/7/09

*IR Gun #2: Oakton, SN 29922500201-0066

*IR Gun #4: Oakton, SN 2372220101-0002

From: Origin ID: ODMA (410) 332-8177
Mike Barsa
CABRERA SERVICES
103 E. Mount Royal Ave
Ste 2B
Baltimore, MD 21202



Ship Date: 04DEC09
ActWgt: 2.0 LB
CAD: 4239785/NET9090
Account#: S *****

Delivery Address Bar Code



12
11

SHIP TO: (970) 490-1511 BILL SENDER


Lance Steere
ALS-Paragon
225 COMMERCE DR

Ref # 08-3800.04, Task 200
Invoice #
PO #
Dept #

FORT COLLINS, CO 80524


MON - 07DEC A2
STANDARD OVERNIGHT

TRK# 7981 9299 5533
0201



XH FTCA

80524
CO-US
DEN



After printing this label:

1. Use the 'Print' button on this page to print your label to your laser or inkjet printer.
2. Fold the printed page along the horizontal line.
3. Place label in shipping pouch and affix it to your shipment so that the barcode portion of the label can be read and scanned.

Warning: Use only the printed original label for shipping. Using a photocopy of this label for shipping purposes is fraudulent and could result in additional billing charges, along with the cancellation of your FedEx account number.

Use of this system constitutes your agreement to the service conditions in the current FedEx Service Guide, available on fedex.com. FedEx will not be responsible for any claim in excess of \$100 per package, whether the result of loss, damage, delay, non-delivery, misdelivery, or misinformation, unless you declare a higher value, pay an additional charge, document your actual loss and file a timely claim. Limitations found in the current FedEx Service Guide apply. Your right to recover from FedEx for any loss, including intrinsic value of the package, loss of sales, income interest, profit, attorney's fees, costs, and other forms of damage whether direct, incidental, consequential, or special is limited to the greater of \$100 or the authorized declared value. Recovery cannot exceed actual documented loss. Maximum for items of extraordinary value is \$500, e.g. jewelry, precious metals, negotiable instruments and other items listed in our Service Guide. Written claims must be filed within strict time limits, see current FedEx Service Guide.



Tritium Case Narrative

Cabrera Services, Inc.

Forest Glen Rad Scoping Survey -- 08-3800.04

Work Order Number: 0912056

1. This report consists of the analytical results and supporting documentation for two wipe samples received by ALS on 12/07/2009.
2. These samples were prepared according to procedure SOP700R10, with procedure modifications outlined in QASS 379412.
3. The samples were analyzed for the presence of tritium according to procedure SOP704R9. The analyses were completed on 12/25/2009.
4. Upon analysis of these samples, it was noted that the quench factor (H#) was outside of the current usable calibration range for the associated QC. Thus, an approximate volume of 10 μ L of nitro-methane was added to the associated QC to bring them into the usable calibration range.
5. The analysis results for these samples are reported on an 'As Received' basis in units of pCi/sample.
6. Sample volume was insufficient to allow preparation of a duplicate and matrix spike. A laboratory control sample duplicate (LCSD) was prepared in lieu of a client sample duplicate.
7. Sample 0912056-1 had a "Window 2" count rate of 27.17 cpm, below the lower control limit of 27.30 cpm, determined from calibration on 12/24/2009 through 12/25/2009. For this analysis, "Window 2" is monitored for high-energy beta contamination, therefore no contamination is observed and the data quality is not believed to be affected.
8. No further anomalous situations were encountered during the preparation or analysis of these samples. All remaining quality control criteria were met.



The data contained in the following report have been reviewed and approved by the personnel listed below. In addition, ALS certifies that the analyses reported herein are true, complete and correct within the limits of the methods employed.

Jean Anderson
Jean Anderson
Radiochemistry Primary Data Reviewer

12/30/09
Date

Michelle Lynn
Radiochemistry Final Data Reviewer

12-30-09
Date



Section 1

CHAIN OF CUSTODY

ALS Laboratory Group -- FC

Sample Number(s) Cross-Reference Table

Paragon OrderNum: 0912056

Client Name: Cabrera Services Inc.

Client Project Name: Forest Glen Rad Scoping Survey

Client Project Number: 08-3800.04

Client PO Number:

Client Sample Number	Lab Sample Number	COC Number	Matrix	Date Collected	Time Collected
SU03-BIAS-02	0912056-1		WIPE	18-Nov-09	12:50
SU03-BIAS-03	0912056-2		WIPE	18-Nov-09	12:55

ALS Laboratory Group

Project Name / No.: Forest Glen Rad Scoping Survey KK, AC, MB, AW (circle one) Turnaround Standard or Rush None Dispose or Return to Client

Report To: Michael Barsa
Phone: 410-332-8177
Fax: 410-332-8183
Company: Cabrera Services
Address: 103 E. Mount Royal Ave, Suite 2B
 Baltimore, MD 21202

Sample ID	Date	Time *	Lab ID	Matrix	No. of Containers	H-3 via Liquid Scintillation E906.0																	
SU03-BIAS-02	11/18/09	12:50	①	SM	1	X																	
SU03-BIAS-03	11/18/09	12:55	②	SM	1	X																	
						X																	
						X																	
						X																	
						X																	
						X																	
						X																	
						X																	
						X																	
						X																	
						X																	
						X																	
						X																	
						X																	
						X																	
						X																	
						X																	
						X																	
						X																	
						X																	
						X																	
						X																	
						X																	

Comments: 08-3800.04, Task 2
 WRAMC Forest Glen Annex Radiological Survey
 SM=Smear
 Additional bias samples collected 11/18, held until necessity of analysis was determined
 Submitted to lab - 12/4/2009

Total number of containers: 2

Relinquished By: <u>[Signature]</u> Signature <u>Michael Barsa</u> Printed Name <u>Michael Barsa</u> Date <u>12/4/09</u> Time <u>1400</u> Company <u>Cabrera</u>	Relinquished By: <u>[Signature]</u> Signature _____ Printed Name _____ Date _____ Time _____ Company _____
Received By: <u>[Signature]</u> Signature <u>Cheryl Trumble</u> Printed Name <u>Cheryl Trumble</u> Date <u>12-7-09</u> Time <u>1140</u> Company <u>ALS</u>	Received By: <u>[Signature]</u> Signature _____ Printed Name _____ Date _____ Time _____ Company _____



CONDITION OF SAMPLE UPON RECEIPT FORM

Client: Cabrera

Workorder No: 0912056

Project Manager: LS

Initials: CDT Date: 12-7-09

1. Does this project require any special handling in addition to standard Paragon procedures?		YES	<input checked="" type="radio"/> NO
2. Are custody seals on shipping containers intact?	NONE	<input checked="" type="radio"/> YES	NO
3. Are Custody seals on sample containers intact?	<input checked="" type="radio"/> NONE	YES	NO
4. Is there a COC (Chain-of-Custody) present or other representative documents?		<input checked="" type="radio"/> YES	NO
5. Are the COC and bottle labels complete and legible?		<input checked="" type="radio"/> YES	NO
6. Is the COC in agreement with samples received? (IDs, dates, times, no. of samples, no. of containers, matrix, requested analyses, etc.)		<input checked="" type="radio"/> YES	NO
7. Were airbills / shipping documents present and/or removable?	DROP OFF	<input checked="" type="radio"/> YES	NO
8. Are all aqueous samples requiring preservation preserved correctly? (excluding volatiles)	<input checked="" type="radio"/> N/A	YES	NO
9. Are all aqueous non-preserved samples pH 4-9?	<input checked="" type="radio"/> N/A	YES	NO
10. Is there sufficient sample for the requested analyses?		<input checked="" type="radio"/> YES	NO
11. Were all samples placed in the proper containers for the requested analyses?		<input checked="" type="radio"/> YES	NO
12. Are all samples within holding times for the requested analyses?		<input checked="" type="radio"/> YES	NO
13. Were all sample containers received intact? (not broken or leaking, etc.)		<input checked="" type="radio"/> YES	NO
14. Are all samples requiring no headspace (VOC, GRO, RSK/MEE, Rx CN/S, radon) headspace free? Size of bubble: _____ < green pea _____ > green pea	<input checked="" type="radio"/> N/A	YES	NO
15. Do perchlorate LCMS-MS samples have headspace? (at least 1/3 of container required)	<input checked="" type="radio"/> N/A	YES	NO
16. Were samples checked for and free from the presence of residual chlorine? (Applicable when PM has indicated samples are from a chlorinated water source; note if field preservation with sodium thiosulfate was not observed.)	<input checked="" type="radio"/> N/A	YES	NO
17. Were the samples shipped on ice?		YES	<input checked="" type="radio"/> NO
18. Were cooler temperatures measured at 0.1-6.0°C? IR gun used*: #2 #4		RAD ONLY	YES <input checked="" type="radio"/> NO
Cooler #: <u>1</u>			
Temperature (°C): <u>DOT 12-7-09</u> <u>18 11.2</u>			
No. of custody seals on cooler: <u>1</u>			
External µR/hr reading: <u>12</u>			
Background µR/hr reading: <u>12</u>			
Were external µR/hr readings ≤ two times background and within DOT acceptance criteria? <input checked="" type="radio"/> YES / NO / NA (If no, see Form 008.)			

Additional Information: PROVIDE DETAILS BELOW FOR A NO RESPONSE TO ANY QUESTION ABOVE. EXCEPT #1 AND #16

If applicable, was the client contacted? YES / NO / NA Contact: _____ Date/Time: _____

Project Manager Signature / Date: [Signature] 12/7/09

*IR Gun #2: Oakton, SN 29922500201-0066

*IR Gun #4: Oakton, SN 2372220101-0002

0912056

From: Origin ID: ODMA (410) 332-8177
Mike Barsa
CABRERA SERVICES
103 E. Mount Royal Ave
Ste 2B
Baltimore, MD 21202



Ship Date: 04DEC09
ActWgt: 2.0 LB
CAD: 4239785/INET9090
Account#: S *****

Delivery Address Bar Code



Ref # 08-3800.04, Task 200
Invoice #
PO #
Dept #

12
11


SHIP TO: (970) 490-1511 BILL SENDER

Lance Steere
ALS-Paragon
225 COMMERCE DR

FORT COLLINS, CO 80524


TRK# 7981 9299 5533
0201

MON - 07DEC A2
STANDARD OVERNIGHT



XH FTCA

80524
CO-US
DEN



After printing this label:

1. Use the 'Print' button on this page to print your label to your laser or inkjet printer.
2. Fold the printed page along the horizontal line.
3. Place label in shipping pouch and affix it to your shipment so that the barcode portion of the label can be read and scanned.

Warning: Use only the printed original label for shipping. Using a photocopy of this label for shipping purposes is fraudulent and could result in additional billing charges, along with the cancellation of your FedEx account number.

Use of this system constitutes your agreement to the service conditions in the current FedEx Service Guide, available on fedex.com. FedEx will not be responsible for any claim in excess of \$100 per package, whether the result of loss, damage, delay, non-delivery, misdelivery, or misinformation, unless you declare a higher value, pay an additional charge, document your actual loss and file a timely claim. Limitations found in the current FedEx Service Guide apply. Your right to recover from FedEx for any loss, including intrinsic value of the package, loss of sales, income interest, profit, attorney's fees, costs, and other forms of damage whether direct, incidental, consequential, or special is limited to the greater of \$100 or the authorized declared value. Recovery cannot exceed actual documented loss. Maximum for items of extraordinary value is \$500, e.g. jewelry, precious metals, negotiable instruments and other items listed in our Service Guide. Written claims must be filed within strict time limits, see current FedEx Service Guide.



Section 2



SAMPLE RESULTS SUMMARY

Tritium Analysis By Liquid Scintillation Sample Results Summary

Client Name: Cabrera Services Inc.

Laboratory Name: ALS Laboratory Group -- FC

Page: 1 of 1

Client Project Name: Forest Glen Rad Scoping Survey

PAI Work Order: 0912056

Reported on: Wednesday, December 30, 2009

Client Project Number: 08-3800.04

7:43:14 AM

Lab Sample ID	Client Sample ID	Sample Type	Nuclide	Result +/- 2 s TPU	MDC	Units	Matrix	Prep Batch	Date Analyzed	Flags
0912056-1	SU03-BIAS-02	Sample	H-3	0.5 +/- 1.8	3.0	pCi/sample	WIPE	3H091209-2	12/24/2009	U
0912056-2	SU03-BIAS-03	Sample	H-3	0.7 +/- 1.8	3.0	pCi/sample	WIPE	3H091209-2	12/24/2009	U

Comments:

Data Package ID: H30912056-1

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- LT - Result is less than Requested MDC, greater than sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
- Y2 - Chemical Yield outside default limits.
- M - The requested MDC was not met.
- M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.

Abbreviations:

- TPU - Total Propagated Uncertainty
- MDC - Minimum Detectable Concentration
- BDL - Below Detection Limit

Date Printed: Wednesday, December 30, 2009

ALS Laboratory Group -- FC

LIMS Version: 6.321A

Page 1 of 1



Section 3

QC RESULTS SUMMARY



Tritium Analysis By Liquid Scintillation

PAI 704 Rev 9

Method Blank Results

Lab Name: ALS Laboratory Group -- FC

Work Order Number: 0912056

Client Name: Cabrera Services Inc.

ClientProject ID: Forest Glen Rad Scoping Survey 08-3800.04

Lab ID: 3H091209-2MB

Sample Matrix: WIPE
Prep SOP: PAI 700 Rev 10
Date Collected: 09-Dec-09
Date Prepared: 09-Dec-09
Date Analyzed: 24-Dec-09

Prep Batch: 3H091209-2
QCBatchID: 3H091209-2-1
Run ID: 3H091209-2A
Count Time: 60 minutes

Final Aliquot: 1.00 sample
Result Units: pCi/sample
File Name: B60_04_122401

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Requested MDC	Lab Qualifier
10028-17-8	H-3	1.3 +/- 2.0	3.2	20	U

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC.
Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
Y2 - Chemical Yield outside default limits.
LT - Result is less than Requested MDC, greater than sample specific MDC.

Abbreviations:

TPU - Total Propagated Uncertainty
MDC - Minimum Detectable Concentration
BDL - Below Detection Limit

M - Requested MDC not met.
B - Analyte concentration greater than MDC.
B3 - Analyte concentration greater than MDC but less than Requested MDC.

Data Package ID: *H30912056-1*

Tritium Analysis By Liquid Scintillation

PAI 704 Rev 9

Laboratory Control Sample(s)

Lab Name: ALS Laboratory Group -- FC

Work Order Number: 0912056

Client Name: Cabrera Services Inc.

ClientProject ID: Forest Glen Rad Scoping Survey 08-3800.04

Lab ID: 3H091209-2LCS	Sample Matrix: WIPE Prep SOP: PAI 700 Rev 10 Date Collected: 09-Dec-09 Date Prepared: 09-Dec-09 Date Analyzed: 25-Dec-09	Prep Batch: 3H091209-2 QCBatchID: 3H091209-2-1 Run ID: 3H091209-2A Count Time: 60 minutes	Final Aliquot: 1.00 sample Result Units: pCi/sample File Name: B60_04_122401
-----------------------	--	--	--

CASNO	Target Nuclide	Results +/- 2s TPU	MDC	Spike Added	% Rec	Control Limits	Lab Qualifier
10028-17-8	H-3	114 +/- 18	3	113	101	85 - 115	P

Comments:

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- LT - Result is less than Requested MDC, greater than sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
- Y2 - Chemical Yield outside default limits.
- L - LCS Recovery below lower control limit.
- H - LCS Recovery above upper control limit.
- P - LCS Recovery within control limits.
- M - The requested MDC was not met.
- M3 - The requested MDC was not met, but thereported activity is greater than the reported MDC.

Abbreviations:

- TPU - Total Propagated Uncertainty
- MDC - Minimum Detectable Concentration

Data Package ID: H30912056-1

Tritium Analysis By Liquid Scintillation

PAI 704 Rev 9

Laboratory Control Sample(s)

Lab Name: ALS Laboratory Group -- FC

Work Order Number: 0912056

Client Name: Cabrera Services Inc.

ClientProject ID: Forest Glen Rad Scoping Survey 08-3800.04

Lab ID: 3H091209-2LCSD

Sample Matrix: WIPE
Prep SOP: PAI 700 Rev 10
Date Collected: 09-Dec-09
Date Prepared: 09-Dec-09
Date Analyzed: 25-Dec-09

Prep Batch: 3H091209-2
QC Batch ID: 3H091209-2-1
Run ID: 3H091209-2A
Count Time: 60 minutes

Final Aliquot: 1.00 sample
Result Units: pCi/sample
File Name: B60_04_122401

CASNO	Target Nuclide	Results +/- 2s TPU	MDC	Spike Added	% Rec	Control Limits	Lab Qualifier
10028-17-8	H-3	114 +/- 18	3	113	101	85 - 115	P

Comments:

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- LT - Result is less than Requested MDC, greater than sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
- Y2 - Chemical Yield outside default limits.
- L - LCS Recovery below lower control limit.
- H - LCS Recovery above upper control limit.
- P - LCS Recovery within control limits.
- M - The requested MDC was not met.
- M3 - The requested MDC was not met, but thereported activity is greater than the reported MDC.

Abbreviations:

- TPU - Total Propagated Uncertainty
- MDC - Minimum Detectable Concentration

Data Package ID: H30912056-1

Tritium Analysis By Liquid Scintillation

PAI 704 Rev 9

Duplicate Sample Results (DER)

Lab Name: ALS Laboratory Group -- FC

Work Order Number: 0912056

Client Name: Cabrera Services Inc.

ClientProject ID: Forest Glen Rad Scoping Survey 08-3800.04

Field ID:
Lab ID: 3H091209-2LCSD

Sample Matrix: WIPE
 Prep SOP: PAI 700 Rev 10
 Date Collected: 09-Dec-09
 Date Prepared: 09-Dec-09
 Date Analyzed: 25-Dec-09

Prep Batch: 3H091209-2
 QCBatchID: 3H091209-2-1
 Run ID: 3H091209-2A
 Count Time: 60 minutes

Final Aliquot: 1.00 sample
 Prep Basis: As Received
 Moisture(%): NA
 Result Units: pCi/sample
 File Name: B60_04_122401

CASNO	Analyte	Sample			Duplicate			DER	DER Lim
		Result +/- 2 s TPU	MDC	Flags	Result +/- 2 s TPU	MDC	Flags		
10028-17-8	H-3	114 +/- 18	3	P	114 +/- 18	3	P	0.00078	2.13

Comments:

Duplicate Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative yield is assumed.
- Y2 - Chemical Yield outside default limits.
- W - DER is greater than Warning Limit of 1.42
- D - DER is greater than Control Limit of 2.13
- LT - Result is less than Request MDC, greater than sample specific MDC
- M - Requested MDC not met.
- M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.
- L - LCS Recovery below lower control limit.
- H - LCS Recovery above upper control limit.
- P - LCS, Matrix Spike Recovery within control limits.
- N - Matrix Spike Recovery outside control limits

Abbreviations:

- TPU - Total Propagated Uncertainty
- DER - Duplicate Error Ratio
- BDL - Below Detection Limit
- NR - Not Reported

Data Package ID: H30912056-1



Section 4

INDIVIDUAL SAMPLE RESULTS

4

Tritium Analysis By Liquid Scintillation

PAI 704 Rev 9
Sample Results

Lab Name: ALS Laboratory Group -- FC
Work Order Number: 0912056
Client Name: Cabrera Services Inc.
ClientProject ID: Forest Glen Rad Scoping Survey 08-3800.04

Field ID: SU03-BIAS-02	Sample Matrix: WIPE	Prep Batch: 3H091209-2	Final Aliquot: 1.00 sample
Lab ID: 0912056-1	Prep SOP: PAI 700 Rev 10	QCBatchID: 3H091209-2-1	Prep Basis: As Received
	Date Collected: 18-Nov-09	Run ID: 3H091209-2A	Moisture(%): NA
	Date Prepared: 09-Dec-09	Count Time: 60 minutes	Result Units: pCi/sample
	Date Analyzed: 24-Dec-09	Report Basis: As Received	File Name: B60_04_122401

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Requested MDC	Lab Qualifier
10028-17-8	H-3	0.5 +/- 1.8	3.0	20	U

Comments:

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
- Y2 - Chemical Yield outside default limits.
- LT - Result is less than Requested MDC, greater than sample specific MDC.
- M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.
- M - The requested MDC was not met.

Abbreviations:

- TPU - Total Propagated Uncertainty
- MDC - Minimum Detectable Concentration
- BDL - Below Detection Limit

Data Package ID: H30912056-1

Tritium Analysis By Liquid Scintillation

PAI 704 Rev 9
Sample Results

Lab Name: ALS Laboratory Group -- FC
Work Order Number: 0912056
Client Name: Cabrera Services Inc.
ClientProject ID: Forest Glen Rad Scoping Survey 08-3800.04

Field ID: SU03-BIAS-03	Sample Matrix: WIPE	Prep Batch: 3H091209-2	Final Aliquot: 1.00 sample
Lab ID: 0912056-2	Prep SOP: PAI 700 Rev 10	QCBatchID: 3H091209-2-1	Prep Basis: As Received
	Date Collected: 18-Nov-09	Run ID: 3H091209-2A	Moisture(%): NA
	Date Prepared: 09-Dec-09	Count Time: 60 minutes	Result Units: pCi/sample
	Date Analyzed: 24-Dec-09	Report Basis: As Received	File Name: B60_04_122401

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Requested MDC	Lab Qualifier
10028-17-8	H-3	0.7 +/- 1.8	3.0	20	U

Comments:

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
- Y2 - Chemical Yield outside default limits.
- LT - Result is less than Requested MDC, greater than sample specific MDC.
- M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.
- M - The requested MDC was not met.

Abbreviations:

- TPU - Total Propagated Uncertainty
- MDC - Minimum Detectable Concentration
- BDL - Below Detection Limit

Data Package ID: H30912056-1



Section 5

RAW DATA



Tritium Analysis By Liquid Scintillation Raw Data Report

Laboratory Name: ALS Laboratory Group -- FC
 PAI Work Order: 0912056

Prep SOP: PAI 700
 Analytical SOP: PAI 704

Reported on: Monday, December 28, 2009
 1:36:14 PM

Sample ID QC Type	Nuclide Type	Sample Date/Time	Prep Batch QC Batch ID	Ingrowth Date/Time	Quench Factor % Lum	Matrix % Moist	Samp Aliq Analy Aliq	Inst ID Det ID	AnRunID File Name	Count Date/Time	GrossCPM BkgCPM	BaseEff ProgEff	CntDur(min) Yield	Activity +/- 2 s TPU	MDC DeclEv	Report/Units ReportBasis	DER RPD	%Spk. Recov Flags
0912056-1	H-3	11/18/2009	3H091209-2	NA	115.8	WIPE	1 S	LS6000	3H091209-2A	12/24/2009	7,070	24.14%	60	0.5	3.0	pCi/sample	NA	NA
SMP	Tgt. Analyte	12:50:00 PM	3H091209-2-1	NA	0.37	NA	1 S	36-2	860_04_122401	8:01 PM	6,829	NA	NA	1.8	NA	As Received	NA	U
0912056-2	H-3	11/18/2009	3H091209-2	NA	117.1	WIPE	1 S	LS6000	3H091209-2A	12/24/2009	7,230	23.95%	60	0.7	3.0	pCi/sample	NA	NA
SMP	Tgt. Analyte	12:55:00 PM	3H091209-2-1	NA	0.32	NA	1 S	36-3	860_04_122401	9:02 PM	6,838	NA	NA	1.8	NA	As Received	NA	U
3H091209-2	H-3	12/9/2009	3H091209-2	NA	127.8	WIPE	1 S	LS6000	3H091209-2A	12/24/2009	7,580	22.75%	60	1.3	3.2	pCi/sample	NA	NA
MB	Tgt. Analyte	10:51:38 AM	3H091209-2-1	NA	0.43	NA	1 S	36-4	860_04_122401	10:03 PM	6,912	NA	NA	2.0	NA	As Received	NA	U
3H091209-2	H-3	12/9/2009	3H091209-2	NA	126.3	WIPE	1 S	LS6000	3H091209-2A	12/25/2009	64,930	22.96%	60	114	3	pCi/sample	NA	101
LCS	Tgt. Analyte	10:51:38 AM	3H091209-2-1	NA	0.21	NA	1 S	36-6	860_04_122401	12:05 AM	6,902	NA	NA	18	NA	As Received	NA	P
3H091209-2	H-3	12/9/2009	3H091209-2	NA	126.3	WIPE	1 S	LS6000	3H091209-2A	12/25/2009	64,920	22.96%	60	114	3	pCi/sample	0.00	101
LCS/D	Tgt. Analyte	10:51:38 AM	3H091209-2-1	NA	0.22	NA	1 S	36-7	860_04_122401	1:07 AM	6,902	NA	NA	18	NA	As Received	NA	P

Comments:

Data Package ID: H30912056-1

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative yield is assumed.
- Y2 - Chemical Yield outside default limits.
- W - DER is greater than Warning Limit of 1.42
- D - DER is greater than Control Limit of 2.13
- + - Duplicate RPD not within limits.
- LT - Result is less than Request MDC, greater than sample specific MDC
- * - Aliquot Basis is 'As Received' while the Report Basis is 'Dry Weight'.
- # - Aliquot Basis is 'Dry Weight' while the Report Basis is 'As Received'

Notes:

- 1) The Tracer results are not yield corrected (i.e. activity measured not activity added).
- 2) Where sample time is not available, 12:00 PM (Mountain) is used for decay correction.

Abbreviations:

- TR - Tracer
- TA - Target Analyte
- TPU - Total Propagated Uncertainty
- MDC - Minimum Detectable Concentration
- DER - Duplicate Error Ratio
- BDL - Below Detection Limit

M - Requested MDC not met

M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.

L - LCS Recovery below lower control limit.

H - LCS Recovery above upper control limit.

P - LCS, Matrix Spike Recovery within control limits.

N - Matrix Spike Recovery outside control limits

NC - Not Calculated for duplicate results less than 5 times MDC

B - Analyte concentration greater than MDC.

63 - Analyte concentration greater than MDC but less than Requested MDC.

ID: 3445 ML, 10 ML 24 DEC 2009 18:57

USER: 4 COMMENT: 456000

PRESET TIME : 60.00
 DATA GATE : CPM H# : YES SAMPLE REPEATS: 1 PRINTER : STD
 COUNT BLANK : NO IC# : NO REPLICATES : 1 RS232 : EDIT
 TWO PHASE : NO AGC : NO CYCLE REPEATS : 1
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0
 LOW LEVEL : YES HALF LIFE CORRECTION DATE: none

CHAN: 50.0 - 250.0 %ERROR: 1.75 FACTOR: 1.000000 BKG. SUB: 0
 CHAN: 450.0 - 900.0 %ERROR: 20.00 FACTOR: 1.000000 BKG. SUB: 0

ALPHA-BETA DISCRIMINATION: NO

SAM NO	POS	TIME MIN	H#	WIND 1		WIND 2		LUMEX %	ELAPSED TIME
				CPM	%ERROR	CPM	%ERROR		
1	36-1	60.00	129.8	6.58	10.06	30.27	4.69	0.48	61.11
2	36-2	60.00	115.8	7.07	9.71	27.17*	4.95	0.37	122.30
3	36-3	60.00	117.1	7.23	9.60	27.90	4.89	0.32	183.45
4	36-4	60.00	127.8	7.58	9.38	28.80	4.81	0.43	244.65
5	36-5	60.00	129.7	6.75	9.94	29.78	4.73	0.44	305.84
6	36-6	60.00	126.3	64.93	3.20	28.10	4.87	0.21	367.02
7	36-7	60.00	126.3	64.92	3.20	27.52	4.92	0.22	428.20
8	36-8	60.00	130.0	7.45	9.46	28.15	4.87	0.48	489.41

860-04-122401
 note 12-28-09

* < LCL of 27.30

B60_04_122401

BSF Version : 3
Instrument Type : LS 6000
Data Capture Date : 24 Dec 2009 18:58:48
User Filename : C:\...\LS WINCONNECTION\DATA\USER04\UNI22401.BSF
User Number : 4
User Id : 3H:5-ML,10-ML
User Comments : LS6000
Preset Count Time : 60.00
Calculation Mode : CPM
H# Selected : YES
Sample Repeats : 1
Printer Output Mode : STD
Blank Count : NO
IC# or SCR selected : NO
Replicates : 1
RS232 Output Mode : EDIT
Two-Phase Selected : NO
AQC Choice : NO
Cycle Repeats : 1
Scintillator Choice : LIQUID
Lumex Selected : NO
Low Sample Reject Count : 0
Low Level Selection : YES
Half Life Correction Date : none
Window Limits Window 1 : 50.00
Preset %Error Iso1 : 1.75
Norm Multiplier Iso1 : 1.00000
Background CPM 1 : 0.00
Window Limits Window 2 : 450.00
Preset %Error Iso2 : 20.00
Norm Multiplier Iso2 : 1.00000
Background CPM 2 : 0.00
Alpha/Beta Discrimination : NO

Sam	Rack	Time	H#	CPM Iso1	%Err1	CPM Iso2	%Err2	LumEX	ElTime
1	36-1	60.00	129.8	6.58	10.06	30.27	4.69	0.48	61.11
2	36-2	60.00	115.8	7.07	9.71	27.17	4.95	0.37	122.30
3	36-3	60.00	117.1	7.23	9.60	27.90	4.89	0.32	183.45
4	36-4	60.00	127.8	7.58	9.38	28.80	4.81	0.43	244.65
5	36-5	60.00	129.7	6.75	9.94	29.78	4.73	0.44	305.84
6	36-6	60.00	126.3	64.93	3.20	28.10	4.87	0.21	367.02
7	36-7	60.00	126.3	64.92	3.20	27.52	4.92	0.22	428.20
8	36-8	60.00	130.0	7.45	9.46	28.15	4.87	0.48	489.41

WR
12-28-09

379857

Instrument ID: LS6000

LSC Run Log

ALS Laboratory Group - Fort Collins

Date	Sample ID	CountTime (min.)	Rack & Position	Test	User #	Batch ID	Position Check	Initials	Comments
12-21-09	091170-63	70	8 - 2	Hb Sm1	8	3H091223-4	OK	ML	N/A
	-63A		- 3						
	-64		- 4						
	-65		- 5						
	3H091223-4C02		- 6						
12-22-09	Daily QC	10	113 -13, 24		113		OK	ML	N/A
12-22-09	091408-1	46.05	47 - 1	Hb 10ml	4	3H091223-1	OK	ML	Cocktail Verification
	-2		- 2						
	-3	45.75	- 3						
	-4	47.80	- 4						
	-5	47.75	- 5						
	-6	48	- 6						
	3H091223-1B1MB	180	- 7						
	-10A MB		- 8						
	-10B MB		- 9						
	-104 MB		- 10						
	-105 MB		- 11						
	-106 MB		- 12						
12-23-09	091170-66	70	8 - 7	Hb Sm1	8	3H091223-4	OK	ML	N/A
	-67		- 8						
	-68		- 9						
	-69		- 10						
	-70		- 11						
	3H091223-4 MB		- 12						
	4LCS	38.50	40 - 1						
	-4C03	70	40 - 2						
12-23-09	Daily QC	10	113 -13, 24		113		OK	ML	N/A
12-23-09	3H091218-5C01	120	32 - 1	Hb 10ml	2	3H091217-5	OK	ML	
	0912143-1		- 2						
	-2		- 3						

MLC 12-24-09

Analyst / Date MLC 12-28-09

FORM 76216.xls (3/7/09)

Note: Each page is copied as completed and included with the workorder/run documentation, reviewed subsequently.

LSC Run Log

Instrument ID: LS6000

379858

ALS Laboratory Group - Fort Collins

Date	Sample ID	Count Time (min.)	Rack & Position	Test	User #	Batch ID	Position Check	Initials	Comments
12-23-09	0912143-3	120	32-4	Hb 10ml	2	3H091218-5	NA	NA	
	↓ -S								
	3H091218-SCB2								
	-SMB								
	-SICS								
	-SICSD								
	-SCB3								
12-24-09	3H091209-2CB1	60	36-1	Hb 5ml	4	3H091209-2	NA	NA	
	0912056-1								
	↓ -2								
	3H091209-2MB								
	-2CB2								
	-2LCS								
	-2LCS.D								
	↓ -2CB3								
12-25-09	3H091117-2CB1	60	42	Hb 5ml	4	3H091117-2	NA	NA	
	0923008-1								
	↓ -2								
	↓ -3								
	↓ -4								
	↓ -5								
	3H091117-2CB2								
	0923008-6								
	3H091117-2MB								
	-2LCS								
	↓ -2LCSA								
	↓ -2CB3								
12-25-09	3H091216-1CB1	120	56-1	Hb 10ml	2	3H091216-1	NA	NA	
	0912092-1								
	↓ -10								

Analyst / Date SKK 12-28-09

Note: Each page is copied as completed and included with the workorder/run documentation; reviewed subsequently.

LSC Run Log

Instrument ID: LS6000

379859

ALS Laboratory Group - Fort Collins

Date	Sample ID	CountTime (min.)	Rack & Position	Test	User #	Batch ID	Position Check	Initials	Comments
11-25-09	0912092-3	120	56 - 4	H3 10m1	2	31091216-1	✓	MA	NA
	↓ 3ms		- 5						
	0912106-9		- 6						
	31091216-1C42		- 7						
	091206-13		- 8						
	↓ 17		- 9						
	↓ 21	27.00	- 10						
	31091216-1M0	120	- 11						
	↓ 1CS		- 12						
	↓ 1C03		4 - 1						
11-26-09	Daily QC	10	1,3 - 1,3,14		113		MA	MA	NA
12									
13									
14									
15									
16									
17									
18									
19									
20									
21									
22									
23									
24									
25									
26									
27									
28									
29									
30									MA 12-28-09

Analyst / Date MA 12-28-09

Note: Each page is copied as completed and included with the workorder/run documentation, reviewed subsequently.



Section 6

QUALITY ASSURANCE SUMMARY REPORTS

6



ALS Laboratory Group - Fort Collins

QUALITY ASSURANCE SUMMARY SHEET

PAR W.O. #/ BATCH 34091209-2
 TEST 3H
 METHOD Prep
 SOP/REV (PREP) 700.10
 SOP/REV (ANAL) _____

Briefly document any QA or other problems or deviations associated with the analysis of samples. Problems could result from: log-in, color, odor, dilution, consistency, scheduling, equipment, or instrumentation, or may include documentation of minor deviations necessary due to unique DQO's or sample characteristics.

GDW 12/09/09
 Samples in work order 0912056 consisted of swipes in a liquid scintillation vial with 5 ml of water. 15ml of UG LLT were added directly to the original vials, all stickers were removed and vials cleaned with methanol before submitting.
 Due to limited sample volume, no sample duplicate or MS were prepared; for additional quality control a LCS duplicate was prepared.
 The CB, MB, LCS and LCS duplicate were made using swipes folded in half and placed into a liquid scintillation vial. 5 ml of DI water were pipetted into the vial using a calibrated pipette and 15 ml of UG LLT added. For the Tritium LCS and LCSD, 0.1 ml of water was removed with a pipette and 0.1ml of Tritium spike was added.

GDW 12/09/09

GDW 12/09/09

GDW 12/09/09

GDW 12-09-09

TECHNICIAN/ANALYST *Carrie Way* DATE 12-09-09

DEPARTMENT MANAGER *Crystal Shaeffer* DATE 12/9/09

379412

FORM 302r6b.doc (4/22/04)



Section 7

LABORATORY BENCH SHEETS



Radiochemistry Instrument Worksheet

ALS Laboratory Group -- FC

Prep Batch: 3H091209-2

Prep Procedure: H3 swipes 30 min Analytical QASS / NCR? Y 10 NT

Prep Num	QC Type	Init Alq	Fin Alq	Units	Report Units	Cnt 1 File/Inst	Cnt 1 Rack-Pos	Cnt 1 Pos Chk By	Cnt 2 File/Inst	Cnt 2 Rack-Pos	Cnt 2 Pos Chk By	Cnt 3 File/Inst	Cnt 3 Rack-Pos	Cnt 3 Pos Chk By	Notes
1	0912056-1	SMP	1	1	sample pCi/samp	15600	36-2	Wa							
1	0912056-2	SMP	1	1	sample pCi/samp		-3								
1	3H091209-2CB1	MB	1	1	sample pCi/samp		-1								
1	3H091209-2CB1	MB	1	1	sample pCi/samp		-5								10 µL nitro added
1	3H091209-2CB1	MB	1	1	sample pCi/samp		-8								
1	3H091209-2	MB	1	1	sample pCi/samp		-4								
1	3H091209-2	LCS	1	1	sample pCi/samp		-6								
1	3H091209-2	LCSD	1	1	sample pCi/samp		-7								10 µL nitro added

Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date	Aliquot Units	Pipet ID
S1	H-3	648.3610.44	2,517.810	DPM/ml	12/09/09	0.1	ml RS-013

Reporting Units

LabID	Isotope/Nuclide	Rpt Units
0912056-1	H3 Edgewood_smears_20pCi	pCi/samp
0912056-2	H3 Edgewood_smears_20pCi	pCi/samp

Sample Barcodes

0912056-1	3H091209-2PS1	0912056-2	3H091209-2PS2
3H091209-2CB1MB	3H091209-2PS3	3H091209-2CB2MB	3H091209-2PS4
3H091209-2CB3MB	3H091209-2PS5	3H091209-2MB	3H091209-2PS6
3H091209-2LCS	3H091209-2PS7	3H091209-2LCSD	3H091209-2PS8

Radiochemistry Prep Worksheet

Prep Batch: 31091209-2

ALS Laboratory Group -- FC

Prep Procedure: H3

Reviewed By: GDW/CDW Review Date: 12/9/2009

Non-Routine Pre-Treatment? Y / N Batch: N/A Re-Prep? Y / N Batch: N/A Prep QASS/ NCR? 0 / N 379412

Prep SOP: PAI 700 Rev: 10 Prep Analyst: Gabriel D. Wagner CDW Cocktaill: UG LLT
 Prep SOP: NONE Prep Date: 12/9/2009 Cocktaill Pipet: T-002
 Matrix Class: solid Prep Dept: RS Aliquot Pipet: RS-015

Balance: 21
 Balance: _____

Sampl Num	Prep Num	LabID	QC Type	Dish No.	Init Aliq sample	Fin Aliq sample	Prep Basis	Water Added(ml)	Moisture(%)	Analysis Vol.(ml)	Standards	Prep Notes
1	1	0912056-1	SMP		1	1	As Received					
2	1	0912056-2	SMP		1	1	As Received					
3	1	3H091209-2CB1	MB		1	1	As Received					
4	1	3H091209-2CB2	MB		1	1	As Received					
5	1	3H091209-2CB3	MB		1	1	As Received					
6	1	3H091209-2	MB		1	1	As Received					
7	1	3H091209-2	LCS		1	1	As Received					
8	1	3H091209-2	LCSD		1	1	As Received					

Comments
 UG LLT #97-090301// DUE TO LIMITED VOLUME A DUPLICATE AND MS WERE NOT PREPARED FOR THIS BATCH. A LCSD WAS PREPARED INSTEAD.

Spiked By: Gabriel D. Wagner CDW Date: 12/9/2009
 Witnessed By: Peter Workman Date: 12/9/2009

Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date	Aliquot Units	Pipet ID
S1	H-3	648.3610.44	2.517.810	DPM/ml	12/09/09	0.1 ml	RS-013

Radiochemistry Prep Worksheet

Prep Batch: 3H091209-2

ALS Laboratory Group -- FC

Prep Batch Not Validated!!!

Prep Procedure: H3

Review Date:

Reviewed By:

Non-Routine Pre-Treatment? Y / N Batch: _____ Re-Prep? Y / N Batch: _____ Prep QASS / NCR? Y / N _____

Prep SOP: PAI 700 Rev: 10 Prep Analyst: Gabriel D. Wagner *GDW* Balance: 21 Cocktail: UG LLT
 Prep SOP: NONE Prep Date: 12/9/2009 Balance: _____ Cocktail Pipet: T-002
 Matrix Class: solid Prep Dept: RS Aliquot Pipet: RS-015

Prep Num	LabID	QC Type	Dish No.	Init Aliq sample	Fin Aliq sample	Prep Basis	Water Added(ml)	Moisture(%)	Analysis Vol.(ml)	Standards	Prep Notes
1	0912056-1	SMP		1	1	As Received					
2	0912056-2	SMP		1	1	As Received					
3	3H091209-2CB1	MB		1	1	As Received					
4	3H091209-2CB2	MB		1	1	As Received					
5	3H091209-2CB3	MB		1	1	As Received					
6	3H091209-2	MB		1	1	As Received					
7	3H091209-2	LCS		1	1	As Received				S1	
8	3H091209-2	LCSD		1	1	As Received				S1	

Comments: 97-090301 *GDW* 12-09-09
 UG LLT #87-000704-11 DUE TO LIMITED VOLUME A DUPLICATE AND MS WERE NOT PREPARED FOR THIS BATCH. A LCSD WAS PREPARED INSTEAD.

Spiked By: Gabriel D. Wagner *GDW* Date: 12-09-09
 Witnessed By: *RSC* Date: 12/9/09

Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date	Alcohol	Units	Pipet ID
S1	H-3	648.3610.44	2.517.810	DP/Minl	12/09/09	0.1	ml	RS-013

exp. 10/21/10

SAMPLE CONDITION FORM (SOLIDS)

ANALYST: **60W**

ANALYSIS DATE: **12-09-2009**

METHOD: **3H**

WORK ORDER	SAMPLE ID	SAMPLE CONDITION		
		Dry/Wet	TEXTURE	Remarks
0912056	1	as rec'd	swipe / solid	none
↓	2	↓	↓	↓
<div style="position: absolute; top: 50%; left: 50%; transform: translate(-50%, -50%); opacity: 0.5;"> <p>(60W) 12-09-09</p> </div>				



Section 8

STANDARDS TRACEABILITY DOCUMENTS



Prepare a working dilution of ~ 5000 dpm/ml of Tritium from 648.2382.75.

1) Density of DI water

Mass of 100 ml Vol flask	66.4336g	26
Mass of flask + water	166.1135g	+
Net Mass	99.6799	↓
$\rho = 0.9968 \frac{g}{ml}$		

2) Mass of Std Transferred

Mass of Empty 500 ml Glass	260.38g	26
Mass of Glass + Std	266.45g	+
Net Mass Transferred	6.07g	↓

3) Dilute to Fin. Vol. w/ DI water

Mass of Glass Std, Diluent	733.7g	26
Mass of Glass	260.38g	↓
Net Mass of New Dilution	473.32g	↓

Final Activity Calc

$$\left(369,530.45 \frac{dpm}{ml} \right) \left(0.9968 \frac{g}{ml} \right) \left(6.07g \right) \left(0.9958 \frac{g}{ml} \right) \left(473.32g \right) = 4743.7 \frac{dpm}{ml}$$

RG 10/23/09

Std ID: 648.3610.44

RG 10/23/09

Description: 3H

Expiration: 10/21/2010

Activity: 4743.70 dpm/mL

2s Uncertainty: 34.15 dpm/mL

Ref. Date: 9/3/1998

Ref Time: N/A

Prep Date: 9/29/2009 Prep by: JD

Matrix/Comp. DI Water

Half Life (y): 1.23E+01

Reverification Log		
Analysis Date	Initials	Expiration Date

RG 10/23/09

RG 10/23/09

Continued on Page

Read and Understood By

Signed

9/29/09

Date

Signed

10/23/09

Date

PROJECT 648.2382.75 3H

Continued From Page

Prepare a ¹⁰ dilution ~~working level~~ dilution (of approximately 2500 DPM/mL) tritium standard using RSD #648 and diluting with DI water.

1) Determine the density of DI water:

Mass of empty 100 mL class A volumetric flask	67.1522 g	Pal #13
Mass of flask + water	116.7351 g	↓
Net mass of 100 mL H ₂ O	99.5829 g	
	$\rho = .9958 \text{ g/mL}$	

2) Transfer contents of Ampule SRM 4927F to a 40 mL VOA vial - 500 mL glass amber bottle

mass of 40 mL VOA vial w/o lid	255.04 g	Pal #26
mass of open ampule + 50 mL beaker	40.9075 g	Pal #12
mass of empty ampule + 50 mL beaker	36.0148 g	↓
Net mass of Std.	4.8927 g	

3) Dilute Std. with DI water

mass of bottle w/o lid	255.04 g	Pal #26
mass of bottle, std, + DI water	757.1 g	↓
Net mass of Std + DI water	502.1 g	

4. Final Activity Calculation.

$$\frac{(634.7 \text{ kBq/g})(1000 \text{ Bq/kBq})(60 \text{ DPM/Bq})(4.8927 \text{ g})(.9958 \text{ g/mL})}{502.1 \text{ g}} = 369,530.45 \frac{\text{DPM}}{\text{mL}}$$

U.S. Department of Commerce
National Institute of Standards
and Technology
SRM 4927F
Hydrogen-3

<4 MBq in distilled water

CAUTION
RADIOACTIVE



Read and Understood By

Chad [Signature]
Signed

3/27/03
Date

Renee [Signature]
Signed

3/27/03
Date

Continued on Page



National Institute of Standards & Technology

Certificate

PA ID 0648
17-04-02

Standard Reference Material 4927F Hydrogen-3 Radioactivity Standard

This Standard Reference Material (SRM) consists of radioactive hydrogen-3, as water, in 5 mL of distilled water. The solution is contained in a flame-sealed NIST borosilicate-glass ampoule. The SRM is intended for the calibration of beta-particle counting instruments and for the monitoring of radiochemical procedures.

Radiological Hazard

The SRM ampoule contains hydrogen-3 with a total activity of approximately 3.2 MBq. Hydrogen-3 decays by beta-particle emission. None of the beta particles escape from the SRM ampoule. During the decay process no photons are emitted. Approximate unshielded dose rates at several distances (as of the reference time) are given in note [a]*. There is no detectable external radiation. The SRM should be used only by persons qualified to handle radioactive material.

Chemical Hazard

The SRM ampoule contains only distilled water. There is no chemical hazard. If the ampoule is to be opened to transfer the solution, the recommended procedure is given on page 2.

Storage and Handling

The SRM should be stored and used at a temperature between 5° and 65 °C. The solution in an unopened ampoule should remain stable and homogeneous until at least September 2008.

The ampoule (or any subsequent container) should always be clearly marked as containing radioactive material. If the ampoule is transported it should be packed, marked, labeled, and shipped in accordance with the applicable national, international, and carrier regulations. The solution in the ampoule is a dangerous good (hazardous material) because of the radioactivity.

Preparation

This Standard Reference Material was prepared in the Physics Laboratory, Ionizing Radiation Division, Radioactivity Group, L.R. Karam, Group Leader. The overall technical direction and physical measurements leading to certification were provided by L.L. Lucas and M.P. Unterweger of the Radioactivity Group.

The support aspects involved in the preparation, certification, and issuance of this SRM were coordinated through the Standard Reference Materials Program by J.W.L. Thomas.

Bert M. Coursey, Chief
Ionizing Radiation Division

Nancy M. Trahey, Chief
Standard Reference Materials Program

Gaithersburg, Maryland 20899
June 1999
Half-life and text revised October 2000

Recommended Procedure for Opening the SRM Ampoule

- 1) If the SRM solution is to be diluted, it is recommended that the diluting solution have a composition comparable to that of the SRM solution.
- 2) Wear eye protection, gloves, and protective clothing and work over a tray with absorbent paper in it. Work in a fume hood.
- 3) Shake the ampoule to wet all of the inside surface of the ampoule. Return the ampoule to the upright position.
- 4) Check that all of the liquid has drained out of the neck of the ampoule. If necessary, gently tap the neck to speed the process.
- 5) Holding the ampoule upright, score the narrowest part of the neck with a scribe or diamond pencil.
- 6) Lightly wet the scored line. This reduces the crack propagation velocity and makes for a cleaner break.
- 7) Hold the ampoule upright with a paper towel, a wiper, or a support jig. Position the scored line away from you. Using a paper towel or wiper to avoid contamination, snap off the top of the ampoule by pressing the narrowest part of the neck away from you while pulling the tip of the ampoule towards you.
- 8) Transfer the solution from the ampoule using a pycnometer or a pipet with dispenser handle. **NEVER PIPETTE BY MOUTH.**
- 9) Seal any unused SRM solution in a flame-sealed glass ampoule, if possible, to minimize the evaporation loss.

See also reference [4]*.

PROPERTIES OF SRM 4927F

Certified values

Solution density	$(0.998 \pm 0.002) \text{ g} \cdot \text{mL}^{-1}$ at 20.0 °C [b]*
Radionuclide	Hydrogen-3
Reference time	1200 EST, 3 September 1998
Massic activity of the solution [c]	$634.7 \text{ kBq} \cdot \text{g}^{-1}$
Relative expanded uncertainty ($k=2$)	0.72% [d] [e]

Uncertified values

Physical Properties:			
Source description	Liquid in flame-sealed NIST borosilicate-glass ampoule		
Ampoule specifications	Body outside diameter	$(16.5 \pm 0.5) \text{ mm}$	
	Wall thickness	$(0.60 \pm 0.04) \text{ mm}$	
	Barium content	Less than 2.5%	
	Lead-oxide content	Less than 0.02%	
	Other heavy elements	Trace quantities	
Solution mass	Approximately 5.0 g		
Chemical Properties:			
Solution composition	Chemical Formula	Concentration ($\text{mol} \cdot \text{L}^{-1}$)	Mass Fraction ($\text{g} \cdot \text{g}^{-1}$)
	H ₂ O HHO	55 6×10^{-7}	1500 1×10^{-5}
Radiological Properties:			
Radionuclidic impurities	None detected [f]		
Half lives used	Hydrogen-3: $(4500 \pm 8) \text{ d}$ [g]		
Calibration method and measuring instrument(s)	4πβ gas counting of SRM 4927E using the NIST length-compensated internal gas proportional counters and intercomparison of SRMs 4927E/4927F using two 4πβ liquid-scintillation counting systems [h]		

EVALUATION OF THE UNCERTAINTY OF THE MASSIC ACTIVITY [d]*

Input Quantity x_i , the source of uncertainty (and individual uncertainty components where appropriate)	Method Used To Evaluate $u(x_i)$, the standard uncertainty of x_i , (A) denotes evaluation by statistical methods, (B) denotes evaluation by other methods	Relative Uncertainty Of Input Quantity, $u(x_i)/x_i$, (%) [j]	Relative Sensitivity Factor, $ \partial y/\partial x_i \cdot$ (x_i/y) [j]	Relative Uncertainty Of Output Quantity, $u_i(y)/y$, (%) [k]
Massic count rate of SRM 4927E, corrected for background and decay [h]	Standard deviation of the mean for 23 sets of gas counting measurements (A)	0.18	1.0	0.18
Gram-mole measurements	Estimated (B)	0.20	1.0	0.20
Live-time [p]	Estimated (B)	0.10	1.0	0.10
Extrapolation of count-rate-versus-energy to zero energy	Estimated (B)	0.20	1.0	0.20
Half life of H-3	Standard uncertainty of the half life (A)	0.18 [m]	0.009 [n]	0.002
Liquid-scintillation intercomparison of SRM 4927F and SRM-4927E	Standard deviation of the mean for 7 sets of liquid-scintillation measurements (A)	0.06	1.0	0.06
Radionuclidic impurities	Limit of detection (B) [q]	100.	0.0005	0.05
Relative Combined Standard Uncertainty of the Output Quantity, $u_c(y)/y$, (%)				0.36
Coverage Factor, k				<u>2</u>
Relative Expanded Uncertainty of the Output Quantity, U/y , (%)				0.72

NOTES

- [a] The Sievert is the SI unit for dose equivalent. See reference [1]. One μSv is equal to 0.1 mrem.
 Distance from Ampoule (cm): Γ 30 100
 Approximate Dose Rate ($\mu\text{Sv/h}$): <0.1 (Not detectable)
- [b] The stated uncertainty is two times the standard uncertainty.
- [c] Massic activity is the preferred name for the quantity activity divided by the total mass of the sample. See reference [1].
- [d] The reported value, y , of massic activity (activity per unit mass) at the reference time was not measured directly but was derived from measurements and calculations of other quantities. This can be expressed as $y = f(x_1, x_2, x_3, \dots, x_n)$, where f is a mathematical function derived from the assumed model of the measurement process.
- The value, x_i , used for each input quantity, i has a standard uncertainty, $u(x_i)$, that generates a corresponding uncertainty in y , $u_i(y) = |\partial y / \partial x_i| \cdot u(x_i)$, called a component of combined standard uncertainty of y .
- The combined standard uncertainty of y , $u_c(y)$, is the positive square root of the sum of the squares of the components of combined standard uncertainty.
- The combined standard uncertainty is multiplied by a coverage factor of $k = 2$ to obtain U , the expanded uncertainty of y .
- Since it can be assumed that the possible estimated values of the massic activity are approximately normally distributed with approximate standard deviation $u_c(y)$, the unknown value of the massic activity is believed to lie in the interval $y \pm U$ with a level of confidence of approximately 95 percent.
- For further information on the expression of uncertainties, see references [2] and [3].
- [e] The value of each standard uncertainty component, and hence the value of the expanded uncertainty itself, is a best estimate based upon all available information, but is only approximately known. That is to say, the "uncertainty of the uncertainty" is large and not well known. This is true for uncertainties evaluated by statistical methods (e.g., the relative standard deviation of the standard deviation of the mean for the massic response is approximately 50%) and for uncertainties evaluated by other methods (which could easily be over estimated or under estimated by substantial amounts). The unknown value of the expanded uncertainty is believed to lie in the interval $U/2$ to $2U$ (i.e., within a factor of 2 of the estimated value).
- [f] The estimated limit of detection for radionuclidic impurities is $300 \text{ Bq}\cdot\text{g}^{-1}$.
- [g] The stated uncertainty is the standard uncertainty. See reference [5].
- [h] Extensive gas-counting measurements were made on the SRM 4927E solution during 1998 and 1999. The SRM 4927F solution was intercompared with the SRM 4927E solution using liquid-scintillation counting.
- [i] Relative standard uncertainty of the input quantity x_i .

- [j] The relative change in the output quantity y divided by the relative change in the input quantity x_i . If $|\partial y/\partial x_i| \cdot (x_i/y) = 1.0$, then a 1% change in x_i results in a 1% change in y . If $|\partial y/\partial x_i| \cdot (x_i/y) = 0.05$, then a 1% change in x_i results in a 0.05% change in y .
- [k] Relative component of combined standard uncertainty of output quantity y , rounded to two significant figures or less. The relative component of combined standard uncertainty of y is given by $u_c(y)/y \approx |\partial y/\partial x_i| \cdot u(x_i)/y = |\partial y/\partial x_i| \cdot (x_i/y) \cdot u(x_i)/x_i$. The numerical values of $u(x_i)/x_i$, $|\partial y/\partial x_i| \cdot (x_i/y)$, and $u_c(y)/y$, all dimensionless quantities, are listed in columns 3, 4, and 5, respectively. Thus, the value in column 5 is equal to the value in column 4 multiplied by the value in column 3. The input quantities are independent, or very nearly so. Hence the covariances are zero or negligible.
- [m] The relative standard uncertainty of $\lambda \cdot t$ is determined by the relative standard uncertainty of λ (i.e., of the half life). The relative standard uncertainty of t is negligible.
- [n] $|\partial y/\partial x_i| \cdot (x_i/y) = |\lambda \cdot t|$
- [p] The live time is determined by counting the pulses from a gated crystal-controlled oscillator.
- [q] The standard uncertainty for each undetected impurity that might reasonably be expected to be present is estimated to be equal to the estimated limit of detection for that impurity, i.e. $u(x_i)/x_i = 100\%$. $|\partial y/\partial x_i| \cdot (x_i/y) = \{(\text{response per Bq of impurity})/(\text{response per Bq of H-3})\} \cdot \{(\text{Bq of impurity})/(\text{Bq of H-3})\}$. Thus $u_c(y)/y$ is the relative change in y if the impurity were present with a massic activity equal to the estimated limit of detection.

REFERENCES

- [1] International Organization for Standardization (ISO), *ISO Standards Handbook - Quantities and Units*, 1993. Available from the American National Standards Institute, 11 West 42nd Street, New York, NY 10036, U.S.A. 1-212-642-4900.
- [2] International Organization for Standardization (ISO), *Guide to the Expression of Uncertainty in Measurement*, 1993. Available from the American National Standards Institute, 11 West 42nd Street, New York, NY 10036, U.S.A. 1-212-642-4900. (Listed under ISO miscellaneous publications as "ISO Guide to the Expression 1993")
- [3] E. N. Taylor and C. E. Kuyatt, *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*, NIST Technical Note 1297, 1994. Available from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20407, U.S.A.
- [4] National Council on Radiation Protection and Measurements Report No. 58, *A Handbook of Radioactivity Measurements Procedures*, Second Edition, 1985. Available from the National Council on Radiation Protection and Measurements, 7910 Woodmont Avenue, Bethesda, MD 20814 U.S.A.
- [5] L.L. Lucas and M.P. Unterwieser, *Comprehensive Review and Critical Evaluation of the Half-Life of Tritium*, J. Res. Natl. Inst. Stand. Technol. 105, 541-549 (2000).



Section 9

ADDITIONAL SUPPORTING DOCUMENTATION



Liquid Scintillation Counter

Instrumentation Calibration

**Initial Efficiency Calibration
Standards Traceability**

H-3 Swipe Quench Curve Background and Efficiency Determination

LS6000
12/28/2009

Polynomial Coefficient		H # Range
Efficiency	Background	
x^0	3.7192E-01	Low = 113.6
x^1	-1.1270E-03	High = 283.3

Calib. Date : 10/26/2009

Sample ID	Pos #	H#	Obs.CPM	Corr. BCPM	CPM Corr.Fact.
3H091209-2CB1	36-1	129.8	6.58	6.972	0.392
3H091209-2CB2	36-5	129.7	6.75	6.971	0.221
3H091209-2CB3	36-8	130.0	7.45	6.973	-0.477
Average=					0.045

Sample ID	Pos #	H #	Efficiency	Background	H # Check
0912056-1	36-2	115.8	0.2414	6.829	OK
0912056-2	36-3	117.1	0.2399	6.838	OK
3H091209-2MB	36-4	127.8	0.2279	6.912	OK
3H091209-2LCS	36-6	126.3	0.2296	6.902	OK
3H091209-2LCSD	36-7	126.3	0.2296	6.902	OK

3H091209-2_H3-5ml_Swipes

H-3 Swipes "Window 2" Control Limits (LS 6000)

The background count rate is determined from the average of the reagent blanks for the batch.

Window 2 control limits are established using the average count rate from the three reagent blanks associated with each prep batch +/- 3X the estimated poisson uncertainty.

Updated 10/29/09 mh

COUNT DATE	#	Sample ID	Count Duration (min.)	Average count Duration (min.)	Count Rate (CPM)	Batch Average Reagent Blank	Lower Control Limit	Upper Control Limit	PASS/ FAIL
12/24/2009	22	3H091209-2CB1	60		30.27				
12/24/2009	23	3H091209-2CB2	60		29.78				
12/25/2009	24	3H091209-2CB3	60	60	28.15	29.40	27.30	31.50	PASS

Tritium Swipe (Glass Vial) Quench Curve

10/26/2009

Beckman LS6000

Standard: 699.3020.95

Ref. Date: 9/3/1998

Spike Act.: 160351.040 dpm/mL

Spike Vol.: 0.10000 mL

Non-Spiked/ Background

Bkg. Coefficients



Ax= 6.9704E-03

B= 6.0669E+00

Sample ID	H#	CPM	Calc. Bkg.	Sigma Diff.	Spiked / Efficiency	Sample ID	H#	CPM	Corr. Bkg. Corr. CPM	DPM	Efficiency	Calc. Eff.	% Diff.	Sam. ID	CPM
3H090422-2B01	113.5	6.58	6.86	0.84	0916022-1	113.6	1992.88	1986.02	8556.38	0.2321	0.2439	5.08%	3H090422-2B01	26.75	
3H090422-2B02	145.1	7.28	7.08	-0.58	0916022-2	142.4	1885.90	1878.84	8556.38	0.2196	0.2114	-3.71%	3H090422-2B02	27.73	
3H090422-2B03	166.9	7.78	7.23	-1.53	0916022-3	159.4	1731.13	1723.95	8556.38	0.2015	0.1923	-4.57%	3H090422-2B03	26.78	
3H090422-2B04	179.4	7.83	7.32	-1.42	0916022-4	179.9	1499.31	1491.99	8556.38	0.1744	0.1692	-2.98%	3H090422-2B04	27.63	
3H090422-2B05	198.1	7.1	7.45	1.01	0916022-5	197.1	1286.78	1278.34	8556.38	0.1494	0.1498	0.26%	3H090422-2B05	27.60	
3H090422-2B06	214.4	7.12	7.56	1.28	0916022-6	215.7	1071.89	1064.32	8556.38	0.1244	0.1288	3.56%	3H090422-2B06	27.55	
3H090422-2B07	232.7	7.17	7.69	1.50	0916022-7	233.0	894.86	887.17	8556.38	0.1037	0.1093	5.44%	3H090422-2B07	29.35	
3H090422-2B08	248.2	7.17	7.80	1.81	0916022-8	250.4	737.97	730.16	8556.38	0.0863	0.0897	5.13%	3H090422-2B08	28.17	
3H090422-2B09	258.7	8.35	7.87	-1.29	0916022-9	261.4	644.34	636.45	8556.38	0.0744	0.0773	3.95%	3H090422-2B09	29.07	
3H090422-2B10	275.8	7.8	7.99	0.53	0916022-10	278.5	528.08	520.07	8556.38	0.0608	0.0580	-4.50%	3H090422-2B10	29.12	
3H090422-2B11	283.3	8.7	8.04	-1.73	0916022-11	287	459.26	451.19	8556.38	0.0527	0.0485	-8.09%	3H090422-2B11	27.43	
3H090422-2B12*	295.9	8.72	8.13	-1.55	0916022-12*	303.6	393.26	385.08	8556.38	0.0450	0.0298	-33.88%	3H090422-2B12*	28.15	

*These data points are not used in the calibration due to poor fit.

(BKG CLs = +/- 3 sigma)
Max 1.81
Min -1.73



Quench Control Limits	
Upper	283.3
Lower	113.6

Analysis Window Settings	
WIN 1 (50-250)	
WIN 2 (450-900)	

Mark Sykes
Instrument Technician

10-29-09
Date

Mark Sykes
Supervisory Review
Date 10/30/09

³H Swipe Efficiency Calibration Verification / Method Blank Verification 10/27/09

Calibration Source Check

L86000
 Analysis Date: 10/27/2009
 Nuclide: ³H
 Half Life: 1.230E+01 yr.

Calibration Check Source:

Spike Standard: 648.3610.05
 Reference Date: 9/3/1998
 Spiked DPM: 5141.07 dpm/mL
 Spike Volume: 0.1 mL
 Spiked into: 1 sample
 Current Spk. Act.: 127.19 pCi/sample

IU= 0.112
 PU= 0.102

Calibration Check Source Count

Sample ID	Rack	Pos	Prep Date	Cnt. Dur.	Anal. Vol.	GrsCPM	BkgCPM	Efficiency	Activity	k (denom.)	Chem. Yield	LCS Recovery:	Pass/Fail	Units	2σ CU	IU	PU	2σ TPU
0916023-1	25	2	4/22/2009	180	1	70.06	7.23	0.2216	131.50	0.478	100%	103.4%	PASS	pCi/sample	2.743	14.72765	13.41268	20.108
0916023-2	25	3	4/22/2009	180	1	47.93	7.67	0.1497	124.73	0.323	100%	98.1%	PASS	pCi/sample	3.444	13.96964	12.72235	19.206
0916023-3	25	4	4/22/2009	180	1	25.13	8.21	0.0622	126.09	0.134	100%	99.1%	PASS	pCi/sample	6.415	14.12263	12.86168	20.150

IU= 0.112
 PU= 0.102

Method Blank Check Count

Sample ID	Rack	Pos	Prep Date	Cnt. Dur.	Anal. Vol.	GrsCPM	BkgCPM	Efficiency	Chem. Yield	k (denom.)	activity	MDC	Pass/Fail	Units	2σ CU	IU	PU	2σ TPU
3H090422-3B01	25	6	4/22/2009	180	1	7.52	7.18	0.2298	100%	0.496	0.6957	1.95	PASS	pCi/sample	1.153	0.07792	0.07096	1.158
3H090422-3B02	25	7	4/22/2009	180	1	7.93	7.68	0.1476	100%	0.318	0.7746	3.11	PASS	pCi/sample	1.850	0.08676	0.07901	1.854
3H090422-3B03	25	8	4/22/2009	180	1	8.12	8.24	0.0574	100%	0.124	-0.9835	8.10	PASS	pCi/sample	4.874	-0.11016	-0.10032	4.876

H-3 Swipe Quench Curve Background and Efficiency Determination

LS6000
10/28/2009

Polynomial Coefficient		H # Range
Efficiency	Background	
x^0	3.7192E-01	Low = 113.6
x^1	-1.1270E-03	High = 283.3

Calib. Date: 10/26/2009

Sample ID	Pos #	H#	Obs.CPM	Corr. BCPM	CPM Corr.Fact.
3H090422-3CB1	25-1	132.3	7.16	6.989	-0.171
3H090422-3CB2	25-5	129.3	7.38	6.968	-0.412
3H090422-3CB3	25-9	128.8	7.07	6.965	-0.105
Average=					-0.229

Sample ID	Pos #	H #	Efficiency	Background	H # Check
0916023-1	25-2	133.4	0.2216	7.226	OK
0916023-2	25-3	197.2	0.1497	7.671	OK
0916023-3	25-4	274.8	0.0622	8.212	OK
3H090422-3B01	25-6	126.1	0.2298	7.175	OK
3H090422-3B02	25-7	199.0	0.1476	7.683	OK
3H090422-3B03	25-8	279.1	0.0574	8.242	OK

ID: 3H: 5-ML, 10-ML

26 OCT 2009 10:12

USER: 4

COMMENT: LS6000

PRESET TIME : 60.00

DATA CALC : CPM H# : YES SAMPLE REPEATS: 1 PRINTER : STD
 COUNT BLANK : NO IC# : NO REPLICATES : 1 RS232 : EDIT
 TWO PHASE : NO AQC : NO CYCLE REPEATS : 1
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0
 LOW LEVEL : YES HALF LIFE CORRECTION DATE: none

CHAN: 50.0 - 250.0 %ERROR: 1.75 FACTOR: 1.000000 BKG. SUB: 0
 CHAN: 450.0 - 900.0 %ERROR: 20.00 FACTOR: 1.000000 BKG. SUB: 0

ALPHA-BETA DISCRIMINATION: NO

SAM NO	POS	TIME MIN	H#	WIND1		WIND2		LUMEX %	ELAPSED TIME
				CPM	%ERROR	CPM	%ERROR		
1	34-1	6.60	113.6	1992.88	1.74	27.73	14.78	0.08	7.20
2	34-2	6.95	142.4	1885.90	1.75	27.34	14.51	0.06	14.85
3	34-3	7.55	159.4	1731.13	1.75	28.48	13.64	0.05	23.07
4	34-4	8.75	179.9	1499.31	1.75	29.37	12.48	0.05	32.52
5	34-5	10.20	197.1	1285.78	1.75	26.96	12.06	0.05	43.43
6	34-6	12.20	215.7	1071.89	1.75	26.07	11.22	0.05	56.36
7	34-7	14.60	233.0	894.86	1.75	30.41	9.49	0.05	71.71
8	34-8	17.75	250.4	737.97	1.75	30.03	8.66	0.06	90.23
9	34-9	20.30	261.4	644.34	1.75	26.80	8.57	0.05	111.34
10	34-10	24.75	278.5	528.08	1.75	27.47	7.67	0.05	136.95
11	34-11	28.45	287.0	459.26	1.75	25.94	7.36	0.05	166.25
12	34-12	33.25	303.6	393.26	1.75	27.34	6.63	0.05	200.41
13	50-1	60.00	113.5	6.58	10.06	26.75	4.99	0.84	261.76
14	50-2	60.00	145.1	7.28	9.57	27.73	4.90	0.63	322.99
15	50-3	60.00	166.9	7.78	9.25	26.78	4.99	0.39	384.17
16	50-4	60.00	179.4	7.83	9.23	27.63	4.91	0.24	445.34
17	50-5	60.00	198.1	7.10	9.69	27.60	4.91	0.22	506.46
18	50-6	60.00	214.4	7.12	9.68	27.55	4.92	0.23	567.60
19	50-7	60.00	232.7	7.17	9.64	29.35	4.77	0.21	628.73
20	50-8	60.00	248.2	7.17	9.64	29.17	4.78	0.24	689.85
21	50-9	60.00	258.7	8.35	8.94	29.07	4.79	0.20	750.98
22	50-10	60.00	275.8	7.80	9.25	29.12	4.79	0.21	812.09
23	50-11	60.00	283.3	8.70	8.75	27.43	4.93	0.19	873.21
24	50-12	60.00	295.9	8.72	8.75	28.15	4.87	0.18	934.33

060-04-102601

WA 10-27-89

B60_04_102601

```

BSF Version           : 3
Instrument Type       : LS 6000
Data Capture Date    : 26 Oct 2009 09:14:01
User Filename        : C:\...\LS WINCONNECTION\DATA\USER04\UN102601.BSF
User Number          : 4
User Id              : 3H:5-ML,10-ML
User Comments        : LS6000
Preset Count Time    : 60.00
Calculation Mode     : CPM
H# Selected          : YES
Sample Repeats       : 1
Printer Output Mode  : STD
Blank Count          : NO
IC# or SCR Selected  : NO
Replicates           : 1
RS232 Output Mode    : EDIT
Two-Phase Selected   : NO
AQC Choice           : NO
Cycle Repeats        : 1
Scintillator choice  : LIQUID
Lumex Selected       : NO
Low Sample Reject Count : 0
Low Level Selection  : YES
Half Life Correction Date : none
Window Limits Window 1 : 50.00
Preset %Error Iso1   : 1.75
Norm Multiplier Iso1 : 1.00000
Background CPM 1     : 0.00
Window Limits Window 2 : 450.00
Preset %Error Iso2   : 20.00
Norm Multiplier Iso2 : 1.00000
Background CPM 2     : 0.00
Alpha/Beta Discrimination : NO

```

Sam	Rack	Time	H#	CPM Iso1	%Err1	CPM Iso2	%Err2	LumEx	ETime
1	34-1	6.60	113.6	1992.88	1.74	27.73	14.78	0.08	7.20
2	34-2	6.95	142.4	1885.90	1.75	27.34	14.51	0.06	14.85
3	34-3	7.55	159.4	1731.13	1.75	28.48	13.64	0.05	23.07
4	34-4	8.75	179.9	1499.31	1.75	29.37	12.48	0.05	32.52
5	34-5	10.20	197.1	1285.78	1.75	26.96	12.06	0.05	43.43
6	34-6	12.20	215.7	1071.89	1.75	26.07	11.22	0.05	56.36
7	34-7	14.60	233.0	894.86	1.75	30.41	9.49	0.05	71.71
8	34-8	17.75	250.4	737.97	1.75	30.03	8.66	0.06	90.23
9	34-9	20.30	261.4	644.34	1.75	26.80	8.57	0.05	111.34
10	34-10	24.75	278.5	528.08	1.75	27.47	7.67	0.05	136.95
11	34-11	28.45	287.0	459.26	1.75	25.94	7.36	0.05	166.25
12	34-12	33.25	303.6	393.26	1.75	27.34	6.63	0.05	200.41
13	50-1	60.00	113.5	6.58	10.06	26.75	4.99	0.84	261.76
14	50-2	60.00	145.1	7.28	9.57	27.73	4.90	0.63	322.99
15	50-3	60.00	166.9	7.78	9.25	26.78	4.99	0.39	384.17
16	50-4	60.00	179.4	7.83	9.23	27.63	4.91	0.24	445.34
17	50-5	60.00	198.1	7.10	9.69	27.60	4.91	0.22	506.46
18	50-6	60.00	214.4	7.12	9.68	27.55	4.92	0.23	567.60
19	50-7	60.00	232.7	7.17	9.64	29.35	4.77	0.21	628.73
20	50-8	60.00	248.2	7.17	9.64	29.17	4.78	0.24	689.85
21	50-9	60.00	258.7	8.35	8.94	29.07	4.79	0.20	750.98
22	50-10	60.00	275.8	7.80	9.25	29.12	4.79	0.21	812.09
23	50-11	60.00	283.3	8.70	8.75	27.43	4.93	0.19	873.21
24	50-12	60.00	295.9	8.72	8.75	28.15	4.87	0.18	934.33

UN 10-27-09

ID: 3H: 5-ML, 10-ML

27 OCT 2009 02:41

USER: 8

COMMENT: LS6000

PRESET TIME : 180.00

DATA CALC : CPM H# : YES SAMPLE REPEATS : 1 PRINTER : STD
 COUNT BLANK : NO IC# : NO REPLICATES : 1 RS232 : EDIT
 TWO PHASE : NO AQC : NO CYCLE REPEATS : 1
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0
 LOW LEVEL : YES HALF LIFE CORRECTION DATE: none

CHAN: 50.0 - 250.0 %ERROR: 1.75 FACTOR: 1.000000 BKG. SUB: 0
 CHAN: 450.0 - 900.0 %ERROR: 20.00 FACTOR: 1.000000 BKG. SUB: 0

ALPHA-BETA DISCRIMINATION: NO

SAM NO	POS	TIME MIN	H#	WIND1		WIND2		LUMEX %	ELAPSED TIME
				CPM	%ERROR	CPM	%ERROR		
1	25-1	180.00	132.3	7.16	5.57	29.98	2.72	0.36	182.24
2	25-2	180.00	133.4	70.06	1.78	30.97	2.68	0.14	364.55
3	25-3	180.00	197.2	47.93	2.15	28.71	2.78	0.13	546.79
4	25-4	180.00	274.8	25.13	2.97	28.56	2.79	0.14	728.98
5	25-5	180.00	129.3	7.38	5.49	28.89	2.77	0.28	911.24
6	25-6	180.00	126.1	7.52	5.44	30.03	2.72	0.26	1093.50
7	25-7	180.00	199.0	7.93	5.29	30.23	2.71	0.17	1275.71
8	25-8	180.00	279.1	8.12	5.23	29.88	2.73	0.14	1457.83
9	25-9	180.00	128.8	7.07	5.61	30.01	2.72	0.17	1640.02

660-08-101701

10-28-09

B60_08_102701

BSF Version : 3
Instrument Type : LS 6000
Data Capture Date : 27 Oct 2009 02:45:15
User Filename : C:\...\LS WINCONNECTION\DATA\USER08\UN102701.BSF
User Number : 8
User Id : 3H:5-ML,10-ML
User Comments : LS6000
Preset Count Time : 180.00
Calculation Mode : CPM
H# Selected : YES
Sample Repeats : 1
Printer Output Mode : STD
Blank Count : NO
IC# or SCR selected : NO
Replicates : 1
RS232 Output Mode : EDIT
Two-Phase Selected : NO
AQC Choice : NO
Cycle Repeats : 1
Scintillator Choice : LIQUID
Lumex Selected : NO
Low Sample Reject Count : 0
Low Level Selection : YES
Half Life Correction Date : none
Window Limits Window 1 : 50.00
Preset %Error Iso1 : 1.75
Norm Multiplier Iso1 : 1.00000
Background CPM 1 : 0.00
Window Limits Window 2 : 450.00
Preset %Error Iso2 : 20.00
Norm Multiplier Iso2 : 1.00000
Background CPM 2 : 0.00
Alpha/Beta Discrimination : NO

Sam	Rack	Time	H#	CPM Iso1	%Err1	CPM Iso2	%Err2	LumEx	ElTime
1	25-1	180.00	132.3	7.16	5.57	29.98	2.72	0.36	182.24
2	25-2	180.00	133.4	70.06	1.78	30.97	2.68	0.14	364.55
3	25-3	180.00	197.2	47.93	2.15	28.71	2.78	0.13	546.79
4	25-4	180.00	274.8	25.13	2.97	28.56	2.79	0.14	728.98
5	25-5	180.00	129.3	7.38	5.49	28.89	2.77	0.28	911.24
6	25-6	180.00	126.1	7.52	5.44	30.03	2.72	0.26	1093.50
7	25-7	180.00	199.0	7.93	5.29	30.23	2.71	0.17	1275.71
8	25-8	180.00	279.1	8.12	5.23	29.88	2.73	0.14	1457.83
9	25-9	180.00	128.8	7.07	5.61	30.01	2.72	0.17	1640.02

OK
10-28-09

LSC Run Log

Instrument ID: LS6000

379825

ALS Laboratory Group - Fort Collins

Date	Sample ID	Count Time (min.)	Rack & Position	Test	User #	Batch ID	Position Check	Initials	Comments
10-21-09	0909050-3	180	3S - 2	FE-SS	12	FE091016-1	NA	NA	
	-3D		- 3						
	-3REP dup		- 4						
	-3Rep2 dup2		- 5						
11-25-09	FE091016-1CB2		- 6						Recount, Lumex > 5% NA
	0909050-3MS	178.10	- 7						
	-3REP	180	- 8						
	-3REP2		- 9						Recount, Lumex > 5% NA
	0922005-1		- 10						
	FE091016-1MB		- 11						
	-1LCS	189.80	- 12						
	FE091016-1CB3	180	17 - 1						
10-23-09	Daily QC	10	173 -131-2		113		NA	NA	
10-23-09	0909050-3Rep2dup2	180	57 - 1	FE-SS	12	FE091016-1	NA	NA	Lumex still > 5% NA
	-3Rep2		- 2						
10-23-09	Daily QC	10	113 -131-2		113		NA	NA	
10-26-09			- 1						
10-26-09	0910022-1	6.60	34 - 1	H3	4	34090422-2	NA	NA	H3 glass vial swipe B. Curie calibration
	-2	6.95	- 2						
	-3	7.55	- 3						
	-4	8.75	- 4						
	-5	10.10	- 5						
	-6	12.10	- 6						
	-7	14.60	- 7						
	-8	17.75	- 8						
	-9	20.30	- 9						
	-10	24.75	- 10						
	-11	28.45	- 11						
	-12	33.25	- 12						
	31090402-2.mbi	60	50 - 1						

Analyst / Date NA 10-28-09

Note: Each page is copied as completed and included with the workorder/run documentation; reviewed subsequently.

LSC Run Log

Instrument ID: LS6000

379826

ALS Laboratory Group - Fort Collins

Date	Sample ID	Count Time (min.)	Rack & Position	Test	User #	Batch ID	Position Check	Initials	Comments
10-26-09	3H090422-2 MB2	60	50 - 2	Hβ	4	3H090422-3	NA	NA	Hβ glass vial B.C. swi p.e. Calibration
	-2 MB3		- 3						
	-2 MB4		- 4						
	-2 MB5		- 5						
	-2 MB6		- 6						
	-2 MB7		- 7						
	-2 MB8		- 8						
	-2 MB9		- 9						
	-2 MB10		- 10						
	-2 MB11		- 11						
	-2 MB12		- 12						
10-27-09	Daily QC	10	11b - 13, 12		1, 3		NA	NA	
10-27-09	3H090422-3 CB1	180	25 - 1	Hβ	8	3H090422-3	NA	NA	Hβ glass vial swipes ICN, TC, B's
	0916 023-1		- 2						
	- 2		- 3						
	- 3		- 4						
	3H090422-3 CB2		- 5						
	- 3 MB1		- 6						
	- 3 MB2		- 7						
	- 3 MB3		- 8						
	- 3 CB3		- 9						
10-28-09	Daily QC	10	11b - 13, 10, 12		1, 3		NA	NA	

Analyst / Date NA 10-28-09

FORM 762r6.xls (3/7/09)

Note: Each page is copied as completed and included with the workorder/run documentation; reviewed subsequently.

Radiochemistry Instrument Worksheet

ALS Laboratory Group -- FC

Prep Batch: 3H090422-2

Prep Procedure: H3 Glass Vial Swipe & Curve Calibration

Analytical QASS / NCR? Y 10 NA

Prep Num	QC Type	Init Aliq	Fin Aliq	Units	Report Units	Cnt 1 File/Inst	Cnt 1 Rack-Pos	Cnt 1 Chk By	Cnt 2 File/Inst	Cnt 2 Rack-Pos	Cnt 2 Chk By	Cnt 3 File/Inst	Cnt 3 Rack-Pos	Cnt 3 Chk By	Notes
1	0916022-1	SMP	1	1	sample pCi/l	LS6000	34-1	NA							
1	0916022-2	SMP	1	1	sample pCi/l		-2								
1	0916022-3	SMP	1	1	sample pCi/l		-3								
1	0916022-4	SMP	1	1	sample pCi/l		-4								
1	0916022-5	SMP	1	1	sample pCi/l		-5								
1	0916022-6	SMP	1	1	sample pCi/l		-6								
1	0916022-7	SMP	1	1	sample pCi/l		-7								
1	0916022-8	SMP	1	1	sample pCi/l		-8								
1	0916022-9	SMP	1	1	sample pCi/l		-9								
1	0916022-10	SMP	1	1	sample pCi/l		-10								
1	0916022-11	SMP	1	1	sample pCi/l		-11								
1	0916022-12	SMP	1	1	sample pCi/l		-12								
1	3H090422-2B01	MB	1	1	sample pCi/l		50-1								
1	3H090422-2B02	MB	1	1	sample pCi/l		-2								
1	3H090422-2B03	MB	1	1	sample pCi/l		-3								
1	3H090422-2B04	MB	1	1	sample pCi/l		-4								
1	3H090422-2B05	MB	1	1	sample pCi/l		-5								
1	3H090422-2B06	MB	1	1	sample pCi/l		-6								
1	3H090422-2B07	MB	1	1	sample pCi/l		-7								
1	3H090422-2B08	MB	1	1	sample pCi/l		-8								
1	3H090422-2B09	MB	1	1	sample pCi/l		-9								
1	3H090422-2B10	MB	1	1	sample pCi/l		-10								
1	3H090422-2B11	MB	1	1	sample pCi/l		-11								
1	3H090422-2B12	MB	1	1	sample pCi/l		-12								

Splice Solution Information					
Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date
S1	H-3	699.3020.95	88,189.122	DPM/ml	04/22/09
				0.1	ml
					RS-008

NA 10-28-09

Radiochemistry Instrument Worksheet

Prep Batch: 3H090422-2

ALS Laboratory Group -- FC

Reporting Units

LatID:	TestGrpName:	RptUnits:
0916022-1	H3	pCi/l
0916022-2	H3	pCi/l
0916022-3	H3	pCi/l
0916022-4	H3	pCi/l
0916022-5	H3	pCi/l
0916022-6	H3	pCi/l
0916022-7	H3	pCi/l
0916022-8	H3	pCi/l
0916022-9	H3	pCi/l
0916022-10	H3	pCi/l
0916022-11	H3	pCi/l
0916022-12	H3	pCi/l

Sample Barcodes

0916022-1 3H090422-2PFS1		0916022-2 3H090422-2PFS2	
0916022-3 3H090422-2PFS3		0916022-4 3H090422-2PFS4	
0916022-5 3H090422-2PFS5		0916022-6 3H090422-2PFS6	
0916022-7 3H090422-2PFS7		0916022-8 3H090422-2PFS8	
0916022-9 3H090422-2PFS9		0916022-10 3H090422-2PFS10	
0916022-11 3H090422-2PFS11		0916022-12 3H090422-2PFS12	
3H090422-2B01MB 3H090422-2PFS13		3H090422-2B02MB 3H090422-2PFS14	
3H090422-2B03MB 3H090422-2PFS15		3H090422-2B04MB 3H090422-2PFS16	
3H090422-2B05MB 3H090422-2PFS17		3H090422-2B06MB 3H090422-2PFS18	
3H090422-2B07MB 3H090422-2PFS19		3H090422-2B08MB 3H090422-2PFS20	
3H090422-2B09MB 3H090422-2PFS21		3H090422-2B10MB 3H090422-2PFS22	

Radiochemistry Instrument Worksheet

ALS Laboratory Group -- FC

Prep Batch: 3H090422-2

3H090422-2B11MB
3H090422-2PS23



3H090422-2B12MB
3H090422-2PS24



Radiochemistry Prep Worksheet

ALS Paragon

Prep Batch: 3H090422-2

Prep Procedure: H3

Reviewed By: DBC *ABC* Review Date: 4/23/2009

Non-Routine Pre-Treatment? Y / N Batch: *N/A* Re-Prep? Y / N Batch: *N/A* Prep QASS / NCR? Y / N

Prep SOP: PAI 700 Rev: 10 Prep Analyst: Derek B. Caduff Balance: Cocktail: UG LLT
 Prep SOP: NONE Prep Date: 4/22/2009 Balance: Cocktail Pipet: T-002
 Matrix Class: liquid Prep Dept: RS Aliquot Pipet: RS-015

Sample Num	Prep Num	LabID	QC Type	Dish No.	Init Alq sample	Fin Alq sample	Prep Basis	Analysis Vol.(ml)	Standards	Prep Notes
1	1	0916022-1	SMP		1	1	As Received		S1	
2	1	0916022-2	SMP		1	1	As Received		S1	0 U/L NITROMETHANE ADDED
3	1	0916022-3	SMP		1	1	As Received		S1	15
4	1	0916022-4	SMP		1	1	As Received		S1	30
5	1	0916022-5	SMP		1	1	As Received		S1	45
6	1	0916022-6	SMP		1	1	As Received		S1	60
7	1	0916022-7	SMP		1	1	As Received		S1	75
8	1	0916022-8	SMP		1	1	As Received		S1	90
9	1	0916022-9	SMP		1	1	As Received		S1	105
10	1	0916022-10	SMP		1	1	As Received		S1	120
11	1	0916022-11	SMP		1	1	As Received		S1	135
12	1	0916022-12	SMP		1	1	As Received		S1	150
13	1	3H090422-2B01	MB		1	1	As Received		S1	165
14	1	3H090422-2B02	MB		1	1	As Received		S1	0
15	1	3H090422-2B03	MB		1	1	As Received		S1	15
16	1	3H090422-2B04	MB		1	1	As Received		S1	30
17	1	3H090422-2B05	MB		1	1	As Received		S1	45
18	1	3H090422-2B06	MB		1	1	As Received		S1	60
19	1	3H090422-2B07	MB		1	1	As Received		S1	75
20	1	3H090422-2B08	MB		1	1	As Received		S1	90
21	1	3H090422-2B09	MB		1	1	As Received		S1	105
22	1	3H090422-2B10	MB		1	1	As Received		S1	120
23	1	3H090422-2B11	MB		1	1	As Received		S1	135
24	1	3H090422-2B12	MB		1	1	As Received		S1	150
										165

Radiochemistry Prep Worksheet

ALS Paragon

Prep Batch: 3H090422.2

Prep Procedure: H3
 Reviewed By: DBC *DK* Review Date: 4/23/2009

Non-Routine Pre-Treatment? Y / N Batch: *MA* Re-Prep? Y / N Batch: *MA* Prep GASS / NCR? Y / N
 Prep SOP: PAI 700 Rev: 10 Prep Analyst: Derek B. Caduff Balance: UG LLT
 Prep SOP: NONE Prep Date: 4/22/2009 Balance: Cocktall Pipet: T-002
 Matrix Class: liquid Prep Dept: RS Alquot Pipet: RS-015

Sampl Num	Prep Num	LabID	QC Type	Dish No.	Init Alq sample	Fin Alq sample	Prep Basis	Analysis Vol.(ml)	Standards	Prep Notes
<p>Comments UG LLT LOT #97-080401</p>										

Spiked By: Derek B. Caduff Date: 4/22/2009
 Witnessed By: Jeff Kujawa Date: 4/22/2009

Spike Solution Information						
Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date	Alquot Units
S1	H-3	699.3020.95	88,189.122	DPM/ml	04/22/09	0.1 ml
						RS-008

Radiochemistry Prep Worksheet

ALS Paragon

Prep Batch: 3H090422-2

Prep Procedure: H3

Prep Batch Not Validated!!!

Reviewed By:

Review Date:

Non-Routine Pre-Treatment? Y / N Batch:

Prep GASS / NCR? Y / N

Prep SOP: PAI 700 Rev: 10

Prep SOP: NONE

Matrix Class: liquid

Prep Analyst: Derek B. Caduff

Prep Date: 4/22/2009

Prep Dept: RS

Balance:

Balance:

Cocktail Pipet:

Cocktail Pipet:

Aliquot Pipet:

Samp Num	Prep Num	LabID	QC Type	Dish No.	Init Aliq sample	Fin Aliq sample	Prep Basis	Analysis Vol.(ml)	Standards	Prep Notes
1	1	0916022-1	SMP		1	1	As Received		S1	
2	1	0916022-2	SMP		1	1	As Received		S1	0 U/L NITROMETHANE ADDED
3	1	0916022-3	SMP		1	1	As Received		S1	15
4	1	0916022-4	SMP		1	1	As Received		S1	30
5	1	0916022-5	SMP		1	1	As Received		S1	45
6	1	0916022-6	SMP		1	1	As Received		S1	60
7	1	0916022-7	SMP		1	1	As Received		S1	75
8	1	0916022-8	SMP		1	1	As Received		S1	90
9	1	0916022-9	SMP		1	1	As Received		S1	105
10	1	0916022-10	SMP		1	1	As Received		S1	120
11	1	0916022-11	SMP		1	1	As Received		S1	135
12	1	0916022-12	SMP		1	1	As Received		S1	150
13	1	3H090422-2B01	MB		1	1	As Received		S1	165
14	1	3H090422-2B02	MB		1	1	As Received		S1	0
15	1	3H090422-2B03	MB		1	1	As Received		S1	15
16	1	3H090422-2B04	MB		1	1	As Received		S1	30
17	1	3H090422-2B05	MB		1	1	As Received		S1	45
18	1	3H090422-2B06	MB		1	1	As Received		S1	60
19	1	3H090422-2B07	MB		1	1	As Received		S1	75
20	1	3H090422-2B08	MB		1	1	As Received		S1	90
21	1	3H090422-2B09	MB		1	1	As Received		S1	105
22	1	3H090422-2B10	MB		1	1	As Received		S1	120
23	1	3H090422-2B11	MB		1	1	As Received		S1	136
24	1	3H090422-2B12	MB		1	1	As Received		S1	150
										165

Radiochemistry Prep Worksheet

ALS Paragon

Prep Batch: SH090422-2

Prep Batch Not Validated!!!

Prep Procedure: H3

Reviewed By:

Review Date:

Non-Routine Pre-Treatment? Y / N Batch: _____

Re-Prep? Y / N Batch: _____

Prep QASS / NCR? Y / N _____

Prep SOP: PAI 700 Rev: 10

Prep SOP: NONE

Matrix Class: liquid

Prep Analyst: Derek B. Caduff *DBC*

Prep Date: 4/22/2009

Prep Dept: RS

Balance:

Balance:

Cocktail:

Cocktail Pipet:

Aliquot Pipet:

QC Type	LabID	Dish No.	Init Alq sample	Fin Alq sample	Prep Basis	Analysis Vol.(ml)	Standards	Prep Notes

Comments

UG LIT LOT # 97-080401

Spiked By: *DBC*

Date: 4/22/09

Witnessed By: *[Signature]*

Date: 4/22/09

Splice Solution Information							
Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date	Aliquot Units	Pipet ID
S1	H-3	699.3020.95	88,189.122	DPM/ml	04/22/09	0.1	RS-008

Exp. 4/1/10

Radiochemistry Instrument Worksheet

ALS Laboratory Group -- FC

Prep Batch: 3H090422-3

Prep Procedure: H3 Glass Vial Swipes ICV's/ICB's

Analytical QASS / NOR? Y N *MA*

Prep Num	LabID	QC Type	Init Aliq	Fin Aliq	Units	Report Units	Cnt 1 File/Inst	Cnt 1 Rack-Pos	Cnt 1 Pos	Cnt 1 Clk By	Cnt 2 File/Inst	Cnt 2 Rack-Pos	Cnt 2 Pos	Cnt 2 Clk By	Cnt 3 File/Inst	Cnt 3 Rack-Pos	Cnt 3 Pos	Cnt 3 Clk By	Notes
1	0916023-1	SMP	1	1	sample	pCi/l	LS6000	25-2	nan										Added 10ul Nitro
1	0916023-2	SMP	1	1	sample	pCi/l		-3											60
1	0916023-3	SMP	1	1	sample	pCi/l		-4											140
1	3H090422-3B01	MB	1	1	sample	pCi/l		-6											10
1	3H090422-3B02	MB	1	1	sample	pCi/l		-7											60
1	3H090422-3B03	MB	1	1	sample	pCi/l		-8											140
1	3H090422-3CB1	MB	1	1	sample	pCi/l		-1											10
1	3H090422-3CB2	MB	1	1	sample	pCi/l		-5											10
1	3H090422-3CB3	MB	1	1	sample	pCi/l		-9											10

Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date	Aliquot	Units	Pipet ID
S1	H-3	648.3810.05	2,827.459	DPM/ml	04/22/09	0.1	ml	RS-008

Reporting Units

LabID	InstGrpName	RptUnits
0916023-1	H3	pCi/l
0916023-2	H3	pCi/l
0916023-3	H3	pCi/l

Sample Barcodes

0916023-1 3H090422-3PS1		0916023-2 3H090422-3PS2	
0916023-3 3H090422-3PS3		3H090422-3B01MB 3H090422-3PS4	
3H090422-3B02MB 3H090422-3PS5		3H090422-3B03MB 3H090422-3PS6	
3H090422-3CB1MB 3H090422-3PS7		3H090422-3CB2MB 3H090422-3PS8	
3H090422-3CB3MB 3H090422-3PS9			

Radiochemistry Prep Worksheet

Prep Batch: 3H090422-3

ALS Paragon

Prep Procedure: H13

Reviewed By: DBC

Review Date: 5/1/2009

Non-Routine Pre-Treatment? Y / N Batch: MA

Re-Prep? Y / N

Prep GASS / NCR? Y / N

Prep SOP: PAI 700 Rev: 10

Prep SOP: NONE

Matrix Class: liquid

Prep Analyst: Derek B. Caduff

Prep Date: 4/22/2009

Prep Dept: RS

Balance:

Balance:

Cocktail: UG LLT

Cocktail Pipet: T-002

Aliquot Pipet: RS-015

Sample Num	Prep Num	LabID	QC Type	Dish No.	Init Aliq sample	Fin Aliq sample	Prep Basis	Analysis Vol.(ml)	Standards	Prep Notes
1	1	0916023-1	SMP		1	1	As Received		S1	
2	1	0916023-2	SMP		1	1	As Received		S1	
3	1	0916023-3	SMP		1	1	As Received		S1	
4	1	3H090422-3B01	MB		1	1	As Received			MA
5	1	3H090422-3B02	MB		1	1	As Received			MA
6	1	3H090422-3B03	MB		1	1	As Received			MA
7	1	3H090422-3CB1	MB		1	1	As Received			MA
8	1	3H090422-3CB2	MB		1	1	As Received			MA
9	1	3H090422-3CB3	MB		1	1	As Received			MA

Comments

UG LLT LOT #97-080401

Spiked By: Derek B. Caduff Date: 4/22/2009

Witnessed By: Jeff Kujawa Date: 4/22/2009

Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date	Aliquot Units	Pipet ID
S1	H-3	648.3610.05	2.827.459	DPW/ml	04/22/09	0.1 ml	RS-008

Radiochemistry Prep Worksheet

ALS Paragon

Prep Batch: 3H090422-3

Prep Batch Not Validated!!!

Prep Procedure: H3

Reviewed By:

Review Date:

Non-Routine Pre-Treatment? Y / N Batch:

Prep GASS / NCR? Y / N

Prep SOP: PAI 700 Rev: 10

Prep Analyst: Derek B. Caduff

Cocktail: UG LLT

Prep SOP: NONE

Prep Date: 4/22/2009

Cocktail Pipet: T-002

Matrix Class: liquid

Prep Dept: RS

Alliquot Pipet: RS-015

Stamp Num	Prep Num	LabID	QC Type	Dish No.	Init Alq sample	Fin Alq sample	Prep Basis	Analysis Vol.(ml)	Standards	Prep Notes
1	1	0916023-1	SMP	1	1	1	As Received		S1	
2	1	0916023-2	SMP	1	1	1	As Received		S1	
3	1	0916023-3	SMP	1	1	1	As Received		S1	
4	1	3H090422-3B01	MB	1	1	1	As Received			
5	1	3H090422-3B02	MB	1	1	1	As Received			
6	1	3H090422-3B03	LCSD	1	1	1	As Received			

Comments

UG LLT LOT # 97-050401

Spiked By: Derek B. Caduff *ops* Date: 4/22/09

Witnessed By: *[Signature]* Date: 4/22/09

Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date	Alliquot Units	Pipet ID
S1	H-3	648.3810.05	2,927.459	DPM/ml	04/22/09	0.1	RS-008

Exp. 4/31/09

Prepare a working dilution ~ 99000 dpm/ml of ³H from 699.3020.18.

1) Determine the density of DI water Bal. #

Mass of empty 100ml volumetric flask: 66.4317g 12

Mass of flask and 100 ml DI water: 106.0659g ✓

Net mass of DI water: 996.335

$\rho = 0.99633 \text{ g/ml}$

2) Transfer standard.

Mass of empty bottle without lid: 97.1195g 12

Mass of bottle and standard: 100.5266g ✓

Net mass of standard: 3.4071g

3) Dilute with DI water.

Mass of empty bottle without lid: 97.1195g 12

Mass of bottle, standard, and DI water: 197.7444g ✓

Net mass of standard and DI water: 100.6249g

4) Final activity calculation:

$$\frac{(4.733,412,57 \text{ dpm/ml})(3.4071 \text{ g})(0.9963 \text{ g/ml})}{(100.6249 \text{ g})(0.9963 \text{ g/ml})} = 15.967757 \text{ dpm/ml}$$

⁰⁰⁰ 7/13/08

160,351,042 dpm/ml

To ensure activity correct for 699.3020.18.

Strid ID: 699.3020.95

MC 7/13/08

Description: H-3

Expiration: 2/26/2009

Activity: 160351.04 dpm/mL

Uncertainty: 1154.53 dpm/mL

Ref. Date: 9/3/1998

Ref Time: N/A

Prep Date: 2/7/2008 Prep by: DC

Matrix/Comp. DI WATER

Half Life (y): 1.23E+01

Reverification Log		
Analysis Date	Initials	Expiration Date
4/1/09	MC	4/1/10

MC 7/13/08

Continued on Page

[Signature]

Signed

7/13/08

Date

Read and Understood By

[Signature]

Signed

7/13/08

Date

Prepare a 10 dilution of Ampoule 699 by diluting with DI water.

1) Determine the density of DI water

Mass of empty Class A ^{100ml} volumetric flask	67.7956 g Bal #12
Mass of flask + 100 ml of DI water	167.3755 g ↓
Net mass of DI water	99.5799 g

P.F. = 0.9958 g/mL

2) Transfer contents of ampoule 699 into a 40 mL amber glass VOA vial.

Mass of VOA vial w/o lid	24.3883g Bal #12
Mass of opened ampoule before transfer	40.8432g ↓
Mass of opened ampoule after transfer	36.0281g ↓
Net mass of Std transferred	4.8151g

3) Add DI water to final dilution

Mass of VOA vial ^{w/ lid} from above	24.3883g Bal #12
Mass of VOA vial std + DI water	62.9648g ↓
Net mass of Std + DI water	38.5765g

4) Final Activity Calculation

$$(634.17 \text{ KBq/g}) (4.8151 \text{ g}) \left(\frac{100 \text{ ppm}}{1000 \text{ g/kg}} \right) \left(\frac{1000 \text{ Bq}}{1 \text{ KBq}} \right) (0.9958 \text{ g/mL})$$

38.5765g

= 4.73341257 $\frac{\text{Bq}}{\text{mL}}$

U.S. Department of Commerce
National Institute of Standards
and Technology
SRM 4927F
Hydrogen-3
~4 MBq in distilled water

CAUTION
RADIOACTIVE



Continued on Page

Chad W. [Signature]
Signed

7/28/03
Date

Read and Understood By
[Signature]
Signed

7/31/03
Date



National Institute of Standards & Technology

Certificate

PA.F ED 08/99
REC'D 5-09-03

Standard Reference Material 4927F Hydrogen-3 Radioactivity Standard

This Standard Reference Material (SRM) consists of radioactive hydrogen-3, as water, in 5 mL of distilled water. The solution is contained in a flame-sealed NIST borosilicate-glass ampoule. The SRM is intended for the calibration of beta-particle counting instruments and for the monitoring of radiochemical procedures.

Radiological Hazard

The SRM ampoule contains hydrogen-3 with a total activity of approximately 3.2 MBq. Hydrogen-3 decays by beta-particle emission. None of the beta particles escape from the SRM ampoule. During the decay process no photons are emitted. Approximate unshielded dose rates at several distances (as of the reference time) are given in note [a]*. There is no detectable external radiation. The SRM should be used only by persons qualified to handle radioactive material.

Chemical Hazard

The SRM ampoule contains only distilled water. There is no chemical hazard. If the ampoule is to be opened to transfer the solution, the recommended procedure is given on page 2.

Storage and Handling

The SRM should be stored and used at a temperature between 5 and 65 °C. The solution in an unopened ampoule should remain stable and homogeneous until at least September 2008.

The ampoule (or any subsequent container) should always be clearly marked as containing radioactive material. If the ampoule is transported it should be packed, marked, labeled, and shipped in accordance with the applicable national, international, and carrier regulations. The solution in the ampoule is a dangerous good (hazardous material) because of the radioactivity.

Preparation

This Standard Reference Material was prepared in the Physics Laboratory, Ionizing Radiation Division, Radioactivity Group, L.R. Karam, Group Leader. The overall technical direction and physical measurements leading to certification were provided by L.L. Lucas and M.P. Unterwieser of the Radioactivity Group.

The support aspects involved in the preparation, certification, and issuance of this SRM were coordinated through the Standard Reference Materials Program by J.W.L. Thomas.

Bert M. Coursey, Chief
Ionizing Radiation Division

Nancy M. Trahey, Chief
Standard Reference Materials Program

Gaithersburg, Maryland 20899
June 1999
Half-life and text revised October 2000

Recommended Procedure for Opening the SRM Ampoule

- 1) If the SRM solution is to be diluted, it is recommended that the diluting solution have a composition comparable to that of the SRM solution.
- 2) Wear eye protection, gloves, and protective clothing and work over a tray with absorbent paper in it. Work in a fume hood.
- 3) Shake the ampoule to wet all of the inside surface of the ampoule. Return the ampoule to the upright position.
- 4) Check that all of the liquid has drained out of the neck of the ampoule. If necessary, gently tap the neck to speed the process.
- 5) Holding the ampoule upright, score the narrowest part of the neck with a scribe or diamond pencil.
- 6) Lightly wet the scored line. This reduces the crack propagation velocity and makes for a cleaner break.
- 7) Hold the ampoule upright with a paper towel, a wiper, or a support jig. Position the scored line away from you. Using a paper towel or wiper to avoid contamination, snap off the top of the ampoule by pressing the narrowest part of the neck away from you while pulling the tip of the ampoule towards you.
- 8) Transfer the solution from the ampoule using a pycnometer or a pipet with dispenser handle. **NEVER PIPETTE BY MOUTH.**
- 9) Seal any unused SRM solution in a flame-sealed glass ampoule, if possible, to minimize the evaporation loss.

See also reference [4]*.

PROPERTIES OF SRM 4927F

Certified values

Solution density	$(0.998 \pm 0.002) \text{ g}\cdot\text{mL}^{-1}$ at 20.0 °C [b]*
Radionuclide	Hydrogen-3
Reference time	1200 EST, 3 September 1998
Massic activity of the solution [c]	$634.7 \text{ kBq}\cdot\text{g}^{-1}$
Relative expanded uncertainty ($k=2$)	0.72% [d] [e]

Uncertified values

Physical Properties:			
Source description	Liquid in flame-sealed NIST borosilicate-glass ampoule		
Ampoule specifications	Body outside diameter	$(16.5 \pm 0.5) \text{ mm}$	
	Wall thickness	$(0.60 \pm 0.04) \text{ mm}$	
	Barium content	Less than 2.5%	
	Lead-oxide content	Less than 0.02%	
	Other heavy elements	Trace quantities	
Solution mass	Approximately 5.0 g		
Chemical Properties:			
Solution composition	Chemical Formula	Concentration ($\text{mol}\cdot\text{L}^{-1}$)	Mass Fraction ($\text{g}\cdot\text{g}^{-1}$)
	H_2O ^3HEO	55 6×10^{-7}	1.00 1×10^{-8}
Radiological Properties:			
Radionuclidic impurities	None detected [f]		
Half lives used	Hydrogen-3: $(4500 \pm 8) \text{ d}$ [g]		
Calibration method and measuring instrument(s)	4 $\pi\beta$ gas counting of SRM 4927E using the NIST length-compensated internal gas proportional counters and intercomparison of SRMs 4927E/4927F using two 4 $\pi\beta$ liquid-scintillation counting systems [h]		

EVALUATION OF THE UNCERTAINTY OF THE MASSIC ACTIVITY [d]*

Input Quantity x_i , the source of uncertainty (and individual uncertainty components where appropriate)	Method Used To Evaluate $u(x_i)$, the standard uncertainty of x_i (A) denotes evaluation by statistical methods (B) denotes evaluation by other methods	Relative Uncertainty Of Input Quantity, $u(x_i)/x_i$, (%) [H]	Relative Sensitivity Factor, $ \partial y/\partial x_i \cdot$ (x_i/y) [I]	Relative Uncertainty Of Output Quantity, $u_i(y)/y$, (%) [K]
Massic count rate of SRM 4927E, corrected for background and decay [h]	Standard deviation of the mean for 23 sets of gas counting measurements (A)	0.18	1.0	0.18
Gram-mole measurements	Estimated (B)	0.20	1.0	0.20
Live-time [p]	Estimated (B)	0.10	1.0	0.10
Extrapolation of count-rate-versus-energy to zero energy	Estimated (B)	0.20	1.0	0.20
Half life of H-3	Standard uncertainty of the half life (A)	0.18 [m]	0.009 [n]	0.002
Liquid-scintillation intercomparison of SRM 4927F and SRM 4927E	Standard deviation of the mean for 7 sets of liquid-scintillation measurements (A)	0.06	1.0	0.06
Radionuclidic impurities	Limit of detection (B) [q]	100.	0.0005	0.05
Relative Combined Standard Uncertainty of the Output Quantity, $u_c(y)/y$, (%)				0.36
Coverage Factor, k				<u>2</u>
Relative Expanded Uncertainty of the Output Quantity, $U(y)/y$, (%)				0.72

NOTES

- [a] The Sievert is the SI unit for dose equivalent. See reference [1]. One μSv is equal to 0.1 mrem.
Distance from Ampoule (cm): 1 30 100
Approximate Dose Rate ($\mu\text{Sv/h}$): <0.1 (Not detectable)
- [b] The stated uncertainty is two times the standard uncertainty.
- [c] Massic activity is the preferred name for the quantity activity divided by the total mass of the sample. See reference [1].
- [d] The reported value, y , of massic activity (activity per unit mass) at the reference time was not measured directly but was derived from measurements and calculations of other quantities. This can be expressed as $y = f(x_1, x_2, x_3, \dots, x_n)$, where f is a mathematical function derived from the assumed model of the measurement process.
- The value, x_i , used for each input quantity i has a standard uncertainty, $u(x_i)$, that generates a corresponding uncertainty in y , $u_i(y) = |\partial f / \partial x_i| \cdot u(x_i)$, called a component of combined standard uncertainty of y .
- The combined standard uncertainty of y , $u_c(y)$, is the positive square root of the sum of the squares of the components of combined standard uncertainty.
- The combined standard uncertainty is multiplied by a coverage factor of $k = 2$ to obtain U , the expanded uncertainty of y .
- Since it can be assumed that the possible estimated values of the massic activity are approximately normally distributed with approximate standard deviation $u_c(y)$, the unknown value of the massic activity is believed to lie in the interval $y \pm U$ with a level of confidence of approximately 95 percent.
- For further information on the expression of uncertainties, see references [2] and [3].
- [e] The value of each standard uncertainty component, and hence the value of the expanded uncertainty itself, is a best estimate based upon all available information, but is only approximately known. That is to say, the "uncertainty of the uncertainty" is large and not well known. This is true for uncertainties evaluated by statistical methods (e.g., the relative standard deviation of the standard deviation of the mean for the massic response is approximately 50%) and for uncertainties evaluated by other methods (which could easily be over estimated or under estimated by substantial amounts). The unknown value of the expanded uncertainty is believed to lie in the interval $U/2$ to $2U$ (i.e., within a factor of 2 of the estimated value).
- [f] The estimated limit of detection for radionuclidic impurities is $300 \text{ Bq} \cdot \text{g}^{-1}$.
- [g] The stated uncertainty is the standard uncertainty. See reference [5].
- [h] Extensive gas-counting measurements were made on the SRM 4927E solution during 1998 and 1999. The SRM 4927F solution was intercompared with the SRM 4927E solution using liquid-scintillation counting.
- [i] Relative standard uncertainty of the input quantity x_i .

- [j] The relative change in the output quantity y divided by the relative change in the input quantity x_i . If $|\partial y/\partial x_i| \cdot (x_i/y) = 1.0$, then a 1% change in x_i results in a 1% change in y . If $|\partial y/\partial x_i| \cdot (x_i/y) = 0.05$, then a 1% change in x_i results in a 0.05% change in y .
- [k] Relative component of combined standard uncertainty of output quantity y , rounded to two significant figures or less. The relative component of combined standard uncertainty of y is given by $u_i(y)/y = |\partial y/\partial x_i| \cdot u(x_i)/y = |\partial y/\partial x_i| \cdot (x_i/y) \cdot u(x_i)/x_i$. The numerical values of $u(x_i)/x_i$, $|\partial y/\partial x_i| \cdot (x_i/y)$, and $u_i(y)/y$, all dimensionless quantities, are listed in columns 3, 4, and 5, respectively. Thus, the value in column 5 is equal to the value in column 4 multiplied by the value in column 3. The input quantities are independent, or very nearly so. Hence the covariances are zero or negligible.
- [m] The relative standard uncertainty of $\lambda \cdot t$ is determined by the relative standard uncertainty of λ (i.e., of the half life). The relative standard uncertainty of t is negligible.
- [n] $|\partial y/\partial x_i| \cdot (x_i/y) = |\lambda \cdot t|$
- [p] The live time is determined by counting the pulses from a gated crystal-controlled oscillator.
- [q] The standard uncertainty for each undetected impurity that might reasonably be expected to be present is estimated to be equal to the estimated limit of detection for that impurity, i.e. $u(x_i)/x_i = 100\%$. $|\partial y/\partial x_i| \cdot (x_i/y) = \{(\text{response per Bq of impurity})/(\text{response per Bq of H-3})\} \cdot \{(\text{Bq of impurity})/(\text{Bq of H-3})\}$. Thus $u_i(y)/y$ is the relative change in y if the impurity were present with a massic activity equal to the estimated limit of detection.

REFERENCES

- [1] International Organization for Standardization (ISO), *ISO Standards Handbook - Quantities and Units*, 1993. Available from the American National Standards Institute, 11 West 42nd Street, New York, NY 10036, U.S.A. 1-212-642-4900.
- [2] International Organization for Standardization (ISO), *Guide to the Expression of Uncertainty in Measurement*, 1993. Available from the American National Standards Institute, 11 West 42nd Street, New York, NY 10036, U.S.A. 1-212-642-4900. (Listed under ISO miscellaneous publications as "ISO Guide to the Expression 1993".)
- [3] B. N. Taylor and C. R. Kuyatt, *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*, NIST Technical Note 1297, 1994. Available from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20407, U.S.A.
- [4] National Council on Radiation Protection and Measurements Report No. 58, *A Handbook of Radioactivity Measurements Procedures*, Second Edition, 1985. Available from the National Council on Radiation Protection and Measurements, 7910 Woodmont Avenue, Bethesda, MD 20814 U.S.A.
- [5] L.L. Lucas and M.P. Unterwieser, *Comprehensive Review and Critical Evaluation of the Half-Life of Tritium*, J. Res. Natl. Inst. Stand. Technol. 105, 541-549 (2000).

Prepare a working dilution of ~ 4,000 dpm/ml of #1,
648.2382.75.

1) Determine density of DI water bulk
 Mass of 100 ml vol. flask 68.2976g 12
 Mass of flask and 100 ml DI H₂O. 168.0291g 12
 Net Mass of water 99.7316g
 $\rho = .9973 \frac{g}{ml}$

2) Transfer Std.
 Mass of empty Amber jar (No lid) 257.75g 26
 Mass of Jar + Std transferred 264.58g 12
 * Net Mass of Std. Transferred 6.8506g 12
 ↳ from below JPO 4/28/08

4) Dilute to final vol. w/ DI water
 Mass of Std., H₂O, Jar 750.9g 26
 Mass of Jar from above 257.75g
 Net Mass of new dilution 493.15g

* Standard was transferred by difference from a plastic cup
 Mass of Cup and ~ 10g of Standard 23.6533g 12
 Mass of Cup and Std. [Not] Transferred 16.8027g 12



Final Activity Calculation

$$\frac{(369,530.45 \frac{dpm}{ml}) (6.8506g) (.9973 \frac{g}{ml})}{(.9958 \frac{g}{ml}) (493.15g)} = 5141.07 \frac{dpm}{ml}$$

Std ID: 648.3610.05

Description: H-3
 Expiration: 4/30/2009
 Activity: 5141.07 dpm/mL
 2s Uncertainty: 37.02 dpm/mL
 Ref. Date: 9/3/1998
 Ref Time: N/A
 Prep Date: 4/30/2008 Prep by: JD
 Matrix/Comp. DI WATER
 Half Life (y): 1.23E+01

Reverification Log		
Analysis Date	Initials	Expiration Date
4/11/09	JDS	4/11/10

 4/28/08 MC 7/13/08  7/13/08
 Signed Date Signed Date

Prepare a ~~working~~ ^{10 dilution} dilution (of approximately 250 ~~ppm~~) with standard using RSD #644B and diluting with DI water.

1) Determine the density of DI water:

Mass of empty 100mL class A volumetric flask	67.1522 g	Bal #13
Mass of flask + water	166.7351 g	↓
Net mass of 100mL H ₂ O	99.5829 g	
	$\rho = .9958 \text{ g/mL}$	

2) Transfer contents of Ampoule SRM 4927F to a 500 mL glass amber bottle

Mass of 500 mL glass amber bottle w/o lid	255.04 g	Bal #26
Mass of open ampoule + 50mL beaker	210.9075 g	Bal #12
Mass of empty ampoule + 50mL beaker	36.0149 g	↓
Net mass of SRM	4.8927 g	

3) Dilute SRM with DI water

Mass of bottle w/o lid	255.04 g	Bal #26
Mass of bottle, stop + DI water	757.1 g	↓
Net mass of SRM + DI water	502.1 g	

4. Final Activity Calculation

$$\left(634.7 \frac{\text{KBq}}{\text{g}} \right) \left(1000 \frac{\text{Bq}}{\text{KBq}} \right) \left(100 \frac{\text{DPM}}{\text{Bq}} \right) \left(4.8927 \text{ g} \right) \left(.9958 \frac{\text{g}}{\text{mL}} \right)$$

$$502.1 \text{ g} = 369,530.45 \frac{\text{DPM}}{\text{mL}}$$

U.S. Department of Commerce
National Institute of Standards
and Technology
SRM 4927F
Hydrogen-3

<4 MBq in distilled water

CAUTION
RADIOACTIVE



Continued on Page

Read and Understood By

Chad [Signature]
Signed

3/27/03
Date

Renee [Signature]
Signed

3/27/03
Date



National Institute of Standards & Technology

Certificate

PA ID 0648
17-01-02

Standard Reference Material 4927F

Hydrogen-3 Radioactivity Standard

This Standard Reference Material (SRM) consists of radioactive hydrogen-3, as water, in 5 mL of distilled water. The solution is contained in a flame-sealed NIST borosilicate-glass ampoule. The SRM is intended for the calibration of beta-particle counting instruments and for the monitoring of radiochemical procedures.

Radiological Hazard

The SRM ampoule contains hydrogen-3 with a total activity of approximately 3.2 MBq. Hydrogen-3 decays by beta-particle emission. None of the beta particles escape from the SRM ampoule. During the decay process no photons are emitted. Approximate unshielded dose rates at several distances (as of the reference time) are given in note [a]*. There is no detectable external radiation. The SRM should be used only by persons qualified to handle radioactive material.

Chemical Hazard

The SRM ampoule contains only distilled water. There is no chemical hazard. If the ampoule is to be opened to transfer the solution, the recommended procedure is given on page 2.

Storage and Handling

The SRM should be stored and used at a temperature between 5° and 65 °C. The solution in an unopened ampoule should remain stable and homogeneous until at least September 2008.

The ampoule (or any subsequent container) should always be clearly marked as containing radioactive material. If the ampoule is transported it should be packed, marked, labeled, and shipped in accordance with the applicable national, international, and carrier regulations. The solution in the ampoule is a dangerous good (hazardous material) because of the radioactivity.

Preparation

This Standard Reference Material was prepared in the Physics Laboratory, Ionizing Radiation Division, Radioactivity Group, L.R. Karam, Group Leader. The overall technical direction and physical measurements leading to certification were provided by L.L. Lucas and M.P. Unterwieser of the Radioactivity Group.

The support aspects involved in the preparation, certification, and issuance of this SRM were coordinated through the Standard Reference Materials Program by J.W.L. Thomas.

Bert M. Coursey, Chief
Ionizing Radiation Division

Nancy M. Trahey, Chief
Standard Reference Materials Program

Gaithersburg, Maryland 20899
June 1999
Half-life and text revised October 2000

Recommended Procedure for Opening the SRM Ampoule

- 1) If the SRM solution is to be diluted, it is recommended that the diluting solution have a composition comparable to that of the SRM solution.
- 2) Wear eye protection, gloves, and protective clothing and work over a tray with absorbent paper in it. Work in a fume hood.
- 3) Shake the ampoule to wet all of the inside surface of the ampoule. Return the ampoule to the upright position.
- 4) Check that all of the liquid has drained out of the neck of the ampoule. If necessary, gently tap the neck to speed the process.
- 5) Holding the ampoule upright, score the narrowest part of the neck with a scribe or diamond pencil.
- 6) Lightly wet the scored line. This reduces the crack propagation velocity and makes for a cleaner break.
- 7) Hold the ampoule upright with a paper towel, a wiper, or a support jig. Position the scored line away from you. Using a paper towel or wiper to avoid contamination, snap off the top of the ampoule by pressing the narrowest part of the neck away from you while pulling the tip of the ampoule towards you.
- 8) Transfer the solution from the ampoule using a pycnometer or a pipet with dispenser handle. NEVER PIPETTE BY MOUTH.
- 9) Seal any unused SRM solution in a flame-sealed glass ampoule, if possible, to minimize the evaporation loss.

See also reference [4]*.

PROPERTIES OF SRM 4927F

Certified values: [b]

Solution density	$(0.998 \pm 0.002) \text{ g} \cdot \text{mL}^{-1}$ at 20.0 °C [b]*
Radionuclide	Hydrogen-3
Reference time	1200 EST, 3 September 1998
Massic activity of the solution [c]	$634.7 \text{ kBq} \cdot \text{g}^{-1}$
Relative expanded uncertainty ($k=2$)	0.72% [d] [e]

Uncertified values

Physical Properties:			
Source description	Liquid in flame-sealed NIST borosilicate-glass ampoule		
Ampoule specifications	Body outside diameter	$(16.5 \pm 0.5) \text{ mm}$	
	Wall thickness	$(0.60 \pm 0.04) \text{ mm}$	
	Barium content	Less than 2.5%	
	Lead-oxide content	Less than 0.02%	
	Other heavy elements	Trace quantities	
Solution mass	Approximately 5.0 g		
Chemical Properties:			
Solution composition	Chemical Formula	Concentration (mol·L ⁻¹)	Mass Fraction (g·g ⁻¹)
	H ₂ O HHO	55 6×10^{-7}	~1800 1×10^{-3}
Radiological Properties:			
Radiometric impurities	None detected [f]		
Half lives used	Hydrogen-3: $(4500 \pm 8) \text{ d}$ [g]		
Calibration method and measuring instrument(s)	4πβ gas counting of SRM 4927E using the NIST length-compensated internal gas proportional counters and intercomparison of SRMs 4927E/4927F using two 4πβ liquid-scintillation counting systems [h]		

EVALUATION OF THE UNCERTAINTY OF THE MASSIC ACTIVITY [q]*

Input Quantity x_i , the source of uncertainty (and individual uncertainty components where appropriate)	Method Used To Evaluate $u(x_i)$, the standard uncertainty of x_i . (A) denotes evaluation by statistical methods. (B) denotes evaluation by other methods	Relative Uncertainty Of Input Quantity, $u(x_i)/x_i$, (%) [i]	Relative Sensitivity Factor, $ dy/dx_i $ (x_i/y) [j]	Relative Uncertainty Of Output Quantity, $u_i(y)/y$, (%) [k]
Massic count rate of SRM 4927E, corrected for background and decay [h]	Standard deviation of the mean for 23 sets of gas counting measurements (A)	0.18	1.0	0.18
Gram-mole measurements	Estimated (B)	0.20	1.0	0.20
Live-time [p]	Estimated (B)	0.10	1.0	0.10
Extrapolation of count-rate-versus-energy to zero energy [r]	Estimated (B)	0.20	1.0	0.20
Half life of H-3	Standard uncertainty of the half life (A)	0.18 [m]	0.009 [n]	0.002
Liquid-scintillation intercomparison of SRM 4927E and SRM-4927E	Standard deviation of the mean for 7 sets of liquid-scintillation measurements (A)	0.06	1.0	0.06
Radiochemical impurities	Limit of detection (B) [q]	100	0.0005	0.05
Relative Combined Standard Uncertainty of the Output Quantity, $u_c(y)/y$, (%)				0.36
Coverage Factor, k				2
Relative Expanded Uncertainty of the Output Quantity, $U(y)/y$, (%)				0.72

NOTES

- [a] The Sievert is the SI unit for dose equivalent. See reference [1]. One μSv is equal to 0.1 mrem.
 Distance from Ampoule (cm): 1 30 100
 Approximate Dose Rate ($\mu\text{Sv/h}$): <0.1 (Not detectable)
- [b] The stated uncertainty is two times the standard uncertainty.
- [c] Massic activity is the preferred name for the quantity activity divided by the total mass of the sample. See reference [1].
- [d] The reported value, y , of massic activity (activity per unit mass) at the reference time was not measured directly but was derived from measurements and calculations of other quantities. This can be expressed as $y = f(x_1, x_2, x_3, \dots, x_n)$, where f is a mathematical function derived from the assumed model of the measurement process.
- The value, x_i , used for each input quantity x_i has a standard uncertainty, $u(x_i)$, that generates a corresponding uncertainty in y , $u_i(y) = |dy/dx_i| \cdot u(x_i)$, called a component of combined standard uncertainty of y .
- The combined standard uncertainty of y , $u_c(y)$, is the positive square root of the sum of the squares of the components of combined standard uncertainty.
- The combined standard uncertainty is multiplied by a coverage factor of $k = 2$ to obtain U , the expanded uncertainty of y .
- Since it can be assumed that the possible estimated values of the massic activity are approximately normally distributed with approximate standard deviation $u_c(y)$, the unknown value of the massic activity is believed to lie in the interval $y \pm U$ with a level of confidence of approximately 95 percent.
- For further information on the expression of uncertainties, see references [2] and [3].
- [e] The value of each standard uncertainty component, and hence the value of the expanded uncertainty itself, is a best estimate based upon all available information, but is only approximately known. That is to say, the "uncertainty of the uncertainty" is large and not well known. This is true for uncertainties evaluated by statistical methods (e.g., the relative standard deviation of the standard deviation of the mean for the massic response is approximately 50%) and for uncertainties evaluated by other methods (which could easily be over estimated or under estimated by substantial amounts). The unknown value of the expanded uncertainty is believed to lie in the interval $U/2$ to $2U$ (i.e., within a factor of 2 of the estimated value).
- [f] The estimated limit of detection for radionuclides impurities is $300 \text{ Bq} \cdot \text{g}^{-1}$.
- [g] The stated uncertainty is the standard uncertainty. See reference [5].
- [h] Extensive gas counting measurements were made on the SRM 4927E solution during 1998 and 1999. The SRM 4927F solution was intercompared with the SRM 4927E solution using liquid-scintillation counting.
- [i] Relative standard uncertainty of the input quantity x_i .

- [f] The relative change in the output quantity y divided by the relative change in the input quantity x_i . If $|\partial y/\partial x_i| \cdot (x_i/y) = 1.0$, then a 1% change in x_i results in a 1% change in y . If $|\partial y/\partial x_i| \cdot (x_i/y) = 0.05$, then a 1% change in x_i results in a 0.05% change in y .
- [k] Relative component of combined standard uncertainty of output quantity, rounded to two significant figures or less. The relative component of combined standard uncertainty of y is given by $u(y)/y = |\partial y/\partial x_i| \cdot u(x_i)/y = |\partial y/\partial x_i| \cdot (x_i/y) \cdot u(x_i)/x_i$. The numerical values of $u(x_i)/x_i$, $|\partial y/\partial x_i| \cdot (x_i/y)$, and $u(y)/y$, all dimensionless quantities, are listed in columns 3, 4, and 5, respectively. Thus, the value in column 5 is equal to the value in column 4 multiplied by the value in column 3. The input quantities are independent, or very nearly so. Hence the covariances are zero or negligible.
- [m] The relative standard uncertainty of $\lambda \cdot t$ is determined by the relative standard uncertainty of λ (i.e., of the half life). The relative standard uncertainty of t is negligible.
- [n] $|\partial y/\partial x_i| \cdot (x_i/y) = |\lambda \cdot t|$
- [p] The live time is determined by counting the pulses from a gated crystal-controlled oscillator.
- [q] The standard uncertainty for each undetected impurity that might reasonably be expected to be present is estimated to be equal to the estimated limit of detection for that impurity, i.e. $u(x_i)/x_i = 100\%$. $|\partial y/\partial x_i| \cdot (x_i/y) = \{(\text{response per Bq of impurity})/(\text{response per Bq of H-3})\} \cdot \{(\text{Bq of impurity})/(\text{Bq of H-3})\}$. Thus $u(y)/y$ is the relative change in y if the impurity were present with a massic activity equal to the estimated limit of detection.

REFERENCES

- [1] International Organization for Standardization (ISO), *ISO Standards Handbook - Quantities and Units*, 1993. Available from the American National Standards Institute, 11 West 42nd Street, New York, NY 10036, U.S.A. 1-212-642-4900.
- [2] International Organization for Standardization (ISO), *Guide to the Expression of Uncertainty in Measurement*, 1993. Available from the American National Standards Institute, 11 West 42nd Street, New York, NY 10036, U.S.A. 1-212-642-4900. (Listed under ISO miscellaneous publications as "ISO Guide to the Expression 1993".)
- [3] B. N. Taylor and C. E. Kuyatt, *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*, NIST Technical Note 1297, 1994. Available from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20407, U.S.A.
- [4] National Council on Radiation Protection and Measurements Report No. 58, *A Handbook of Radioactivity Measurements Procedures*, Second Edition, 1985. Available from the National Council on Radiation Protection and Measurements, 7910 Woodmont Avenue, Bethesda, MD 20814 U.S.A.
- [5] L.L. Lucas and M.P. Unterwiesing, *Comprehensive Review and Critical Evaluation of the Half-Life of Tritium*, J. Res. Natl. Inst. Stand. Technol. 105, 541-549 (2000).

LS6000 H-3 Swipes Background Determination

Interim control limits are established from the initial calibration for the geometry of interest. Limits are +/- 3 standard deviations from the initial unquenched calibration blank data. Once enough historical data is acquired, new historical limits are set as follows: Control limits for reagent blanks are established from 30 individual historical data points (10 batches). Limits are +/- 3 standard deviations from 30 individual historical data points. Individual reagent blanks and the average of reagent blanks from each batch are in control if the Count Rate (CPM) is within the established control limits.

CURRENTLY UNDER INTERIM LIMITS!

COUNT DATE	#	Sample ID	Count Duration (m)	Count Rate (CPM)	Updated 10/29/09 mh			Average of Reagent Blanks			
					Total Cts.	Mean	LCL	UCL	Pass ?	LCL	UCL
10/27/2009	1	3H090422-3CB1	180	7.16	1288.8	3.49	8.24	PASS			
10/27/2009	2	3H090422-3CB2	180	7.38	1328.4	3.49	8.24	PASS			
10/28/2009	3	3H090422-3CB3	180	7.07	1272.6	3.49	8.24	PASS	3.49	8.24	PASS

H-3 Swipes "Window 2" Control Limits (LS 6000)

The background count rate is determined from the average of the reagent blanks for the batch.

Window 2 control limits are established using the average count rate from the three reagent blanks associated with each prep batch +/- 3X the estimated poisson uncertainty.

Updated 10/29/09 mh

COUNT DATE	#	Sample ID	Count Duration (min.)	Average count Duration (min.)	Count Rate (CPM)	Batch Average Reagent Blank	Lower Control Limit	Upper Control Limit
10/27/2009	1	3H090422-3CB1	180		29.98			
10/27/2009	2	3H090422-3CB2	180		28.89			
10/28/2009	3	3H090422-3CB3	180	180	30.01	29.63	28.41	30.84

DAILY INSTRUMENT PERFORMANCE CHECKS - LS6000 (LL OFF, LUMEX OFF)

Daily IPCs consist of the following standards;

Efficiency Check -

Beckman Tritium Standard

Beckman C-14 Standard

Lot HNZ0202

Lot CNZ3112

101900.00 dpm

98500.0 dpm

2/17/2005 REF

2/17/2005 REF

2/17/2010 EXP

2/17/2010 EXP

INSTRUMENT RE-CALIBRATED FOR ALL TESTS STARTING 04/08/09. mbc

Historical Control Limits

as of 06/12/09 MH

Decay Corrected Tritium

Carbon-14

UCL 70748.54

80439.21

Mean Value 67379.56

76608.77

LCL 64010.58

72778.33

Decay Corrected

Obs	Date	H-3 CPM	H-CPM	PASS?	C-14 CPM	PASS?
122	10/26/2009	51452.50	67035.95	OK	76690.2	OK
123	10/27/2009	51352.20	66915.61	OK	76616.2	OK
124	10/28/2009	51270.40	66819.35	OK	76543.7	OK

DAILY CHECK LL ON ⁹⁹Tc SOURCE- LS6000

⁹⁹ Tc standard	SPIKE
836.3020.70	KNOWN ACTIVITY
7/25/2008 REF	58000.38 dpm/g
7/25/2009 EXP	58000.38 dpm

INSTRUMENT RE-CALIBRATED FOR ALL TESTS STARTING 04/08/09. mbc

Historical Control Limits 6/12/2009

	blank	Blank Quench #	spike
UCL	21.36	52.5	11454.01
Mean Value	17.11	50.0	10187.61
LCL	12.85	47.5	8921.21

Obs #	Date	Blank C.R.	Pass ?	Quench #	Pass	Spiked C.R.	Pass ?
122	10/26/2009	16.9	OK	49	OK	10005	OK
123	10/27/2009	18.9	OK	48.4	OK	10076	OK
124	10/28/2009	16.2	OK	50.2	OK	10045.7	OK



Liquid Scintillation Counter

Quality Control Data

Daily Instrument Performance Checks

DAILY INSTRUMENT PERFORMANCE CHECKS - LS6000 (LL OFF, LUMEX OFF)

Daily IPCs consist of the following standards;

Efficiency Check -

Beckman Tritium Standard

Beckman C-14 Standard

Lot HNZ0202

Lot CNZ3112

101900.00 dpm

98500.0 dpm

2/17/2005 REF

2/17/2005 REF

2/17/2010 EXP

2/17/2010 EXP

INSTRUMENT RE-CALIBRATED FOR ALL TESTS STARTING 04/08/09. mbc

Historical Control Limits

as of 06/12/09 MH

Decay Corrected Tritium

Carbon-14

UCL 70748.54

80439.21

Mean Value 67379.56

76608.77

LCL 64010.58

72778.33

Decay Corrected

<u>Obs</u>	<u>Date</u>	<u>H-3 CPM</u>	<u>H-CPM</u>	<u>PASS?</u>	<u>C-14 CPM</u>	<u>PASS?</u>
160	12/23/2009	50835.30	66828.14	OK	76503.9	OK
161	12/26/2009	50846.80	66874.25	OK	76138.9	OK

DAILY CHECK LL ON ⁹⁹Tc SOURCE- LS6000

⁹⁹ Tc standard		SPIKE	
836.3020.70		KNOWN ACTIVITY	
12/1/2009	REF	58000.38	dpm/g
12/1/2010	EXP	58000.38	dpm

INSTRUMENT RE-CALIBRATED FOR ALL TESTS STARTING 04/08/09. mbc
 Historical Control Limits 6/12/2009

	<u>blank</u>	<u>Blank Quench #</u>	<u>spike</u>
UCL	20.61	55.9	12600.00
Mean Value	16.50	53.2	11206.90
LCL	12.40	50.5	9813.80

Obs #	Date	Blank C.R.	Pass ?	Quench #	Pass	Spiked C.R.	Pass ?
160	12/23/2009	15.9	OK	52.2	OK	11055.7	OK
161	12/26/2009	16.9	OK	53	OK	10874.6	OK