



**UNITED STATES
NUCLEAR REGULATORY COMMISSION**

REGION III
2443 WARRENVILLE ROAD, SUITE 210
LISLE, ILLINOIS 60532-4352

MAY 20 2011

Gary Schwerin
Radiation Safety Officer
City of Mt. Pleasant
Division of Public Works
Public Works Building
1303 N. Franklin Street
Mt. Pleasant, MI 48858

Dear Mr. Schwerin:

It has come to our attention that an omission existed to your NRC Material License No. 21-24504-01. This omission has been corrected by the addition of a total possession limit in Item 8.A. and 8.B. in the enclosed Corrected Copy.

Due to a change in NRC's licensing policies, we must obtain possession limits for all materials listed on your license. As we discussed by telephone on March 15, 2011, we have listed the total possession limit for cesium-137 as 9 millicuries in Item 8.A. and americium-241 as 44 millicuries in Item 8.B. as listed in the Sealed Source and Device Registry for your Troxler Model 3450 moisture density gauge.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the NRC's Agencywide Documents Access and Management System (ADAMS), accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Please contact me at (623) 214-5283, if you have any questions.

Sincerely,


James R. Mullauer, M.H.S.
Materials Licensing Branch
Region III

License No. 21-24504-01
Docket No. 030-28673

Enclosure: Corrected copy of License No. 21-24504-01