



UNITED STATES  
NUCLEAR REGULATORY COMMISSION

REGION III  
2443 WARRENVILLE ROAD, SUITE 210  
LISLE, ILLINOIS 60532-4352

JUN 15 2011

Saiyid M. Shah, Ph. D.  
Radiation Safety Officer  
St. Mary's Medical Center  
Nuclear Medicine  
Evansville, IN 47750

Dear Dr. Shah:

In your letter dated March 21, 2011, a request was made to add to License No. 13-03226-04, Kathryn Bryan, M.D. as an authorized user for 10 CFR 35.100 and 35.200. The training and experience information provided, including her American Board of Radiology certification, occurred more than 20 years ago and is not in accordance with 10 CFR 35.59, Recentness of Training. Therefore, Dr. Bryan cannot be added to your license at this time and your request has been voided.

If you wish to add Dr. Bryan to your license in the future, please review 10 CFR 35.290, 35.59 and provide recent training and experience as it applies to her use of byproduct materials under 10 CFR 35.100 and 35.200.

If you have any questions regarding this matter, please contact me at (630) 829-9842.

In accordance with 10 Code of Federal Regulations 2.390 of the NRC's "Rules of Practice," a copy of this letter will be available electronically for public inspection in the NRC Public Document Room or from the NRC's Agencywide Documents Access and Management System (ADAMS), accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Sincerely,

A handwritten signature in black ink that reads "Toye L. Simmons".

Toye L. Simmons  
Materials Licensing Branch

License No. 13-03226-04  
Docket No. 030-20812