



DEPARTMENT OF THE ARMY  
WALTER REED ARMY MEDICAL CENTER  
WASHINGTON, DC 20307-5001

REPLY TO  
ATTENTION OF:

MCHL-HP

23 June 1997

MEMORANDUM FOR Record

SUBJECT: Decommissioning Survey of Building 149A Forest Glen Annex

1. A decommissioning survey was conducted in building 149A, Forest Glen Annex on 25 April 1997. The building was surveyed with survey meters for detectable contamination. In addition swipes were collected for each square meter of the floor area and up 5 feet on every inside wall for removable contamination.
2. The building was used as a short term package storage location for shipments that arrived after normal duty hours or on weekends. No unsealed radioactive material was used at this location. The packages were stored in a freezer, refrigerator, or at room temperature depending on the package label recommendation. Based on the knowledge of the site history and previous survey information this building was not expected to contain any residual radioactive contamination.
3. A historical review indicated that no spills or unusual occurrences happened at this location.
4. The floor of the building was grided off in 3 foot squares. The grided areas were labeled as indicated in enclosure 1. The inside walls were grided from the floor to 3 feet and labeled as the floor grid coordinate plus a reference direction (N=north, E=east, S=south, W=west) and a number 1. From 3 feet up to 5 feet as the floor coordinate, reference direction and number 2. An example of a wall coordinate is A1N1 for the area on the wall adjacent to the floor coordinate A1, in the north reference direction, and the lower grid coordinate from the floor up to 3 feet.
5. The survey instrumentation was chosen to provide the highest sensitivity for detecting alpha, beta, or gamma radiation. The calibration and sensitivity data for all instrumentation used in this report is included in enclosure 2.
6. Every square yard of the floor of building 149A was grided and directly surveyed for gamma contamination using the Eberline ESP-1, serial number 975. The results of that survey are

MCHL-HP

SUBJECT: Decommissioning Survey of Building 149A Forest Glen  
Section

included in columns 1 and 4 of enclosure 3. The meter was checked for proper operation and calibrated against a NIST traceable check source. The check source was 1.81  $\mu\text{Ci}$  of Cs-137 as of December 1972 source number 475. The decayed source strength at the time of the survey was about 1.03  $\mu\text{Ci}$  of Cs-137. The measured activity of this check source was 8.0E4 CPM. The measured background level was 1.45E3 CPM. Eighteen grid locations exceeded 200 DPM above background, and six locations exceeded 2000 DPM above background (enclosure 4). The swipes from these locations indicated no contamination. All areas that had readings in excess of 200 DPM were resurveyed on 24 June 1997 and no measurements exceeding 200 DPM above background were found (enclosure 12).

7. A Ludlum L-3 with an unshielded beta probe was used to directly survey the floor of building 149A for beta or gamma contamination. The results of the survey were recorded in columns 2 and 5 of enclosure 3. The meter was checked for proper operation and calibrated against a NIST traceable check source. The check source was 0.0064  $\mu\text{Ci}$  of Am-241 as of February 1970. The decayed source strength at the time of the survey was about 0.0064  $\mu\text{Ci}$  of Am-241. The measured activity of this check source was 18E3 CPM. The measured background level was 80 CPM. The results of this survey indicated that no beta contamination was present in building 149A (enclosure 5). Note that C-14 and tritium which emit very low energy beta radiation may not be detectable using this survey meter.

8. The AN/PDR-77, serial number 798A was used to directly survey the floor of building 149A for alpha contamination. The results of the survey were recorded in columns 3 and 6 of enclosure 3. The meter was checked for proper operation and calibrated against a NIST traceable check source. The check source was 0.0064  $\mu\text{Ci}$  of Am-241 as of February 1970. The decayed source strength at the time of the survey was about 0.0064  $\mu\text{Ci}$  of Am-241. The measured activity of this check source was 10,200 CPM. The measured background level was 140 CPM. The results of this survey indicated that no alpha contamination was present in building 149A (enclosure 6).

9. Swipes were taken in each grid location and on every inside wall for removable contamination. The swipes were put in marked vials and analyzed in a Packard A5530 automatic gamma counter. The background was counted for 10 minutes and the samples were

MCHL-HP

SUBJECT: Decommissioning Survey of Building 149A Forest Glen  
Section


counted for 2 minutes each. The results are background subtracted and printed in enclosure 7. A diagram of the floor plan of the building with the walls is included as enclosure 8. The results of this survey indicate that no removable gamma contamination was present in building 149A.

10. After the swipes were analyzed in the auto-gamma counter, 10 ml of liquid scintillation fluid was added to each vial and the vials were counted in a Packard 2500LX automatic liquid scintillation counter. The background was counted for 10 minutes and the samples were counted for 2 minutes each. The results are background subtracted and printed in enclosure 9. A diagram of the floor plan of the building with the walls is included as enclosure 10. The results of this survey indicate that no removable alpha or beta contamination was present in building 149A.


11. A final survey was performed on a refrigerator that was used to temporarily store packages on 25 April 1997. The item was a Pfeifer & Son refrigerator, MMCN B8864. A Ludlum L-3 with an unshielded pancake probe, serial number 18103, as well as swipes were used to check for any radioactive contamination. The survey meter, the auto-gamma counter, and the liquid scintillation counter all indicated no contamination was present on the refrigerator (enclosure 11).

12. Based on the results of this survey, building 149A, Forest Glen Annex, can be removed from the WRAMC NRC license and is considered free from radioactive contamination.

12 Encls  
as

  
ARTHUR R. MORTON  
CPT, MS  
Chief, Operations Branch, HPO

# Building 149A Bunker



TSC-ABERDEEN, APC. MD.

# SURVEY INSTRUMENT CALIBRATION REPORT AN/PDR-77

Serial Number: 0798A

Radionuclide: Cs-137

OUIC: WADHDL

PUIC: W459QC

Battery Check: OK

Beta/Gamma Probe Low Range Detector (DT-616/VDR-2)			
Applied	Read		Correction Factor
2 R/hr	1.98	K mR/hr	1.01
200 mR/hr	200	mR/hr	1.00
80 mR/hr	79.8	mR/hr	.999
20 mR/hr	19.6	mR/hr	1.008
8 mR/hr	8.3	mR/hr	.96
0.8 mR/hr	.803	mR/hr	.999
0.08 mR/hr	.078	mR/hr	1.02
Beta/Gamma Probe High Range Detector (DT-616/VDR-2)			
Applied	Read		Correction Factor
200 R/hr	203	K mR/hr	1.02
80 R/hr	83.3	K mR/hr	1.05
20 R/hr	21.8	K mR/hr	1.09
8 R/hr	8.29	K mR/hr	.96

Calibration Report Number: WADHDL012C

Date: 9 MAY 97

CALIBRATION DUE DATE 5 NOV 97

Radionuclide: Pu-239

SERIAL NUMBER 01798H

Alpha Probe (DT-669/PDR-77)		
Applied	Read	Correction Factor
130 K cpm	131 K cpm	.99
10.7 K cpm	11.4 K cpm	.94
.877 K cpm	.860 K cpm	1.08
* Check Source	9.18 K cpm	N/A

\* check source (Th-232) measurement obtained with Alpha side up, centered, and flush against detector.

Radionuclide: Am-241

X-ray Probe (DT-674/PDR-77)				
Energy/Position	Instructions	Applied	Read	Correction Factor
17 KeV	* Center	24.1 K cpm	23.9 K cpm	1.00
17 KeV	**	check source	1.15 K cpm	N/A
60 KeV	Center	45.6 K cpm	K cpm	1.08
60 KeV	**	check source	2.39 K cpm	N/A

\* Center measurements were obtained 12 inches directly above source.

\*\* Check source (Th-232) measurement obtained with X-ray side up, centered, and flush against detector.

Calibration Report Number: WADH01012C

Date: 9 MAY 97

CALIBRATION DUE DATE 5 NOV 97

In Charge of Calibration

Reviewer

AREA NUCLEONICS LABORATORY  
REPORT OF CALIBRATION  
FOR  
RADIAC SET/SURVEY METER

MODEL:ESP-2  
SERIAL NO.:975

SUBMITTED BY:

W2DH01

The measurements were performed Under ambient conditions of Approximately 22.7 Degrees C and 40 Percent relative humidity.

This instrument was calibrated in accordance with TB9-6665-285-15, Army Calibration Program for Radiac Meters.


The reference standard for this calibration is an Atlan-Tech Cesium 137 source, serial number 0393-102, Model GC60, Calibrated 9 DEC.1996.

Calibration uncertainty including measuring errors and accuracy of the reference standard is +/-10%.

This calibration is traceable to and compatible with National Institute of Standards and Technology (NIST) measurements.

In charge of Test:

  
THOMAS KACZOROWSKI

  
DAVID A. JAMISON  
Radiation Protection Officer  
US Army DTSC-Aberdeen

Calibration Report No.:W2DH01190C  
Calibration Date: 22 MAR 97  
Calibration Due Date: 20 JUL 97  
Page 1 Of 2 pages

# RADIAC INSTRUMENT DATA SHEET

HIGH VOLTAGE=1000V

DT= 1.40-05

CC= 1.00+00

INSTRUMENT MAKE: EBERLINE

MODEL NO.: ESP-2

SER.NO.: 975

PROBE TYPE: HB63K/5

SERIAL NO: LS177R

RATE: 50KCPM/mR/HR



CALIBRATION GEOMETRY:

SCALE	SOURCE	ATTN	DISTANCE CM	INSTRUMENT METER READING		
				RATE mR/HR	ADJUSTMENT CPM BEFORE	ADJUSTMENT CPM AFTER
DIGITAL	150mci	X2,X10	146.89	1	44.1K	
	150mci	X2,X10	208.58	.5	21.1K	
	150mci	X4,X100	121.66	.1	3.99K	

Background Reading= .278K



1. JOB ORDER NUMBER <b>97064-6</b>		2. PRIORITY <b>13</b>		3. ID CODE <b>L7827</b>		4. TB <b>MPR MANUAL</b>		5. TM	
6. SCHEDULE UIC <b>W459QC</b>		7. OWNER UIC <b>W2DH01</b>		8. SERIAL NUMBER <b>975</b>		9. MODEL NUMBER <b>ESP2</b>		10. MFG CODE <b>04596</b>	
11. NATIONAL STOCK NUMBER <b>6665013483931</b>				12. DUE DATE <b>8 Mar 97</b>		13. INT <b>D</b>		14. SYS CODE <b>505</b>	
15. NOMENCLATURE <b>METR RADIAC</b>				16. WORK C/T		17. LEV <b>S</b>		18. ACT DATE <b>97081</b>	
19. PERFORM UIC <b>W459QC</b>		20. TYPE STANDARD <b>MW ELEC PHYS IR RAD E</b>				21. ERC		22. UPDATE <b>ADD CHAN-GE DE-LETE</b>	
23. REPORT		24. CHANGE DATA							
25. DEFICIENCIES/SYMPTOMS									
FUND CODE: <b>D</b> FBC: <b>7R1001</b>									
26. STATUS				27. CALIBRATION CONDITION		28. REPAIR/ADJUSTMENT		29. REPAIR	
PUIC CODE DATE INIT <b>QC N 97081 13</b>				<input checked="" type="checkbox"/> WITHIN TOLERANCE <input type="checkbox"/> OUT OF TOLERANCE		<input type="checkbox"/> BY TEAM/LAB <input type="checkbox"/> DEFERRED MAINT <input type="checkbox"/> NRTS/ <input type="checkbox"/> CODE OUT		TECH CODE HOURS <b>02 5</b>	
				31. PARTS ADJUSTED*					
				32. PARTS REPLACED/					
				33. REMARKS @					
34. PARTS REQUISITIONED									
REF DESIG	NOMENCLATURE	MFG CODE	NSN/PN	QTY	UNIT ISSUE	UNIT COST	DOCUMENT NUMBER	RECEIVED	
35. SUBMITTED				36. RECEIVED UNDER LIMITED TECHNICAL INSPECTION			37. ACCEPTED		
BY: <i>John Depina</i> DATE: <b>5/3/97</b>				BY: <i>[Signature]</i> DATE: <b>97064</b>			DATE: <b> </b>		

DEPARTMENT OF THE ARMY

US ARMY TMDE/SUPPORT CENTER-ABERDEEN  
ABERDEEN PROVING GROUNDS, MD.

AREA NUCLEONICS LABORATORY  
REPORT OF CALIBRATION  
FOR  
RADIAC SET/SURVEY METER

MODEL: ANALYST  
SERIAL NO.: A948P

SUBMITTED BY:

W2DH01

The measurements were performed under ambient conditions of approximately 22.7 Degrees C and 40 percent relative humidity.

This instrument was calibrated in accordance with TB9-6665-285-15, Army Calibration Program for Radiac Meters.

The reference standard for this Calibration is a model 500-2 LUDLUM Pulser, serial number 129168, Calibrated 5 MAR. 1997.

Calibration uncertainty including measuring errors and the accuracy of the reference standard is  $\pm 10\%$ .

This calibration is traceable to and compatible with National Institute of standards and technology (NIST) measurements.

In charge of test:

STLAN-TECH, Model 5006, Cs-137, Used to check BAYPA probe only.  
SCALE: SOURCE: ATTN: DISTANCE: DAVID A. JAMISON

X1A: 150mCi X2: X13 146.85

DAVID A. JAMISON  
Radiation Protection Officer  
US ARMY DTSC-Aberdeen

AN/UDM-6, PU-239, used to check ALPHA probe only.

AN/UDM-6, PU-239, used to check ALPHA probe only.

Calibration Report No: W2DH01124C

Calibration Date: 29 MAY 97

Calibration Due Date: 25 NOV 97

Page 1 of 2 pages

# RADIAC INSTRUMENT DATA SHEET

INSTRUMENT MAKE: BICRON

MODEL NO.: ANALYST

SER. NO.: A948P

PROBE TYPE: G5

SERIAL NO.: A3760

RATE: 200KCPM/mR/HR

CALIBRATION GEOMETRY:



SCALE SOURCE ATTN DISTANCE  
CM

INSTRUMENT METER READING  
RATE CPM BEFORE ADJUSTMENT AFTER

X1K 400K 400K  
100K 100K

X100 40K 39K  
10K 10K

X10 4K 3.9K  
1K 1K

X1 400 398  
100 100

ATLAN-TECH, Model GC60, Cs-137, Used to check GAMMA probe only.

SCALE SOURCE ATTN DISTANCE  
CM

X1K 150mCi X2, X10 146.89 mR/HR CPM  
1 125K

AN/UDM-6, PU-239, used to check ALPHA probe only.

X100 P-1647 AN/UDM6 16.5K NOT USED

SENSITIVITY: 100 mV

HIGH VOLTAGE: 1200 V

CALIBRATION REPORT NO.: W2DH01124C

DATE: 29 MAY 97

PAGE 2 OF 2

1. JOB ORDER NUMBER <b>DC 127-24</b>		2. PRIORITY		3. ID CODE <b>A5160</b>		4. TB		5. TM																																																																																																																	
6. SCHEDULE UIC <b>W459UC</b>		7. OWNER UIC <b>W2DH01</b>		8. SERIAL NUMBER <b>A948P</b>		9. MODEL NUMBER <b>ANALYST</b>		10. MFG CODE <b>10617</b>																																																																																																																	
11. NATIONAL STOCK NUMBER <b>6665000000000</b>				12. DUE DATE <b>15 Feb 97</b>		13. INT <b>F</b>		14. SYS CODE <b>U82</b>																																																																																																																	
15. NOMENCLATURE <b>METR RADIAC</b>				16. WORK CTR		17. LEV <b>F</b>		18. ACT DATE <b>9 MAY 97</b>																																																																																																																	
19. PERFORM UIC <b>W459UC</b>		20. TYPE STANDARD <b>MW ELEC PHYS IR RAD E</b>				21. ERC		22. UPDATE <b>ADD</b>																																																																																																																	
23. REPORT		24. CHANGE DATA		25. DEFICIENCIES/SYMPTOMS		26. FUND CODE <b>D PBC1</b>		27. FUND CODE <b>FR1001</b>																																																																																																																	
28. STATUS				29. CALIBRATION CONDITION		30. REPAIR/ADJUSTMENT		31. REPAIR																																																																																																																	
<table border="1" style="width:100%;"> <tr> <th>PUIC</th> <th>CODE</th> <th>DATE</th> <th>INIT</th> </tr> <tr><td> </td><td> </td><td> </td><td> </td></tr> <tr><td> </td><td> </td><td> </td><td> </td></tr> <tr><td> </td><td> </td><td> </td><td> </td></tr> <tr><td> </td><td> </td><td> </td><td> </td></tr> <tr><td> </td><td> </td><td> </td><td> </td></tr> <tr><td> </td><td> </td><td> </td><td> </td></tr> <tr><td> </td><td> </td><td> </td><td> </td></tr> <tr><td> </td><td> </td><td> </td><td> </td></tr> <tr><td> </td><td> </td><td> </td><td> </td></tr> </table>				PUIC	CODE	DATE	INIT																																					<table border="1" style="width:100%;"> <tr> <th colspan="2">WITHIN TOLERANCE</th> <th colspan="2">OUT OF TOLERANCE</th> </tr> <tr><td>A</td><td> </td><td>B</td><td> </td></tr> <tr><td>B</td><td> </td><td>C</td><td> </td></tr> </table>		WITHIN TOLERANCE		OUT OF TOLERANCE		A		B		B		C		<table border="1" style="width:100%;"> <tr> <th colspan="2">BY TEAM/LAB</th> <th colspan="2">DEFERRED MAINT</th> <th colspan="2">NRTS/ CODE OUT</th> </tr> <tr><td>A</td><td> </td><td>B</td><td> </td><td>C</td><td> </td></tr> <tr><td>B</td><td> </td><td>C</td><td> </td><td>D</td><td> </td></tr> </table>		BY TEAM/LAB		DEFERRED MAINT		NRTS/ CODE OUT		A		B		C		B		C		D		<table border="1" style="width:100%;"> <tr> <th>TECH CODE</th> <th>HOURS</th> </tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> </table>		TECH CODE	HOURS																			<table border="1" style="width:100%;"> <tr> <th>TECH CODE</th> <th>HOURS</th> </tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> </table>		TECH CODE	HOURS																		
PUIC	CODE	DATE	INIT																																																																																																																						
WITHIN TOLERANCE		OUT OF TOLERANCE																																																																																																																							
A		B																																																																																																																							
B		C																																																																																																																							
BY TEAM/LAB		DEFERRED MAINT		NRTS/ CODE OUT																																																																																																																					
A		B		C																																																																																																																					
B		C		D																																																																																																																					
TECH CODE	HOURS																																																																																																																								
TECH CODE	HOURS																																																																																																																								
32. PARTS ADJUSTED				33. PARTS REPLACED																																																																																																																					
34. PARTS ADJUSTED				35. PARTS REPLACED																																																																																																																					
36. REMARKS @				37. REMARKS @																																																																																																																					
38. PARTS REQUISITIONED				39. PARTS REQUISITIONED																																																																																																																					
<table border="1" style="width:100%;"> <tr> <th>REF DESIG</th> <th>NOMENCLATURE</th> <th>MFG CODE</th> <th>NSN/PN</th> <th>QTY</th> <th>UNIT ISSUE</th> <th>UNIT COST</th> <th>DOCUMENT NUMBER</th> <th>RECEIVED</th> </tr> <tr><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td></tr> <tr><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td></tr> <tr><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td></tr> <tr><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td></tr> <tr><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td></tr> <tr><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td></tr> <tr><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td></tr> <tr><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td></tr> <tr><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td></tr> </table>				REF DESIG	NOMENCLATURE	MFG CODE	NSN/PN	QTY	UNIT ISSUE	UNIT COST	DOCUMENT NUMBER	RECEIVED																																																																																		<table border="1" style="width:100%;"> <tr> <th>DATE</th> <th>BY</th> </tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> <tr><td> </td><td> </td></tr> </table>						DATE	BY																				
REF DESIG	NOMENCLATURE	MFG CODE	NSN/PN	QTY	UNIT ISSUE	UNIT COST	DOCUMENT NUMBER	RECEIVED																																																																																																																	
DATE	BY																																																																																																																								
35. SUBMITTED				36. RECEIVED UNDER LIMITED TECHNICAL INSPECTION				37. ACCEPTED																																																																																																																	
BY:                      DATE:				BY:                      DATE:				BY:                      DATE:																																																																																																																	

US ARMY TMDE SUPPORT CENTER-ABERDEEN  
ABERDEED PROVING GROUNDS,MD.

AREA NUCLEONICS LABORATORY  
REPORT OF CALIBRATION  
FOR  
RADIAC SET/SURVEY METER

MODEL:5  
SERIAL NO.:18103

SUBMITTED BY:

W2DH01

The measurements were performed under ambient conditions of approximately 22.7 Degrees C and 40% percent relative humidity.



This instrument was calibrated In accordance with TB9-6665-285-15, Army Calibration Program for Radiac Meters.

The reference standard for this Calibration is a model 500-Z LUDLUM Pulser, serial number 129168, Calibrated 30 JAN. 1996.

Calibration uncertainty including measuring errors and the accuracy Of the reference standard is +/- 10%.

This calibration is traceable to and compatible with National Institute of standards and technology (NIST) measurements.

In charge of test:

  
JUNG K. PARK  
  
DAVID A. JAMISON  
Radiation Protection Officer  
US. ARMY DTSC-Aberdeen

Calibration Report No.: W2DH01059C  
Calibration Date: 29 OCT 1996  
Calibration Due Date: 24 OCT 1997  
Page 1 of 2 pages

# RADIAC INSTRUMENT DATA SHEET

INSTRUMENT MAKE: LUDLUM

MODEL NO. 3

SER.NO.18103

PROBE TYPE: 44-9

SERIAL NO.: 18103-1

RATE:3300CPM/mR/HR

CALIBRATION GEOMETRY:

Check Source Reading at surface Level =18K

SCALE	SOURCE MODE	DISTANCE CM	INSTRUMENT METER READING	
			RATE CPM	ADJUSTMENT BEFORE AFTER
X100			400K	400K
			100K	100K
X10			40K	40K
			10K	10K
X1			4K	4K
			1K	1K
X.1			400	400
			100	100

Probe checked with= Americium-241

Activity: (2.65 kbq) 71.6 nci +/-10%.

Count rate in the alpha-plateau at 5 mm distance from the  
Surface of the source= 79,524 CPM.

	CPM	BEFORE	AFTER
X10	79,524	26K	

PROBE EFFICIENCY= 32.7%

CALIBRATION REPORT NO.W2DH01059C  
PAGE 2 OF 2

DATE 29 OCT 1996

TOTAL P.03

### Instrument Detection Sensitivity

For an integrated measurement over a preset time, the minimum detectable activity (MDA) for surface activity can be approximated by:

$$MDA = \frac{2.71 + 4.65 \sqrt{B_r \cdot t}}{t \cdot E \cdot \frac{1}{100}}$$

where,

MDA = activity level in DPM/100 cm<sup>2</sup>  
B<sub>r</sub> = measured background rate in CPM  
t = counting time in minutes  
E = detector efficiency in counts/disintegrations  
A = active area of the probe in cm<sup>2</sup>

Meter	SN	B <sub>r</sub> (CPM)	t (min)	E	A (cm <sup>2</sup> )	MDA (DPM)
L-3	18,103	80	1	0.327	20.27	668.0
AN/PDR-77	798A	14	1	0.990	129.00	15.7
Bicron	A948P	550	1	0.600	506.70	37.0
Eberline	975	1,450	1	0.600	506.70	59.1

Coordinate	Alpha	Beta	Gamma	Coordinate	Alpha	Beta	Gamma
A1	42.4	100	1.89 E3	D12	23.7	150	6.43 E3
A2	39.1	110	1.43 E3	D13	24.2	110	7. E3
A3	32.4	110	1.49 ..	E1	12.8	160	1.35 ..
A4	12.6	90	1.3 ..	F2	11.1	160	1.55 ..
A5	14.3	100	1.25 ..	E3	13.8	140	1.25 ..
A6	16.7	90	1.2 ..	E4	25.1	220	1.35 ..
A7	17.8	40	1.3 ..	E5	28	120	1.40 ..
A8	23	40	1.2 ..	E6	20.6	150	1.51 ..
A9	22.4	85	1.26 ..	E7	24.3	120	1.33 ..
A10	16.3	95	1.35 ..	E8	17.9	120	1.27 ..
A11	19.2	50	1.5 ..	E9	16.8	160	1.42 ..
A12	22.9	70	1.9 ..	E10	10.6	130	1.43 ..
A13	19.3	95	1.7 ..	E11	19.2	150	1.5 ..
B1	25.4	105	1.76 E3	E12	25	180	1.4 ..
B2	30.7	90	1.35 E3	E13	32.7	130	8.6 ..
B3	32.9	100	1.59 ..	F1	19.9	80	1.22 ..
B4	15.7	105	1.2 ..	F2	25.5	115	1.20 ..
B5	11.4	120	1.3 ..	F3	20.4	90	1.40 ..
B6	19.5	140	1.3 ..	F4	22.3	105	1.47 ..
B7	23	120	1.2 ..	F5	22.2	100	1.51 E3
B8	23	120	1.9 ..	F6	27.4	70	1.36 E3
B9	19.3	105	1.3 ..	F7	20.8	75	1.41 ..
B10	12.4	200	1.5 ..	F8	10.4	100	1.44 ..
B11	22.6	130	1.75 ..	F9	19.2	110	1.32 ..
B12	19.6	70	1.71 ..	F10	11.9	100	1.10 ..
B13	22.3	160	1.58 ..	F11	21	120	1.45 ..
C1	26.2	70	1.65 E3	F12	11.8	80	1.24 ..
C2	27.8	60	1.25 E3	F13	13.5	80	1.21 ..
C3	21.5	120	1.57 ..	G1	23.4	120	1.4 ..
C4	20.5	180	1.32 ..	G2	20.6	110	1.4 ..
C5	20.6	200	1.42 ..	G3	28	105	1.3 ..
C6	11.4	160	1.58 ..	G4	26.9	120	1.45 ..
C7	28.5	95	1.55 ..	G5	20.9	100	1.4 E3
C8	17.6	140	1.72 ..	G6	29.8	110	1.35 E3
C9	16.8	80	1.85 ..	G7	17.2	90	1.42 ..
C10	16.6	70	2.15 ..	G8	14.0	90	1.53 ..
C11	16.5	120	2.1 ..	G9	16.4	110	1.68 ..
C12	16.9	110	1.6 ..	G10	14.4	70	1.62 ..
C13	19.8	100	8.1 E3	G11	17.9	70	1.32 ..
D1	21	120	4.4 E3	G12	13.5	90	1.28 ..
D2	13.8	160	4.45 ..	G13	12.3	110	1.43 ..
D3	20.9	110	1.45 ..				
D4	25.5	140	1.45 ..				
D5	27.6	150	4.6 ..				
D6	15.1	240	1.5 ..				
D7	25.6	200	1.58 ..				
D8	20.5	150	1.4 ..				
D9	19.1	110	1.6 ..				
D10	12.2	110	1.5 ..				
D11	21.3	130	1.6 ..				

Gamma Meter: Electronic datalogger  
Beta Meter: 1.3 Ludlum  
Alpha Meter: AMPOR 77

Cs-137  
Chk Source: Am 241  
Chk Source: Am 241

8.0 E4  
Msr: 1.45 E3 Bkgd: 1.45 E3  
Msr: 1.45 E3 Bkgd: 80  
Msr: 1.45 E3 Bkgd: 140 cpm  
15000 cpm  
10,200 cpm

Am-241 0.0064  $\mu$ Li 2/70 616-3

Cs-137 1.81  $\mu$ Li 12/72 475

ENCL 3



Gamma Meter	Source	Activity	Date	Source	Bkgd
Eberline Dat	Cs-137	uCi	Dec 1972	CPM	CPM
		1.81		80,000	1.45E+03

## Building 149A Forest Glen Section (Bunker)

ENCL 4

Gamma Meter Readings (CPM) - Not Background Subtracted

	1	2	3	4	5	6	7	8	9	10	11	12	13	
A	1.89E+03	1.43E+03	1.49E+03	1.3E+03	1.25E+03	1.2E+03	1.3E+03	1.2E+03	1.26E+03	1.35E+03	1.50E+03	1.8E+03	1.7E+03	A
B	1.76E+03	1.35E+03	1.6E+03	1.2E+03	1.3E+03	1.3E+03	1.2E+03	1.9E+03	1.3E+03	1.5E+03	1.75E+03	1.71E+03	1.58E+03	B
C	1.65E+03	1.25E+03	1.37E+03	1.32E+03	1.42E+03	1.5E+03	1.55E+03	1.72E+03	1.85E+03	2.15E+03	2.1E+03	1.6E+03	8.1E+03	C
D	4.4E+03	4.45E+03	1.45E+03	1.45E+03	4.6E+03	1.5E+03	1.5E+03	1.4E+03	1.6E+03	1.5E+03	1.6E+03	1.43E+03	7E+03	D
E	1.35E+03	1.55E+03	1.25E+03	1.35E+03	1.4E+03	1.51E+03	1.33E+03	1.27E+03	1.42E+03	1.43E+03	1.5E+03	1.4E+03	8.6E+03	E
F	1.22E+03	1.26E+03	1.4E+03	1.47E+03	1.51E+03	1.36E+03	1.41E+03	1.44E+03	1.32E+03	1.1E+03	1.05E+03	1.24E+03	1.21E+03	F
G	1.4E+03	1.4E+03	1.3E+03	1.45E+03	1.40E+03	1.35E+03	1.42E+03	1.53E+03	1.6E+03	1.62E+03	1.32E+03	1.2E+03	1.43E+03	G

Gamma Meter Readings (CPM) - Background Subtracted

	1	2	3	4	5	6	7	8	9	10	11	12	13	
A	440	0	40	0	0	0	0	0	0	0	50	350	250	A
B	310	0	140	0	0	0	0	450	0	50	300	260	130	B
C	200	0	0	0	0	50	100	270	400	700	650	150	6650	C
D	2950	3000	0	0	3150	50	50	0	150	50	150	0	5550	D
E	0	100	0	0	0	60	0	0	0	0	50	0	7150	E
F	0	0	0	20	60	0	0	0	0	0	0	0	0	F
G	0	0	0	0	0	0	0	80	150	170	0	0	0	G

Beta	Source	Activity	Date	Source	Bkgd
Meter		uCi		CPM	CPM
Ludlum L-3	Am-241	0.0064	Feb 1970	18,000	80

## Building 149A Forest Glen Section (Bunker)

ENCLOSURE

Beta Meter Readings (CPM) - Not Background Subtracted

	1	2	3	4	5	6	7	8	9	10	11	12	13	
A	100	110	110	80	100	90	40	40	85	95	50	70	95	A
B	105	90	100	105	120	140	120	120	105	200	130	70	160	B
C	70	60	100	100	200	160	95	140	80	70	120	110	100	C
D	120	160	110	140	150	240	200	150	110	110	130	150	110	D
E	160	160	180	220	120	150	120	120	160	130	150	180	130	E
F	80	115	90	105	100	70	75	100	110	100	120	80	80	F
G	120	110	105	120	100	110	90	90	110	70	70	90	110	G

Beta Meter Readings (CPM) - Background Subtracted

	1	2	3	4	5	6	7	8	9	10	11	12	13	
A	20	30	30	0	20	10	0	0	5	15	0	0	15	A
B	25	10	20	25	40	60	40	40	25	120	50	0	80	B
C	0	0	20	20	120	80	15	60	0	0	40	30	20	C
D	40	80	30	60	70	160	120	70	30	30	50	70	30	D
E	80	80	100	140	40	70	40	40	80	50	70	100	50	E
F	0	35	10	25	20	0	0	20	30	20	40	0	0	F
G	40	30	25	40	20	30	10	10	30	0	0	10	30	G

Alpha Meter	Source	Activity	Date	Source	Bkgd
ANPDR-77	Am-241	uCi	Feb 1970	CPM	CPM
		0.0064		10,200	14

## Building 149A Forest Glen Section (Bunker)

9776

Alpha Meter Readings (CPM) - Not Background Subtracted

	1	2	3	4	5	6	7	8	9	10	11	12	13	
A	42.4	39.1	32.4	12.6	14.3	16.7	17.8	23	22.4	16.3	19.2	22.9	19.3	A
B	25.4	30.7	32.9	15.7	11.4	19.5	23	23	19.3	12.4	22.6	19.6	22.3	B
C	26.2	27.8	21.5	20.5	20.6	11.4	28.5	17.6	16.8	10.6	10.5	16.9	19.8	C
D	21	13.8	20.9	25.5	27.6	15.1	25.6	20.5	19.1	12.2	21.3	23.7	24.2	D
E	12.8	11.1	13.8	25.1	28	20.6	24.3	17.9	16.8	10.6	19.2	25	32.7	E
F	19.9	25.5	20.4	22.3	22.2	27.4	20.8	10.9	19.2	11.9	21	11.8	13.5	F
G	23.4	20.6	28	26.9	20.9	29.8	17.2	14	16.4	14.4	17.9	13.5	12.3	G

Alpha Meter Readings (CPM) - Background Subtracted

	1	2	3	4	5	6	7	8	9	10	11	12	13	
A	28.4	25.1	18.4	0	0.3	2.7	3.8	9	8.4	2.3	5.2	8.9	5.3	A
B	11.4	16.7	18.9	1.7	0	5.5	9	9	5.3	0	8.6	5.6	8.3	B
C	12.2	13.8	7.5	6.5	6.6	0	14.5	3.6	2.8	0	0	2.9	5.8	C
D	7	0	6.9	11.5	13.6	1.1	11.6	6.5	5.1	0	7.3	9.7	10.2	D
E	0	0	0	11.1	14	6.6	10.3	3.9	2.8	0	5.2	11	18.7	E
F	5.9	11.5	6.4	8.3	8.2	13.4	6.8	0	5.2	0	7	0	0	F
G	9.4	6.6	14	12.9	6.9	15.8	3.2	0	2.4	0.4	3.9	0	0	G

PROGRAM #.

02/20/01 17:02

REGION A: LL= 15 UL= 85 BKG= 0 %SIGMA= .00  
 REGION B: LL= 85 UL= 150 BKG= 0 %SIGMA= .00  
 REGION C: LL= 150 UL= 275 BKG= 0 %SIGMA= .00  
 REGION D: LL= 275 UL= 400  
 REGION E: LL= 400 UL= 900  
 REGION F: LL= 900 UL= 1400  
 TIME= 2.00 SCREENING LIMITS= 0 0

P#	S#	TIME	CPMA	CPMB	CPMC	CPMD	CPME	CPMF	FLAGS	MIN
5	0	10.00	32	25	55	34	67	30	B	10
5	1	2.00	0	0	0	1	0	0		13
5	2	2.00	0	0	0	5	0	0		15
5	3	2.00	0	1	0	2	2	0		17
5	4	2.00	2	0	0	3	0	0		19
5	5	2.00	0	0	0	5	0	0		21
5	6	2.00	0	0	0	0	0	0		23
5	7	2.00	1	4	0	2	0	0		25
5	8	2.00	0	5	0	7	5	0		28
5	9	2.00	0	3	0	9	0	1		30
5	10	2.00	0	2	3	0	0	0		32
5	11	2.00	4	0	0	1	0	0		34
5	12	2.00	0	5	0	0	0	3		36
5	13	2.00	1	10	0	0	0	1		38
5	14	2.00	0	0	0	0	0	0		41
5	15	2.00	2	1	0	0	4	3		43
5	16	2.00	1	0	0	0	0	0		45
5	17	2.00	0	0	0	4	0	0		47
5	18	2.00	0	1	0	0	2	2		49
5	19	2.00	1	5	0	2	0	0		51
5	20	2.00	0	0	0	0	0	0		53
5	21	2.00	0	0	0	6	0	0		56
5	22	2.00	5	3	0	0	1	4		58
5	23	2.00	0	6	0	0	0	0		60
5	24	2.00	0	1	0	5	0	1		62
5	25	2.00	2	0	4	0	0	0		64
5	26	2.00	0	0	0	0	0	0		66
5	27	2.00	6	0	8	0	0	0		68
5	28	2.00	0	0	0	3	0	0		71
5	29	2.00	0	0	0	0	0	0		73
5	30	2.00	1	1	0	4	0	0		75
5	31	2.00	0	0	0	0	0	0		77
5	32	2.00	8	4	0	3	0	0		79
5	33	2.00	0	4	0	1	2	0		81
5	34	2.00	1	0	0	4	0	0		84
5	35	2.00	0	2	0	0	0	0		86
5	36	2.00	6	0	0	0	0	0		88
5	37	2.00	0	2	0	0	0	0		90
5	38	2.00	0	0	0	0	0	0		92
5	39	2.00	0	3	0	0	0	0		94
5	40	2.00	0	3	0	1	0	1		96
5	41	2.00	0	3	0	0	0	0		99
5	42	2.00	0	3	0	3	0	0		101

ENCL 7

ST	ST	TIME	CFMA	CFMB	CFMC	CFMD	CFME	CFMF	CFMF	CFMF	MIN
5	43	2.00	0	0	0	2	2	0	0	0	103
5	44	2.00	0	0	0	4	0	0	0	0	105
5	45	2.00	3	3	0	0	0	0	0	0	107
5	46	2.00	0	0	0	0	0	0	0	0	109
5	47	2.00	5	6	1	1	0	0	0	0	112
5	48	2.00	3	0	0	0	0	0	0	0	114
5	49	2.00	6	9	0	0	0	0	0	0	116
5	50	2.00	0	2	0	0	0	0	0	0	118
5	51	2.00	1	0	0	3	0	0	0	0	120
5	52	2.00	5	0	0	0	2	0	0	0	122
5	53	2.00	0	2	0	0	4	0	0	0	125
5	54	2.00	5	2	0	0	2	0	0	0	127
5	55	2.00	0	4	0	0	3	0	0	0	129
5	56	2.00	1	4	0	2	0	0	0	0	131
5	57	2.00	0	0	0	0	3	4	0	0	133
5	58	2.00	1	6	0	0	0	0	0	0	135
5	59	2.00	0	5	0	6	0	0	0	0	137
5	60	2.00	0	0	0	0	0	0	0	0	139
5	61	2.00	1	0	4	2	1	0	0	0	142
5	62	2.00	0	4	0	0	0	0	0	0	144
5	63	2.00	0	4	0	2	0	2	0	0	146
5	64	2.00	0	0	0	7	0	1	0	0	148
5	65	2.00	0	0	0	0	0	0	0	0	150
5	66	2.00	0	1	0	0	0	0	0	0	152
5	67	2.00	0	0	0	0	0	0	0	0	155
5	68	2.00	1	0	0	0	0	3	0	0	157
5	69	2.00	0	0	0	0	0	0	0	0	159
5	70	2.00	2	2	0	0	2	1	0	0	161
5	71	2.00	0	1	0	1	0	0	0	0	163
5	72	2.00	3	1	0	5	0	0	0	0	165
5	73	2.00	0	2	0	0	0	0	0	0	168
5	74	2.00	0	5	0	0	4	0	0	0	170
5	75	2.00	1	0	0	0	0	0	0	0	172
5	76	2.00	1	0	0	5	0	0	0	0	174
5	77	2.00	1	4	0	0	4	0	0	0	176
5	78	2.00	0	1	0	3	0	0	0	0	178
5	79	2.00	0	4	0	6	4	0	0	0	180
5	80	2.00	0	0	0	0	0	0	0	0	182
5	81	2.00	0	0	0	0	0	0	0	0	185
5	82	2.00	2	6	0	0	0	0	0	0	187
5	83	2.00	0	0	0	6	12	5	0	0	189
5	84	2.00	2	1	0	0	5	0	0	0	191
5	85	2.00	0	0	0	0	0	0	0	0	193
5	86	2.00	0	3	4	4	0	4	0	0	195
5	87	2.00	0	0	0	2	0	0	0	0	198
5	88	2.00	0	4	0	0	0	0	0	0	200
5	89	2.00	2	0	0	3	0	0	0	0	202
5	90	2.00	1	3	0	0	0	0	0	0	204
5	91	2.00	0	1	0	2	0	0	0	0	206
5	92	2.00	0	0	0	0	1	0	0	0	208
5	93	2.00	0	1	0	0	0	0	0	0	211

ST	SN	TIME	CFNA	CFNB	CFNC	CFND	CFNE	CFNF	FLAGS	MIN
5	94	2.00	0	0	0	0	0	1	C4	213
5	95	2.00	3	0	0	3	4	0	E1	215
5	98	2.00	2	0	4	0	2	0	E2	217
5	97	2.00	0	4	0	0	0	0	C5	219
5	98	2.00	4	0	1	0	0	0	E1	221
5	99	2.00	0	4	0	0	0	0	E2	223
5	100	2.00	0	8	0	1	3	5	C6	226
5	101	2.00	2	3	0	2	1	4	E1	228
5	102	2.00	0	3	0	0	3	0	E2	230
5	103	2.00	0	0	0	0	0	0	C7	232
5	104	2.00	0	1	0	0	0	0	E1	234
5	105	2.00	0	7	0	2	0	0	E2	236
5	106	2.00	0	4	0	2	1	0	C8	238
5	107	2.00	0	3	0	0	16	4	C1	241
5	108	2.00	0	0	0	1	0	0	E2	243
5	109	2.00	1	4	0	0	0	0	C9	245
5	110	2.00	0	1	0	12	7	0	C10	247
5	111	2.00	1	0	0	5	3	0	E1	249
5	112	2.00	0	1	0	0	2	0	E2	251
5	113	2.00	0	0	0	0	3	0	C11	254
5	114	2.00	7	0	0	0	0	0	E1	256
5	115	2.00	3	1	0	0	1	0	E2	258
5	116	2.00	0	6	0	3	0	0	D1	260
5	117	2.00	2	3	0	0	0	0	N1	262
5	118	2.00	1	6	0	0	0	0	N2	264
5	119	2.00	0	10	0	3	0	0	D2	266
5	120	2.00	0	4	0	4	0	1	D3	268
5	121	2.00	0	0	0	0	0	0	D4	271
5	122	2.00	1	7	0	3	19	0	D5	273
5	123	2.00	0	0	6	0	0	0	D6	275
5	124	2.00	0	0	0	0	0	2	D7	277
5	125	2.00	0	0	6	4	0	3	D8	279
5	126	2.00	0	0	0	7	6	3	D9	282
5	127	2.00	2	2	0	2	0	0	D10	284
5	128	2.00	0	1	0	2	0	0	D11	286
5	129	2.00	0	4	1	0	0	0	D12	288
5	130	2.00	0	2	7	0	2	0	D13	290
5	131	2.00	2	0	0	0	0	0	E1	292
5	132	2.00	6	0	0	3	0	0	E2	295
5	133	2.00	4	2	0	0	3	0	E1	297
5	134	2.00	4	0	0	0	2	0	N1	299
5	135	2.00	0	0	0	0	0	0	N2	301
5	136	2.00	4	6	0	4	0	4	E2	303
5	137	2.00	5	3	0	0	7	0	E3	305
5	138	2.00	3	2	3	1	0	0	E4	307
5	139	2.00	7	0	0	2	6	3	E5	309
5	140	2.00	1	3	2	4	3	0	E6	312
5	141	2.00	6	8	0	6	1	0	E7	314
5	142	2.00	0	0	0	1	0	0	E8	316
5	143	2.00	2	1	0	0	0	0	E9	318
5	144	2.00	0	2	12	2	0	3	E10	320

FW	SW	TIME	CPMA	CPMB	CPMC	CPMD	CPME	CPMF	FLAGS	MIN
5	145	2.00	0	0	0	0	0	0	E11	322
5	146	2.00	0	8	0	4	0	0	E12	325
5	147	2.00	4	0	0	6	13	0	E13	327
5	148	2.00	4	0	0	0	0	0	S1	329
5	149	2.00	0	4	0	2	0	0	S2	331
5	150	2.00	5	1	0	0	8	0	F1	333
5	151	2.00	0	0	0	1	0	0	N1	335
5	152	2.00	4	0	0	0	2	0	N2	338
5	153	2.00	0	9	0	0	6	0	F2	340
5	154	2.00	0	2	0	6	0	0	F3	342
5	155	2.00	1	1	0	2	4	0	F4	344
5	156	2.00	0	0	3	3	0	0	F5	346
5	157	2.00	8	5	2	0	0	4	F6	348
5	158	2.00	0	5	0	0	3	0	F7	350
5	159	2.00	5	5	1	0	0	0	F8	352
5	160	2.00	2	7	0	2	0	0	F9	355
5	161	2.00	0	6	0	0	0	0	F10	357
5	162	2.00	0	0	0	7	0	0	F11	359
5	163	2.00	0	1	0	0	0	0	F12	381
5	164	2.00	0	1	0	2	1	0	F13	383
5	165	2.00	3	4	0	1	0	1	S1	385
5	166	2.00	0	0	5	0	1	0	S2	388
5	167	2.00	0	12	0	1	8	1	G1	370
5	168	2.00	0	3	0	6	0	0	W1	372
5	169	2.00	5	5	0	2	0	0	N1	374
5	170	2.00	0	0	2	0	0	0	N2	376
5	171	2.00	4	1	0	0	8	0	W2	378
5	172	2.00	1	8	0	8	0	0	G2	381
5	173	2.00	0	8	0	2	8	2	W1	383
5	174	2.00	1	0	0	0	0	0	W2	385
5	175	2.00	0	2	0	4	0	0	G3	387
5	176	2.00	0	4	1	0	0	0	W1	389
5	177	2.00	6	6	1	0	0	0	W2	391
5	178	2.00	0	0	5	0	0	0	G4	393
5	179	2.00	0	0	3	7	0	0	W1	396
5	180	2.00	0	2	0	0	9	0	W2	398
5	181	2.00	0	0	0	1	5	1	G5	400
5	182	2.00	1	0	0	4	0	1	W1	402
5	183	2.00	0	3	0	9	1	0	W2	404
5	184	2.00	5	6	0	4	3	0	G6	406
5	185	2.00	0	1	0	1	1	12	W1	408
5	186	2.00	0	0	0	3	0	0	W2	411
5	187	2.00	6	0	0	0	0	0	G7	413
5	188	2.00	4	5	0	0	0	0	W1	415
5	189	2.00	6	3	0	6	2	0	W2	417
5	190	2.00	3	0	0	1	0	0	G8	419
5	191	2.00	0	1	0	0	0	0	W1	421
5	192	2.00	1	0	0	4	10	0	W2	424
5	193	2.00	0	2	0	11	0	4	G9	426
5	194	2.00	2	0	0	5	2	0	W1	428
5	195	2.00	0	7	0	3	11	0	W2	430

5 196	2.00	0	6	2	0	0	5 610	432
5 197	2.00	0	0	0	2	5	0 W1	434
5 198	2.00	0	1	0	0	0	0 W2	436
5 199	2.00	0	2	0	0	2	0 611	439
5 200	2.00	1	0	0	6	0	0 W1	441
5 201	2.00	2	3	0	7	0	0 W2	443
5 202	2.00	1	3	0	7	1	0 612	445
5 203	2.00	2	10	0	12	0	0 W1	447
5 204	2.00	1	13	11	0	0	0 W2	449
5 205	2.00	6	2	0	5	1	0 613	452
5 206	2.00	0	5	0	0	0	2 S1	454
5 207	2.00	0	0	0	3	3	0 S2	456
5 208	2.00	1	0	10	1	3	0 W1	458
5 209	2.00	1	3	0	0	1	0 W2	460



Time: 2.00

Data Mode: Dual DPM

Nuclides: 3H-14C-UG

Quench Sets

Background Subtract: 1st Vial

Low Energy: 3H-UG

High Energy: 14C-U

	LL	UL	LCR	2S%	BKG
Region A:	0.0 - 12.0		0	0.0	12.70
Region B:	12.0 - 156		0	0.0	14.80
Region C:	156 - 2000		0	0.0	13.20

Quench Indicator: tSIE/AEC

Ext Std Terminator: Count

Coincidence Time(ns): 18

Delay Before Burst(ns): Normal

S#	TIME	CPMA	A:2S%	CPMB	B:2S%	CPMC	DPM1	DPM2	tSIE	FLAG
1	10.00	12.70	17.75	14.80	16.44	13.20			693.81	B
2	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	598.56	A1
3	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	615.89	N1
4	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	597.20	N2
5	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	606.70	E1
6	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	620.66	E2
7	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	568.76	A2
8	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	579.04	E1
9	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	587.37	E2
10	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	592.88	A3
11	2.00	0.00	0.00	0.04	16696	0.00	0.00	0.04	594.32	E1
12	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	581.84	E2
13	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	531.13	S1
14	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	577.77	S2
15	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	614.45	A4
16	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	599.50	N1
17	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	602.22	N2
18	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	616.67	E1
19	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	578.91	E2
20	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	613.17	A5
21	2.00	13.48	56.23	0.00	0.00	0.00	30.86	0.00	603.28	E1
22	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	607.34	E2
23	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	591.49	A6
24	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	625.81	E1
25	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	606.82	E2
26	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	613.79	A7
27	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	633.70	E1
28	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	600.31	E2
29	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	606.09	A8
30	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	630.42	E1
31	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	620.49	E2
32	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	593.63	S2
33	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	627.84	S1
34	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	588.34	A9
35	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	553.26	N1
36	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	570.85	N2
37	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	603.68	E1
38	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	602.13	E2
39	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	541.85	A10
40	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	566.86	E1

ENCL

S#	TIME	CPMA	A:2S%	CPMB	B:2S%	CPMC	DPM1	DPM2	tSIE	FLAG
41	2.00	0.00	0.00	2.20	287.1	0.00	0.00	2.72	577.67	A10 E2
42	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	561.37	A11
43	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	582.34	E1
44	2.00	0.00	0.00	6.61	105.7	0.00	0.00	8.16	595.40	E2
45	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	557.10	A12
46	2.00	0.80	708.9	0.00	0.00	0.00	1.83	0.00	602.73	E1
47	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	616.02	E2
48	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	557.33	A12
49	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	559.99	E1
50	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	631.21	E2
51	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	575.11	S1
52	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	598.51	S2
53	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	600.48	B1
54	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	606.88	N1
55	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	595.23	N2
56	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	562.56	B2
57	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	549.72	B2
58	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	535.59	E1
59	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	563.26	E2
60	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	599.95	B4
61	2.00	0.00	0.00	0.20	2981	0.00	0.00	0.25	588.41	N1
62	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	584.17	N2
63	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	585.88	W1
64	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	586.61	W2
65	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	602.64	B3
66	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	600.97	W1
67	2.00	0.00	0.00	0.00	0.00	0.80	0.00	0.00	602.68	W2
68	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	584.86	B6
69	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	619.14	W1
70	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	611.58	W2
71	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	622.30	B7
72	2.00	0.00	0.00	1.49	416.3	0.00	0.00	1.84	606.10	W1
73	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	593.75	W2
74	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	603.60	B8
75	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	589.28	S1
76	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	609.58	S2
77	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	601.58	W1
78	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	607.97	W2
79	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	528.48	B9
80	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	576.21	N1
81	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	593.02	N2
82	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	565.15	E10
83	2.00	16.36	48.60	0.00	0.00	0.00	39.54	0.00	545.81	B11
84	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	558.76	B12
85	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	572.03	B12
86	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	565.08	S1
87	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	607.02	S2
88	2.00	3.30	184.6	0.00	0.00	0.00	7.75	0.00	575.75	C1
89	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	557.28	N1
90	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	603.73	N2
91	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	555.81	C2
92	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	566.75	C3
93	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	576.75	E1
94	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	621.18	E2
95	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	559.93	C4
96	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	554.90	E1

[illegible]

S#	TIME	CPMA	A:2S%	CPMB	B:2S%	CPMC	DPM1	DPM2	tSIE	FLAG
153	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	580.36	F1 N2
154	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	549.70	F2
155	2.00	0.00	0.00	1.16	531.7	0.00	0.00	1.43	552.62	F3
156	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	537.70	F4
157	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	563.33	F5
158	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	537.18	F6
159	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	551.73	F7
160	2.00	0.00	0.00	0.20	2981	0.00	0.00	0.25	552.90	F8
161	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	563.57	F9
162	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	556.53	F10
163	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	547.94	F11
164	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	576.59	F12
165	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	561.05	F13
166	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	614.93	S1
167	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	626.63	S2
168	2.00	0.00	0.00	0.20	2981	0.00	0.00	0.25	588.28	G1
169	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	585.45	W1
170	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	589.53	W1
171	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	594.11	N2
172	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	583.88	W2
173	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	544.49	G2
174	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	572.30	W1
175	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	586.18	W2
176	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	565.51	G3
177	2.00	0.00	0.00	2.20	287.1	0.00	0.00	2.72	564.77	U1
178	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	587.63	W2
179	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	559.70	G4
180	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	574.85	W1
181	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	602.69	W2
182	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	547.81	G5
183	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	580.13	W1
184	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	601.46	W2
185	2.00	0.00	0.00	0.00	0.00	0.80	0.00	0.00	568.00	G6
186	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	581.20	W1
187	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	592.69	W2
188	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	570.14	G7
189	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	596.82	W1
190	2.00	20.80	40.82	0.00	0.00	0.00	47.13	0.00	614.57	W2
191	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	529.68	G8
192	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	583.95	U1
193	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	587.38	W2
194	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	522.71	G9
195	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	592.88	W1
196	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	590.79	W2
197	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	585.13	G10
198	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	604.26	W1
199	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	570.26	W2
200	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	601.84	G11
201	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	589.08	W1
202	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	601.52	W2
203	2.00	0.00	0.00	1.20	512.7	0.00	0.00	1.49	557.11	G12
204	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	586.85	W1
205	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	603.39	W2
206	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	604.32	G13
207	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	632.77	S1
208	2.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	639.53	S2

PROTOCOL #. J

ROUTINE ..

S#	TIME	CPMA A:2S%	CPMB B:2S%	CPMC	DPM1	DPM2	tSIE	FLAG
209	2.00	0.00 0.00	0.00 0.00	0.00	0.00	0.00	605.72	613 W1
210	2.00	0.00 0.00	0.00 0.00	0.00	0.00	0.00	597.27	W2

SYSTEM NORMALIZED

C14 IPA DATA PROCESSED - 28-Apr-97 08:08

C14 Eff (0-156 keV) = 96.39 %

H3 IPA DATA PROCESSED - 28-Apr-97 08:09

H3 Eff (0-18.6 keV) = 64.00 %

WARNING: Questionable H3 Efficiency value - Please rerun quench curves  
& view historic data