

Michael G. Herr, CHP  
Materials License Branch  
United States Nuclear Regulatory Commission  
Region III  
2443 Warrenville Road Suite 210  
Lisle, IL 60532-4352

06/20/2011

Michael

Reference mail control # 574801

This is confirmation that I have never had a leaking source while licensed by NRC license 48-32333-01. Attached are most recent leak test for the two gauges covered by my license. Leaks test have been performed at 6 month intervals if gauges were actively being used.


Wisconsin License  
License Number 097-1136-01

Old Address on license  
1800 Blueberry Dr.  
Plover, WI 54467

**New address**  
**10513 Boulder Lane**  
**Boulder Junction, WI 54512**

Any questions please let me know.

Joe Keyzers  
RSO  
Metso Fabrics  
10513 Boulder Lane  
Boulder Junction, WI 54512



RECEIVED JUN 24 2011

NRC FORM 314  
(12-2010)  
10 CFR 30.36(j)(1); 40.42(j)(1);  
70.38(j)(1); and 72.54(k)(5)(1)(1)

U.S. NUCLEAR REGULATORY COMMISSION

APPROVED BY OMB: NO. 3150-0028

EXPIRES: 10/31/2013

### CERTIFICATE OF DISPOSITION OF MATERIALS

Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the Information Services Branch (T-5 F53), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to [Infocollects.Resource@nrc.gov](mailto:Infocollects.Resource@nrc.gov), and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

LICENSEE NAME AND ADDRESS

Keyzers Joseph M.  
1800 Blueberry Dr.  
Plover, WI 54467

LICENSE NUMBER

48-32333-01

DOCKET NUMBER

0303577

LICENSE EXPIRATION DATE

07/31/11

#### A. LICENSE STATUS (Check the appropriate box)

- This license has expired.  This license has not yet expired; please terminate it.

#### B. DISPOSAL OF RADIOACTIVE MATERIAL

(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

1. No radioactive materials have ever been procured or possessed by the licensee under this license.
2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner.
- a. Transfer of radioactive materials to the licensee listed below:  
Joseph M. Keyzers Transferred to State of Wisconsin License # 097-1136-01 Have License in State of WI since 7/7/06
- b. Disposal of radioactive materials:
1. Directly by the licensee:
2. By licensed disposal site:
3. By waste contractor:
- c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

#### C. SURVEYS PERFORMED AND REPORTED

1. A radiation survey was conducted by the licensee. The survey confirms:
- a. the absence of licensed radioactive materials
- b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.
2. A copy of the radiation survey results:
- a. is attached; or  b. is not attached (Provide explanation); or  c. was forwarded to NRC on: \_\_\_\_\_ Date
3. A radiation survey is not required as only sealed sources were ever possessed under this license, and
- a. The results of the latest leak test are attached; and/or  b. No leaking sources have ever been identified.

The person to be contacted regarding the information provided on this form:

NAME	TITLE	TELEPHONE (Include Area Code)	E-MAIL ADDRESS
Joseph M Keyzers	Owner Keyzers Enterprise LLC	715-892-7724	joe.keyzers@metso.com

Mail all future correspondence regarding this license to:  
10513 Boulder Lane, Boulder Junction, WI 54512

#### C. CERTIFYING OFFICIAL

I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE

Joseph M. Keyzers

SIGNATURE

*Joseph M Keyzers*

DATE

6/20/11

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

Send all kits and requests to:

**SUNTRAC Services, Inc.**  
**1818 East Main Street**  
**League City, TX 77573**  
**(281) 338-2133**

ATTN: SIT-KIT

---

CAUTION: Conduct a survey on the outside of each package placed in the U.S. Mails. Any reading over 0.5 mR/hr at contact with the envelope or package shall not be mailed.

---

**LEAK TEST INVENTORY/REPORT FORM**

Company Name: Tamfelt Inc.

Address: 1800 Blueberry Dr. City: Plover State: WI Zip Code: 54467

Isotope: Am 241 Activity: 80 mCi

Source Serial No: N/A Leak Test Date: 04/01/11

Device Serial No: 1271

Manufacturer: Not Provided Model No.: 104P

Smear Taken By: Seth Ponschock


----- DO NOT WRITE BELOW THIS LINE -----

This is to certify that the above described smear/swab has been assayed at our facilities for indication of source leakage.

Our findings show the leakage to be:

ALPHA	BETA-GAMMA
<.0001	<.0001

 $\mu\text{Ci (Wet)}$ 

Certified By:  Date: 04/05/11

Jessica Ochoa

**Register Number: 1636911**

Send all kits and requests to:

**SUNTRAC Services, Inc.**  
**1818 East Main Street**  
**League City, TX 77573**  
**(281) 338-2133**

ATTN: SIT-KIT

CAUTION: Conduct a survey on the outside of each package placed in the U.S. Mails. Any reading over 0.5 mR/hr at contact with the envelope or package shall not be mailed.

**LEAK TEST INVENTORY/REPORT FORM**

Company Name: Tamfelt Inc. (12975)

Address: 1800 Blueberry Dr. City: Plover State: WI Zip Code: 54467

Isotope: Am 241 Activity: 25 mCi

Source Serial No: NDC 104P Leak Test Date: 12/27/10

Device Serial No: 11891

Manufacturer: NDC System Model No.: 104P

Smear Taken By: Roy Probandt

----- DO NOT WRITE BELOW THIS LINE -----

This is to certify that the above described smear/swab has been assayed at our facilities for indication of source leakage.

Our findings show the leakage to be:

ALPHA	BETA-GAMMA
<.0001	<.0001

µCi (Wet)

Certified By:  Date: 01/03/11


SUNTRAC Services, Inc. Representative  
(Texas Radioactive License No. L03062)

1636911

Joe Keyzers  
16513 Boulder Lane  
Boulder Junction, WI 54912

CENTRAL WI P&DF 544

21 JUN 2011 PM 1 T



Michael Her CHP  
Motorist License Branch  
US - NRC Region III  
2443 Warrenville Rd. Suite 210  
Lisle, IL 60532-4352

Mail control # 574801

60532+4352 