



Topic 4

Additional Molybdenum Breakthrough Testing and Reporting Requirements

Neelam Bhalla

**Office of Federal and State Materials and Environmental
Management Programs**

June 21, 2011

Acronyms

- **IN - Information Notice**
- **Mo-99 - Molybdenum-99**

Background

- **Currently, only 1st eluate is required to be tested**
- **Prior to 2002, *each* eluate required testing**
- **Mo-99 breakthrough was considered to be a rare event**
- **Manufacturers recommend testing *each* eluate.**

Background

- **Licensees reported failures for both the 1st elution, and in some cases, subsequent elutions**
- **In response, NRC issued IN 2008-22 “Molybdenum-99 Breakthrough in Molybdenum-99m/Technetium-99m Generators”**

Topic 4 Discussion Point

- **The NRC is considering amending Part 35 to:**
 - **Require Mo-99 testing of each eluate**
 - **Adding reporting requirement if the regulatory limit is exceeded**

Reporting of a Failed Test

- **If reporting requirements are adopted, how soon after discovery of a failed test should the NRC be notified?**