

MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

03620

575387

<p>Licensee</p> <p>1. Ferris State University Environmental Health and Safety Office</p> <p>2. 111 W. Knollview Drive Big Rapids, MI 49307-2742</p>	<p>In accordance with the letter dated June 14, 2011,</p> <p>3. License number 21-15237-01 is amended in its entirety to read as follows:</p> <hr/> <p>4. Expiration date April 30, 2015</p> <hr/> <p>5. Docket No. 030-08783 Reference No.</p>
---	--

6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form	8. Maximum amount that licensee may possess at any one time under this license
A. Barium-133	A. Sealed sources	A. 1 millicurie
B. Cadmium-109	B. Sealed sources	B. 1 millicurie
C. Manganese-54	C. Sealed sources	C. 1 millicurie
D. Cobalt-60	D. Sealed sources	D. 20 microcuries
E. Europium-152	E. Sealed sources	E. 20 microcuries
F. Iodine-129	F. Sealed sources	F. 20 microcuries
G. Cesium-137	G. Sealed sources	G. 10 millicuries
H. Molybdenum 99/Technetium 99m	H. Technetium 99m Generators	H. 1 curie
I. Cesium-137	I. Sealed sources (Troxler Dwg. No. A-102112)	I. 20 millicuries
J. Americium-241	J. Sealed sources (Troxler Dwg. No. A-102451)	J. 100 millicuries
K. Cobalt-60	K. Any	K. Not to exceed a total of 40 microcuries
L. Sodium-22	L. Sealed sources (New England Nuclear, North American Scientific Model MED 3400; Isotope Products Labs GF-290)	L. Not to exceed a total of 1 millicurie

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License Number
21-15237-01

Docket or Reference Number
030-08783

Amendment No. 36

- | | | |
|--|---|---|
| <p>6. Byproduct, source, and/or special nuclear material</p> <p>M. Cobalt-57</p> | <p>7. Chemical and/or physical form</p> <p>M. Sealed sources (New England Nuclear, Dupont Models CR-164C, 486E, NES 137T; North American Scientific models MED 3400 & MED 3504; Isotope Products Labs Model GF-290)</p> | <p>8. Maximum amount that licensee may possess at any one time under this license</p> <p>M. Not to exceed a total of 1 millicurie</p> |
|--|---|---|

9. Authorized Use:

- A. through H, L. and M. To be used for student instruction and instrument calibration.
- I. and J. To be used in Troxler Model 3440 gauging devices for student instruction in measuring moisture and density of construction materials.
- K. For storage only incident to disposal.

CONDITIONS

10. A. Licensed material shall be used only at the licensee's facilities located at the campus of Ferris State University, 200 Ferris Drive, Big Rapids, Michigan; 220 Ferris Drive, Big Rapids, Michigan; 1020 East Maple Street, Big Rapids, Michigan; and Ferris State University – Grand Rapids, 151 Fountain Street NE, Grand Rapids, Michigan.
- B. Licensed material listed in Subitems 6.I. and 6.J. may be used at temporary job sites of the licensee anywhere in the United States where the U.S. Nuclear Regulatory Commission maintains jurisdiction for regulating the use of licensed material.
11. A. Licensed material listed in Subitems 6.A. through 6.H. shall be used by, or under the supervision of Sheila MacEachron, Timothy Vander Laan and Tracy L. Glentz.
- B. Licensed material listed in Subitems 6.I. through 6.J. shall be used by, or under the supervision and in the physical presence of **Thomas C. Larabel** or other individuals who have successfully completed one of the training courses described in the section entitled "Training for Individuals Working In or Frequenting Restricted Areas in NUREG-1556, Vol. 1, Rev. 1, dated November 2001.
- C. The Radiation Safety Officer for this license is Brad McCormick.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License Number

21-15237-01

Docket or Reference Number

030-08783

Amendment No. 36

12. A. Sealed sources shall be tested for leakage and/or contamination at intervals not to exceed the intervals specified in the certificate of registration issued by NRC under 10 CFR 32.210 or by an Agreement State.
- B. In the absence of a certificate from a transferor indicating that a leak test has been made within the intervals specified in the certificate of registration issued by NRC under 10 CFR 32.210 or by an Agreement State prior to the transfer, a sealed source received from another person shall not be put into use until tested.
- C. Sealed sources need not be tested if they are in storage and are not being used. However, when they are removed from storage for use or transferred to another person, and have not been tested within the required leak test interval, they shall be tested before use or transfer. No sealed source shall be stored for a period of more than 10 years without being tested for leakage and/or contamination.
- D. The leak test shall be capable of detecting the presence of 0.005 microcurie (185 becquerels) of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie (185 becquerels) or more of removable contamination, a report shall be filed with the U.S. Nuclear Regulatory Commission in accordance with 10 CFR 30.50(c)(2), and the source shall be removed immediately from service and decontaminated, repaired, or disposed of in accordance with Commission regulations.
- E. Tests for leakage and/or contamination shall be performed by persons specifically licensed by the Commission or an Agreement State to perform such services. In addition, the licensee is authorized to collect leak test samples but not perform the analysis: analysis of leak samples must be performed by persons specifically licensed by the Commission or an Agreement State to perform such services.
13. Sealed sources or source rods containing licensed material shall not be opened or sources removed or detached from source rods or gauges by the licensee, except as specifically authorized.
14. Maintenance, repair, cleaning, replacement, and disposal of foils contained in detector cells shall be performed only by the device manufacturer or other persons specifically authorized by the Commission or an Agreement State to perform such services.
15. The licensee is authorized to transport licensed material only in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."
16. When performing tests at temporary job sites, the authorized user shall not leave the moisture/density gauge unattended. Upon completion of tests the device shall be locked in the licensee's vehicle or a secure building to prevent unauthorized use, loss, or theft.
17. Except for maintaining labeling as required by 10 CFR Part 20 or 71, the licensee shall obtain authorization from NRC before making any changes in the sealed source, device, or source-device combination that would alter the description or specifications as indicated in the respective Certificates of Registration issued either by the Commission pursuant to 10 CFR 32.210 or by an Agreement State.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**License Number
21-15237-01Docket or Reference Number
030-08783

Amendment No. 36

18. Each portable nuclear gauge shall have a lock or outer locked container designed to prevent unauthorized or accidental removal of the sealed source from its shielded position. The gauge or its container must be locked when in transport. A minimum of two independent physical controls that form tangible barriers to secure portable gauges from unauthorized removal whenever the portable gauge is not under the control and constant surveillance of the licensee are required.
19. The licensee shall conduct a physical inventory every 6 months, or at other intervals approved by NRC, to account for all sources and/or devices received and possessed under the license.
20. Any cleaning, maintenance, or repair of the gauges that requires detaching the source or source rod from the gauge shall be performed only by the manufacturer or other persons specifically licensed by the Commission or an Agreement State to perform such services.
21. A. If the licensee uses unshielded sealed sources extended more than 3 feet below the surface, the licensee shall use surface casing that extends from the lowest depth to 12 inches above the surface and other appropriate procedures to reduce the probability of the source or probe becoming lodged below the surface. If it is not feasible to extend the casing 12 inches above the surface, the licensee shall implement procedures to ensure that the cased hole is free of obstruction before making measurements.
- B. If a sealed source or a probe containing sealed sources becomes lodged below the surface and it becomes apparent that efforts to recover the sealed source or probe may not be successful, the licensee shall notify the U. S. Nuclear Regulatory Commission and submit the report required by 10 CFR 30.50(b)(2) and (c). The licensee shall not abandon the sealed source or probe without obtaining the Commission's prior written consent.
22. The licensee shall develop, implement and maintain operating and emergency procedures that meets the criteria in the section entitled "Radiation Safety Program - Operating and Emergency Procedures" in NUREG-1556, Vol. 1, Rev. 1, dated November 2001.
23. The licensee is authorized to hold radioactive material with physical half-life of less than or equal to 120 days for decay-in-storage before disposal in ordinary trash provided:
- A. Before disposal as ordinary trash, the waste shall be surveyed at the container surface with the appropriate survey instrument set on its most sensitive scale and with no interposed shielding to determine that its radioactivity cannot be distinguished from background. All radiation labels shall be removed or obliterated.
- B. A record of each such disposal permitted under this license condition shall be retained for 3 years. The record must include the date of disposal, the date on which the byproduct material was placed in storage, the radionuclides disposed, the survey instrument used, the background dose rate, the dose rate measured at the surface of each waste container, and the name of the individual who performed the disposal.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**License Number
21-15237-01Docket or Reference Number
030-08783


Amendment No. 36

24. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.

A. Application dated November 12, 2004; and

B. Letters dated April 8, 2005, January 27, 2009 (excluding all references to NARM), August 24, 2010.

FOR THE U.S. NUCLEAR REGULATORY COMMISSION

Date JUN 28 2011
JUN 28 2011By 
William P. Reichhold
Materials Licensing Branch
Region III