

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

<p>1. LICENSEE/LOCATION INSPECTED: New Milford Hospital 21 Elm Street New Milford, Connecticut 06776-2915</p> <p>REPORT Nos 2011-001</p>	<p>2. NRC/REGIONAL OFFICE</p> <p>U.S. Nuclear Regulatory Commission Region I, 475 Allendale Road King of Prussia, Pennsylvania 19406-1415</p>	
<p>3. DOCKET NUMBER(S) 030-13596</p>	<p>4. LICENSE NUMBER(S) 06-17892-01</p>	<p>5. DATE(S) OF INSPECTION <i>June 6-7, 2011</i></p>

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

1. Based on the inspection findings, no violations were identified.

2. Previous violation(s) closed.

3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

10 CFR 35.633 (b)(5) requires in part that a licensee authorized to use a remote afterloader unit for medical use shall perform full calibration measurements which include a determination of the timer accuracy and linearity over the typical range of use. Contrary to the above, the licensee did not perform timer accuracy and linearity over the typical range of use. Specifically, as of June 7, 2011, timer accuracy and linearity was determined over a four minute period. However, the typical range of use is 10 minutes. Timer accuracy + linearity was performed on June 7, 2011 over a 10 minute period.

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE	Deborah Weymouth	<i>Deborah Weymouth</i>	6/7/11
NRC INSPECTOR	Janice Nguyen	<i>Janice Nguyen</i>	6-7-11
BRANCH CHIEF	Marc S. Ferdas	<i>Marc S. Ferdas</i>	6/28/11