Oakwood Heritage Hospital

10000 Telegraph Rd. Taylor, Michigan 48180-3349 313,295,5000





June 30, 2011

United States Nuclear Regulatory Commission Region III, Office of Materials Licensing 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352

RE:

Amendment to NRC License #: 21-32587-01

Millennium Medical Center South

Dear Sir/Madam:

We would like to amend our current NRC license to reflect the following changes.

Millennium Medical Center South has been purchased by Oakwood Health Care System effective July, 2011. The new name of this facility is Oakwood Imaging Center Taylor.

Therefore we wish to amend our license to reflect these changes.

- File Transfer of Control papers to reflect the "Change of Ownership".
- Change the name of NRC License #21-32587-01 to reflect this "Change of Ownership".

Thank you for your cooperation. If you have any questions or require additional information, please contact our RSO, James Botti at 734-662-3197 or by email at jbotti@mpcphysics.com

Sincerely,

Kelly Smith

COO

TRANSFER OF CONTROL PER INFORMATION APPENDIX G, NUREG-1556, VOL. 9, REV. 9

- 1. NRC License # <u>21-32587-01</u>
- 2. Current name: Millennium Medical Center South
- 3. The new name will be: Oakwood Imaging Center Taylor
- 4. The license contact name is: <u>Jamie Heldt at 313-295-6704</u>

Transfer of stocks by sale will be made. <u>Oakwood Health Care System</u> will obtain possession of all radioactive materials, sealed sources and radiation safety equipment for the following locations:

8850 South Telegraph Taylor, MI 48180

- 5. All currently approved users for the licensee will remain without change. The Radiation Safety Officer for this license will be <u>James M. Botti, MS</u>
- 6. <u>Millennium Medical Center South</u> will no longer operate once the transfer is complete. All activities will be under Oakwood Imaging Center Taylor.
- 7. All required records regarding the use and possession of licensed materials will be current at the time of the transfer. The status of compliance will be maintained before, during and after the merger. All of these records are maintained on site. NO records regarding licensed material will be removed from the licensed site.
- 8. All required documents and records are currently kept in accordance with the NRC license conditions. These include, but are not limited to the annual calibration of survey instruments, semi-annual leak testing of non-exempt sealed radioactive sources, radiation monitoring and personnel exposure monitoring reports. Radiation monitoring is performed as required by the NRC licensee on a daily basis for acceptable radiation levels and weekly for removable contamination.

- 9. We confirm the areas of storage, preparation, handling, and administration of radioactive materials are surveyed daily for acceptable radiation levels and weekly for removable contamination. If contamination of these areas is detected during these daily or weekly surveys, decontamination is undertaken and completed at that time. The unsealed radio-nuclides in use at this location are all short-lived (i.e. half lives less than 65 days, most less than 10 days). Therefore, radioactive contamination is highly unlikely at this facility. If contamination were present, it would be discovered prior to the transfer. Record will be maintained on site of inventories of sealed sources and spills of radioactive materials.
- 10. <u>Oakwood Imaging Center Taylor</u> agrees to abide by all commitments and representations previously made to the NRC by <u>Millennium Medical Center South</u>.
- 11. <u>Oakwood Imaging Center Taylor</u> accepts full responsibility for the site. This facility is exempt from providing financial assurance for decommissioning because of the type of radioactive materials used (i.e. short-lived, unsealed sources in small quantities only). Routine radiation monitoring of the facility ensures that contamination will not be transferred during any subsequent change of ownership.
- 12. All parties involved with the transfer are aware and agree to the planned transfer of ownership.

Forster, Sara

From:

Heldt, Jamie [Jamie.Heldt@oakwood.org]

Sent:

Friday, July 01, 2011 10:07 AM

To:

Forster, Sara

Subject:

FW: Scan from a Xerox WorkCentre

Attachments:

img-701072235-0001.pdf

----Original Message----

From: Heldt, Jamie

Sent: Friday, July 01, 2011 10:37 AM

To: 'Sarah.Forster@nrc.gov.'

Subject: FW: Scan from a Xerox WorkCentre

----Original Message----

From: Heldt, Jamie

Sent: Friday, July 01, 2011 7:29 AM

To: 'Sarah.Forster@nrc.gov'

Subject: FW: Scan from a Xerox WorkCentre

----Original Message----

From: Heldt, Jamie

Sent: Friday, July 01, 2011 7:28 AM

To: 'Sarah.Forster@nrc.gov.'

Subject: FW: Scan from a Xerox WorkCentre

----Original Message----

From: Heldt, Jamie

Sent: Friday, July 01, 2011 7:27 AM

To: 'Sara Forster at'

Subject: FW: Scan from a Xerox WorkCentre

Please amend our current NRC license to reflect the attached changes.

Thank you for all of your help in this matter!

Jamieann Heldt Director of Cardiology and Radiology Oakwood Heritage Hospital 313-295-6704