

U.S. NUCLEAR REGULATORY COMMISSION**TELEPHONE CONVERSATION RECORD**

Date: 7/12/2011

Time:

Mail Control
or Report No(s).

575395

License No(s). 09-31177-01

Docket No(s). 03037316

Name of Licensee: Ambergris, LLC
dba 21st Century Oncology

Name of Participant(s): Leandro Barreca

Telephone No. 239-768-7377

Subject: RAI

(NOTE: This will be used as the
Documents Title in ADAMS)

In a letter dated 4/11/2011, 21st Century Oncology requested a change to their current HDR unit. The new unit is the Nucletron microSelectron Model 106.990. This unit has the ability to be operated in a pulse mode. I spoke with Leandro Barreca on 7/12/11 to confirm that their intentions are NOT to use the HDR in the pulse mode. L. Barreca stated that they have no intention of using the unit in any other mode than the high dose rate mode.

Action Required: Issue the amendment

Document Availability:

Publicly Available

Non-Publicly Available

 Non-Sensitive
 Sensitive – Privacy Act/ PII
 Immediate Release Non-Sensitive Copyright
 Sensitive – Internal
Normal Release Date: 7/20/2011 Sensitive- Proprietary
 Sensitive – Security-Related
Delayed Release Date:

SUNSI Review Completed By: TLWeidner / RA /

Document Accession #: ML111930210