

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:
West Michigan Heart
709 S. Greenville West Drive,
Greenville, Michigan
904 S. Washington, Suite 102,
Holland, Michigan
6105 Wilson SW, Suite 101,
Wyoming, Michigan
REPORT NUMBER(S) 2011-001

2. NRC/REGIONAL OFFICE
Region III
2443 Warrenville Rd.
Lisle, IL 60532

3. DOCKET NUMBER(S)
030-33375

4. LICENSEE NUMBER(S)
21-26543-01

5. DATE(S) OF INSPECTION
July 13 and 14, 2011

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied

_____ Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Dennis Lawyer		7/14/2011
Branch Chief	Tamara E. Bloomer		7/28/11

Docket File Information
SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE West Michigan Heart 2900 Bradford St. N.E. Grand Rapids, MI 49525 REPORT NUMBER(S) 2011-001	2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Rd. Lisle, IL 60532
---	---

3. DOCKET NUMBER(S) 030-33375	4. LICENSE NUMBER(S) 21-26543-01	5. DATE(S) OF INSPECTION July 13 and 14, 2011
----------------------------------	-------------------------------------	--

6. INSPECTION PROCEDURES 87130	7. INSPECTION FOCUS AREAS 02.01-02.08
-----------------------------------	--

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM 2201	2. PRIORITY 5	3. LICENSEE CONTACT David Langholz, M.D., RSO	4. TELEPHONE NUMBER 616-486-6947
--------------------	------------------	--	-------------------------------------

- Main Office Inspection Next Inspection Date: July 2016
- Field Office Inspection 709 S. Greenville West Drive, Greenville, Michigan; 904 S. Washington, Suite 102, Holland, Michigan; and 6105 Wilson SW, Suite 101, Wyoming, Michigan
- Temporary Job Site Inspection

PROGRAM SCOPE

Private Medical Practice performing cardiac studies. Greenville location normally performs nuclear studies on six patients a day on Monday, Tuesday, and Wednesdays. They have one camera, one treadmill and one technician. The Holland location normally performs nuclear studies on seven patients a day, Monday through Thursday. They have one camera, one treadmill and two nuclear technicians. The Wyoming location normally performs nuclear studies on five patients a day with one camera, one treadmill and one technician. The Grand Rapids location was not inspected. The program was the same at each facility. Each facility maintained their records and nothing was expected to be different at the Grand Rapids location. A physicist visits each location quarterly and performs training, instrument checks, and an audit.

Performance Observations

The nuclear medicine technicians at each location demonstrated receipt of material, daily checks of potable instruments, well counter, and dose calibrator, daily survey, weekly wipe survey, and release of decayed material. The technicians were knowledgeable about the handling of radioactive material and described emergency procedures. The inspector observed dosimetry, waste handling, and material security methods.