

**U.S. NUCLEAR REGULATORY COMMISSION**

Date: 8/1/11

**TELEPHONE CONVERSATION RECORD**

Time:

Mail Control  
or Report No(s).

575554

License No(s). 06-06941-01

Docket No(s). 03001267

Name of Licensee: Norwalk Hospital

Name of Participant(s): Balan Kodery, RSO

Telephone No. 203-852-2783

Subject: Amendment not needed

(NOTE: This will be used as the  
Documents Title in ADAMS)

On July 28, 2012 I spoke to B. Kodery about his amendment request to add a rubidium-82 generator (CardioGen-82). According to an FDA correspondence dated 7/26/11, the CardioGen-82 generator is being voluntarily recalled by the manufacturer, Bracco. At that time B. Kodery stated that he would like to withdraw his amendment request. After further discussions with RI staff, it was determined that an amendment was not necessary because Rb-82 falls under 35.200 for imaging and localization. On 8.1.11 informed B. Kodery that an amendment was not necessary because their license currently authorizes 35.200 use at the PET facility at 148 East Avenue, Norwalk, CT.

**Action Required:** Send a letter stating that an amendment is not needed.

Document Availability:

Publicly Available

Non-Publicly Available

Non-Sensitive

Non-Sensitive Copyright

Sensitive- Proprietary

 Sensitive – Privacy Act/ PII

Sensitive – Internal

 Immediate Release

Normal Release Date: 8/9/2011

Sensitive – Security-Related

Delayed Release Date:

SUNSI Review Completed By: TLWeidner

/ RA /

Document Accession #: ML112130448