

Manufacturer/Model: KEITHLEY / 614 **Description:** ELECTROMETER Serial Number: 0555368 Asset Number: 002792 Procedure: KEITHLEY 614 - 21 MAR 09

Date Calibrated: 17-Jan-2011 * Date Due : 17-Jan-2012 ** Results: FOUND-LEFT Temperature: 74.0 °F Humidity: 40 %RH

Calibration Laboratory

Certificate #0972-01

Barometer: N/A

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U.S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Data type found in this certificate or attached measurement report must be interpreted as: Found-left - adjustment and/or repair was not performed, As-found - data is before unit is adjusted and/or repaired, As-left - data is after adjusted and/or repaired was performed. The customer has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance.

Measurement uncertainty calculated in accordance with the method described in the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM), for a confidence level of approximately 95 percent using a coverage factor of k=2.

Remarks: nC Not Calibrated.

Standards Used

<u>Asset #</u>	<u>Manufacturer</u>	Model	Description	<u>Cal Date</u>	Due Date
000101	BIDDLE	72-6346-1	DECADE RESISTOR	30-Aug-2010	30-Aug-2011
000182	FLUKE	5700A/EP	MULT-FUNCTION CALIBRATOR	3-Nov-2010	3-Feb-2011
000201	FLUKE	5725A	MULT-FUNCTION CALIBRATOR AMPLIFIE	3-Nov-2010	3-Nov-2011
001022	FLUKE	742A-10K	STANDARD RESISTOR	5-May-2010	5-May-2011
014917	FLUKE	742A-10M	STANDARD RESISTOR	21-Sep-2010	21-Sep-2011

<u>e</u>ú Ware

Laboratory Manager \mettracksvr\metcal\A2LA OCT_08.rpt

Bob Trollinger Metrology Technician

Southwest Research Institute Calibration Laboratory Measurement Report

Work Order: Asset No.	303099182 002792	Mfr. Model	KEITHLEY 614		Technician Type Data:	blt Found-	eft					
Bemarka: nC not calib	USSS308	ype	Electrometer			17-Jan-	• 1 [
Function/Range	Test Point	TI Reading	Difference	± Limits	± Uncertainty	Result	% Limit					
Zero	Volts	Volts	Volts	Volts	Volts							
0.2 V	0.19000	0.18999	-0.00001	0.00017	0.000010	Pass	6%					
	-0.19000	-0.18998	0.00002	0.00017	0.000010	Pass	12%					
2 Volt	1.9000	1.9000	0.0000	0.0016	0.00010	Pass	0%					
	-1.9000	-1.8998	0.0002	0.0016	0.00010	Pass	12%					
20 Volt	19.000	18.997	-0.003	0.016	0.0010	Pass	19%					
	-19.000	-18.995	0.005	0.016	0.0010	Pass	31%					
DC Amps	µAmps	µAmps	µAmps	µAmps	µAmps							
200 uAmp	190.0	190.0	0.0	0.7	0.055	Pass	0%					
	nAmps	nAmps	nAmps	nAmps	nAmps							
200 nAmp	190.0	190.1	0.1	1.0	0.055	Pass	10%					
	pAmps	pAmps	pAmps	pAmps	pAmps							
2000 pAmp	1900	1896	-4	29	0.50	Pass	14%					
Resistance	kohm	kohm	kohm	kohm	kohm							
20 kohm	19.00	18.99	-0.01	0.11	0.0058	Pass	9%					
200 kohm	190.0	189.9	-0.1	1.1	0.058	Pass	9%					
	Mohm	Mohm	Mohm	Mohm	Mohm							
20 Mohm	10.00	10.01	0.01	0.10	0.0058	Pass	10%					
	Gohm	Gohm	Gohm	Gohm	Gohm							
20 Gohm	9.97	9.95	-0.02	0.04	0.0058	Pass	47%					
		END	OF REPORT									