SOUTHWEST RESEARCH INSTITUTE®
6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834

Certificate of Calibration

A2LA ACCREDITED
Calibration Laboratory
Certificate #0972-01

Cost Center / Customer: DIV20 / DON BANNON
Mail Stop: B57
Manufacturer/Model: SWRI / 1000 OHM
Description: RESISTOR BOX
Serial Number: 171001
Asset Number: 003096
Procedure: RESISTORS 3458A METHOD - 26 FEB 09

Work Order: 303101953
Date Issued: 13-Jun-2011
Date Calibrated: 13-Jun-2011
*** Date Due :** 13-Dec-2011
**** Results:** FOUND-LEFT
Temperature: 72.0 °F
Humidity: 45 %RH
Barometer: N/A

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Data type found in this certificate or attached measurement report must be interpreted as: Found-left - adjustment and/or repair was not performed, As-found - data is before unit is adjusted and/or repaired, As-left - data is after adjusted and/or repaired was performed. The customer has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance.

Measurement uncertainty calculated in accordance with the method described in the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM), for a confidence level of approximately 95 percent using a coverage factor of $k=2$.

Remarks: ± 1.0%

Standards Used

Asset #	Manufacturer	Model	Description	Cal Date	Due Date
015482	HEWLETT-PACKARD	3458A/OPT 002	MULTIMETER	4-May-2011	4-May-2012

120 --- Q2011052700G6
Instrument calibration record for SwRI
Resistor Box, Model # 1000 Ohm, Serial # 171001, Asset No.: 003096 (06/13/2011)

Walt Hill
Laboratory Manager

Clint Rowe
Metrology Technician

m:\\A2LA OCT_08.rpt

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order:	303101953	Mfr:	SWRI	Technician:	CER
Asset No:	003096	Model:	1000 Ohm	Type Data:	Found-left
Serial No:	171001	Type:	Resistor Box	Cal Date:	13-Jun-11
Remarks: Limits of ± 1.0 % per customer request.					

Function/Range	Test Point	TI Reading	Difference	+/- Limit	+/- Uncertainty	Result	% Limit
Resistance	Ω	Ω	Ω	Ω	Ω		
Green to White	1000.00	1000.48	0.48	10	0.016	Pass	5%

END OF REPORT