TELECON & FAX TRANSMITTAL

U.S.NRC		TO: <u>file</u>		
United States Nuclear Regulatory Communication Protecting People and the Environment		COMPANY: _	N/A, see below .	
NUCLEAR REGULATORY COMMIS	SSION			
REGION III		# PAGES: <u>N/A</u>	EMAIL: N/A	
2443 WARRENVILLE ROAD				
LISLE, ILLINOIS 60532-4351				
(630) 829-9892 FAX: (630) 515-1	078			
CONVERSATION RECORD		TIME	IDATE	
		1:30 pm	6/14/2011	
NAME OF PERSON(S) CONTACTED	TELEPHONE NO.	ORGANIZATION		
James Botti, Physicist Consultant	(810) 599-8817	MPC Physics		
REPRESENTED PERSON or PERSONS		ORGANIZATION		
Allan Fraiberg, Radiation Safety Officer		Providence Hospital		
SUBJECT				
ILicense No.: 21-02802-03		Control No : 5	7/858	

SUMMARY

We have conducted a preliminary review your <u>license amendment request</u> dated March 28, 2011, and find that we are unable to continue this action until we have received information regarding the following:

- In our discussion surrounding the most recent deficiency request, you have indicated that not all injections of PET isotopes will be conducted in the mobile trailer. Please provide a statement indicating that the limits stated in 10 CFR Part 20 will be adhered to. RESPONSE: In a response dated June 15, 2011, the licensee confirmed that the limits specified in 10 CFR Part 20 will be adhered to. No additional information is required.
- In our conversation surrounding the most recent deficiency request, you indicated that the
 mobile trailer will be limited to the use of materials authorized under 10 CFR 35.200. Please
 provide a written statement indicating the same.
 RESPONSE: In response dated June 15, 2011, the licensee confirmed that materials
 used in the trailer will be limited to those permitted by 10 CFR 35.200. No additional
 information is required.
- 3. As discussed in our phone conversation, we have determined that the March 28, 2011, amendment request adequately describes the shielding. However, it was noted that, for future submissions, PET shielding analysis materials should be streamlined, to better convey relevant information. The request for a summary of shielding results was withdrawn. RESPONSE: Response is adequate; no additional information is required.

No additional information is required at this time. Please reference the Control No. 574858, listed at the top of this memo, with any additional correspondence, relating to this action. Please direct any questions you have to me at **(630) 829-9892.**

NAME OF PERSON DOCUMENTING CONVERSATION	SIGNATURE	DATE	
Sara A.B. Forster	Ama a.B.	Froter 4/15/2011	
•			//
	Jry	& Sinerian H 6	[1]],