

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED:

CzarTech Analytical, Inc.  
42910 W. Ten mile road  
Complex A-6  
Novi, MI 48375  
REPORT NUMBER(S): 11-01

2. NRC/REGIONAL OFFICE

U.S. Nuclear Regulatory Commission, Region III  
2443 Warrenville Road, Suite 210  
Lisle, Illinois 60532

3. DOCKET NUMBER(S)

030-35697

4. LICENSEE NUMBER(S)

21-26093-02

5. DATE(S) OF INSPECTION

07/29/2011

**LICENSEE:**

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied

\_\_\_\_\_ Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11

**Statement of Corrective Actions**

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Bob Hays Bill C. Lin		8/2/11
Branch Chief	Tamara Bloomer		8/3/11

*Docket File Information*  
**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

<b>1. LICENSEE</b> CzarTech Analytical, Inc. 42910 W. Ten mile road Complex A-6 Novi, MI 48375	<b>2. NRC/REGIONAL OFFICE</b> U.S. Nuclear Regulatory Commission, Region III 2443 Warrenville Road, Suite 210 Lisle, Illinois 60532
--	--

REPORT NUMBER(S) 11-01

<b>3. DOCKET NUMBER(S)</b> 030-35697	<b>4. LICENSEE NUMBER(S)</b> 21-26093-02	<b>5. DATE(S) OF INSPECTION</b> 07/29/2011
---	---	---

<b>6. INSPECTION PROCEDURES</b> 87126	<b>7. INSPECTION FOCUS AREAS</b> 3.01-3.07
--	---

**SUPPLEMENTAL INSPECTION INFORMATION**

<b>1. PROGRAM</b> 03123	<b>2. PRIORITY</b> T	<b>3. LICENSEE CONTACT</b> Bruce Czarnecki	<b>4. TELEPHONE NUMBER</b> 248-348-2300
----------------------------	-------------------------	---	--

- Main Office Inspection
- Field Office Inspection \_\_\_\_\_
- Temporary Job Site Inspection \_\_\_\_\_

Next Inspection Date: 07/2016

**PROGRAM SCOPE**

The licensee was a small laboratory with one gas chromatograph (GC) and two employees at Novi Michigan. The licensee uses a Hewlett-Packard 5790A Series GC for sample analysis. The licensee was a telephone contact priority code.

**Performance Observations**

The licensee was a telephone contact priority. The inspector conducted a telephone interview on 06/21/2011. After the initial interview, the inspector was unable to reach the licensee to obtain the necessary leak tests and inventory records for review. The inspector was concerned that the materials had been abandoned; therefore, the inspector included CzarTech Analytical Inc. as part of an inspection trip. The inspectors were able to inspect the licensee and gain access to the necessary leak tests and inventory records for review. No violation was noted.