

**U.S. NUCLEAR REGULATORY COMMISSION****TELEPHONE CONVERSATION RECORD**

Date: 8/8/2011

Time: 1:25 PM

Mail Control  
or Report No(s).

575679

License No(s).

08-30577-01

Docket No(s).

03035409

Name of Licensee: Medstar Georgetown Medical Center, Inc.

Name of Participant(s): Maxwell Amurao, RSO  
Dennis Lawyer, DNMS

Telephone No. 202-444-4049

Subject: RAI, Response to Question 1

(NOTE: This will be used as the  
Documents Title in ADAMS)

I discussed with Dr. Amurao that the point of question 1 of my July 28, 2011 electronic mail. I asked if he had performed decontamination of the hot spots. He stated that he had used Radiac wash, however, he didn't have instrumentation to show if he was effective in removing the spots or not. Also the decontamination was after it was identified by the vendor. We discussed different methods to determine if he was effective in decontamination and he was going to try to decontaminate again and measure the effectiveness.

**Action Required: Licensee will communicate the results of decontamination attempt.**

Document Availability:

 Publicly Available Non-Publicly Available Non-Sensitive  
 Sensitive – Privacy Act/ PII  
 Immediate Release Non-Sensitive Copyright  
 Sensitive – Internal  
Normal Release Date: 08/16/11 Sensitive- Proprietary  
 Sensitive – Security-Related  
Delayed Release Date:SUNSI Review Completed By: Dennis Lawyer / RA /Document Accession #: ML112200555