

TELECON & FAX TRANSMITTAL



NUCLEAR REGULATORY COMMISSION
 REGION III
 2443 WARRENVILLE ROAD
 LISLE, ILLINOIS 60532-4351
 (630) 829-9892 FAX: (630) 515-1078

TO: file

COMPANY: N/A, see below

PAGES: N/A EMAIL: N/A

CONVERSATION RECORD

		TIME	DATE
		10:15 am	6/13/2011
NAME OF PERSON(S) CONTACTED	TELEPHONE NO.	ORGANIZATION	
Maury Neuweg, Consultant	(217) 786-3839	Radiation Protection Services, Ltd.	
REPRESENTED PERSON or PERSONS	ORGANIZATION		
Daniel Abodeely, Radiation Safety Officer	Des Peres Hospital		
SUBJECT			
License No.: 24-32195-01		Control No.: 574912	

SUMMARY

We have conducted a preliminary review your license amendment request dated March 18, 2011, and find that we are unable to continue this action until we have received information regarding the following:

1. In order to remove the authorization for materials authorized by 10 CFR 31.11, additional information surrounding the 31.11 use history is required. For example, please include information related to the last date materials were used and/or disposed of.
RESPONSE: Licensee contact person indicated that last such use had been several years ago. Licensee indicated, in a written statement dated June 14, 2011, that 10 CFR 31.11 procedures were discontinued prior to June 1, 2011. No further response is required.
2. Based on the information provided, Dr. Cizek's RSO training is limited to 35.100, 35.200, and 35.300 (limited to oral administration of 33 millicuries or less of sodium iodide iodine-131). However, the license indicates that some Authorized Users may be involved in other types of 35.300 materials use. To be approved as RSO, 10 CFR 35.50 (e), Dr. Cizek must have training for all types of use for which the licensee seeks approval. Please indicate the types of 35.300 use being performed, and whether all 35.300 use is limited to oral administration of 33 millicuries or less of sodium iodide iodine-131.
RESPONSE: Licensee contact person indicated that 35.300 use is limited as stated above. In the written response dated June 14, 2011, licensee indicated that material use under 10 CFR 35.300 was limited to 33 mCi or less of I-131. No further response is required.

Please submit the items discussed above:

- **Statement describing last use of material under 10 CFR 31.11; and**
- **Statement specifying 35.300 limitation.**

to my attention via facsimile no. (630) 515-1078, or as a pdf file attached to an email addressed to sara.forster@nrc.gov. Reference Control No. 574912, listed at the top of this memo, with any additional correspondence, relating to this action. Please direct any questions you have to me at (630) 829-9892.

NOTE: Response received via facsimile/letter dated June 14, 2011, is adequate to fulfill the request. No further action is required.

NAME OF PERSON DOCUMENTING CONVERSATION	SIGNATURE	DATE
Sara A.B. Forster	<i>Sara A.B. Forster</i>	06/22/2011

J. Simon 7/5/11