



1980 NOV 26 PM 3 02

US NRC
Reproduction Services
Branch 50-331

IMMEDIATE ACTION LETTER
UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION III
799 ROOSEVELT ROAD
GLEN ELLYN, ILLINOIS 60137

November 21, 1980

Iowa Electric Light and Power
Company
ATTN: Mr. Duane Arnold
President
IE Towers
Post Office Box 351
Cedar Rapids, IA 52406

Gentlemen:

Based on a discussion during the Health Physics Appraisal exit meeting between Mr. Larry Root, Assistant Vice President, Iowa Electric Corporation, and A. B. Davis of this office, we understand that you will calibrate and make otherwise operable by 6:00 p.m., November 21, 1980, portable survey instruments necessary to meet emergency requirements as specified in your procedure No. PPIP-7, Revision 6, Emergency Assignment Board Tag Duties.

If our understanding of this matter as described above is not in accordance with the actual plans and actions being implemented, please contact this office immediately.

Sincerely,

James G. Keppler
James G. Keppler
Director

cc w/encl:
Mr. D. Mineck, Chief
Engineer
✓ Central Files
Reproduction Unit NRC 20b
AEOD
Resident Inspector, RIII
PDR
Local PDR
NSIC
TIC

MA 4

IMMEDIATE ACTION LETTER

8012020657