

Oakland Hills**CARDIOLOGY**

Harold Z. Friedman, M.D., F.A.C.C. • Thomas A. Mladsi, M.D., F.A.C.C.

FAXDate: 7-5-11Number of pages including cover sheet: 2

To:

James Mullamer
NRC

Phone:

Fax Phone:

623-214-5213

CC:

From:

Harold Z. Friedman, M.D.

Thomas A. Mladsi, M.D.

Julie Ebbing-Maher, N.P.

Peggy PassiakThank You!

The information contained in this facsimile is confidential and may also be physician-patient privileged. The information is intended only for the use of the individual or entity to whom it is addressed. If you are not the intended recipient, any use, dissemination, distribution, or copying of this communication is strictly prohibited. If you have received this facsimile in error, please notify this office immediately by calling us collect by telephone and return this material (all copies) to us at the address above via the U.S. Postal Service. On request, we will reimburse you for any cost of return. Thank-you.



Sealed Source Leak Test

Licensee: Oakland Hills CardiologyDate: 04/05/11Performed by: Laura Luna

Nuclide	Type	Calibration Activity	Calibration Date	Location	M/N	S/N
Cs-137	Vial	223.2 uCi	05/01/02	Hot Lab	IPL	787-84-10
Current Activity: 181.58uCi						
Ba-133	Vial	287.4 uCi	05/01/02	Hot Lab	IPL	788-63-4
Current Activity: 161.48uCi						
Co-57	Flood	10 mCi	05/01/09	Hot Lab	PF24R-057-10M	1373-052
Current Activity: 1.651mCi						

Comment: The sources listed above were leak tested using a dry wipe technique and were found to have less than 0.005 uCi removable activity. The following Minimum Detectable Activities are based upon a background at the indicated value. Background was at or below these levels when the above tests were completed.

Well Counter: Captus 3000

Nuclide	MDA	Background
Cs-137	2.0×10^{-4} uCi	145 counts/1 min
Ba-133	8.8×10^{-5} uCi	204 counts/1 min
Co-57	1.7×10^{-5} uCi	53 counts/1 min

RADIATION SAFETY OFFICER: _____