

From: Gabriel, Sandra
Sent: Friday, September 02, 2011 5:41 PM
To: dberkley@ccmh.org
Subject: Additional information for NRC license amendment request for Camden-Clark Memorial Hospital Corporation, mail control 575796

Licensee: Camden-Clark Memorial Hospital Corporation
License No: 47-09772-02
Docket No.: 03003390
Mail Control No.: 575796

To: Daniel Berkley, Radiation Safety Officer

Please send a return e-mail to confirm that you received this message.

I have been assigned to work on the amendment request dated August 10, 2011. Two additional pieces of information are needed regarding the request to add new brachytherapy source models to item 7D of your NRC license.

- 1) Is the current possession limit in item 8D sufficient? If not, please specify a new value.
- 2) What release surveys and release criteria do you plan to use for patients undergoing Cs-131 implants to assure compliance with 10 CFR 35.75 (i.e., to assure that no individual member of the public is likely to receive more than 500 millirem total effective dose equivalent)? The model guidance in Appendix U to NUREG-1556, Vol. 9, Rev. 2 does not address Cs-131.

In addition, please note that the amended license will list the model numbers of these sources in accordance with those currently listed in the Sealed Source and Device Registry: "Model 200" will be added to the listing for the Pd-103 source and the Cs-131 source will be listed as "Model CS-1."

Please submit the information described above within 30 days, under signature of hospital senior management. If a response is not received within 30 days, we will presume that you do not wish to pursue this request. This information may be mailed to the Region I office or submitted by fax to my attention at 610-337-5269, referencing mail control 575796.

Please feel free to contact me with any questions. I am frequently away from the office, but will check e-mail and voicemail daily except for the time period September 12 through 17.

Thank you,

Sandy Gabriel
Senior Health Physicist
Medical Branch
NRC Region I
610-337-5182