



NUCLEAR SOLUTIONS

A DIVISION OF THE TOMAYKO GROUP WWW.TOMAYKOGROUP.COM

Br. 1

RECEIVED
REGION 1
2011 SEP 07 PM 12:40

September 6, 2011

Nuclear Regulatory Commission
Region 1
475 Allendale Road
King of Prussia, PA 19406

03031697

RE: RML # 37-28520-01 Amendment 42 (Please Expedite)

Dear Inspector:

I would like to amend our radioactive materials license as follows:

1. Please remove the following facility location from condition 10.

260 Crest Rd, Cobblestone Building, Suite 206, St Albans , Vermont.

Nuclear Medicine testing ceased at this location on August 11, 2011.
All radioactive waste has been decayed to background and removed.
All radioactive sources have been relocated to another TTG base facility.
Final surveys including wipes are attached.

2. Remove Luis F. Gonzalez III, M.D. as an authorized user.


To expedite the receipt of the amended license could you please fax it to - 412-490-6908 or email a copy to ashtond@ttgnuclear.com

If you need any more information please contact me at 412-427-9284 or 412-490-0302 ex 112.

Sincerely,

Donald Ashton, CNMT
Executive Director
The Tomayko Group/TTG Nuclear Solutions

575953
NMSS/RGN1 MATERIALS-002



NUCLEAR SOLUTIONS

Final Closeout Surveys and Wipes

Radioactive Materials License -NRC-37-28520-01

TTG- Cobblestone Building
260 Crest Road
Suite 206
St Albans, VT 05478

All radioactive waste was decayed to background level.
All sealed sources were transferred to another TTG Base facility.
All surveys and wipes are equal to background. Date performed 8-22-2011

Area Surveys: Background 0.01 mR/Hr

Ludlum Model 14C S/N 103934
Ck Source:ok

Rad Waste Area: 01 mR/Hr
Dose Storage Area: 01 mR/Hr
LBlock: 0.02 mR/Hr
Floor / Doorways : 0.01 mR/Hr
Treadmill Area: 0.01 mR/Hr
Nuclear Camera Area: 0.01 mR/Hr

Area Wipes: Caprac Well s/n 000207 Efficiency used for wipes- 90% or CF 1.1
Numerous wipe samples were taken throughout the facility. Bkg: 282 cpm

Area Wipe:	Net(DPM)
Rad Waste Area:	0
Dose Storage Area	0
LBlock:	0
Floor/ Doorways:	0
Treadmill Area:	0
Nuclear Camera Area	0
Multiple Random Wipes	0

Completed By: Amy Kutchukian, CNMT
Reviewed By: Donald Ashton , CNMT, Executive Director

August 22, 2011
August 23, 2011

This is to acknowledge the receipt of your letter/application dated

9/6/2011, and to inform you that the initial processing which includes an administrative review has been performed.

Amendment (37-28520-01)
There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 575953.
When calling to inquire about this action, please refer to this control number.
You may call us on (610) 337-5398, or 337-5260.