

U.S. NUCLEAR REGULATORY COMMISSION

Date: 9/13/2011

TELEPHONE CONVERSATION RECORD

Mail Control or Report No(s). 575628 License No(s). 06-02388-01 Docket No(s). 03001250

Name of Licensee: The Hospital of Central Connecticut

Name of Participant(s): Stu Korchin - RSO

Telephone No. 203-494-2852

Subject: RAI (575628) dated 9/13/2011

(NOTE: This will be used as the Documents Title in ADAMS)

In the letter dated 7/11/2011, HCC requested the addition of a PET/CT Suite. The suite had been approved by the State of CT a number of years ago. J. Nguyen inspected this facility and realized that floor plans for the PET/CT were not included in our licensing documentation. As part of the inspection, J. Nguyen requested that the licensee provide information on the PET/CT suite in an amendment.

On 9/13/2011 I contacted the RSO, S. Korchin for additional information concerning the suite.

1. Describe the amount of lead in the walls and ceiling of the Hot Lab and scanner room.
2. Calculations or a survey for the Hot Lab to confirm public dose limit is not exceeded.

Action Required: Wait for information

Document Availability:

Publicly Available

Non-Publicly Available

Non-Sensitive

Non-Sensitive Copyright

Sensitive- Proprietary

Sensitive - Privacy Act/ PII

Sensitive - Internal

Sensitive - Security-Related

Immediate Release

Normal Release Date: 9/21/2011

Delayed Release Date:

SUNSI Review Completed By: TLWeidner / RA /

Document Accession #: ML112560230