



September 23, 2011

Licensing Section  
U.S. Nuclear Regulatory Commission Region III  
2443 Warrenville Road  
Suite 210  
Attn: Toye Simmons  
Lisle, IL 60532-4352

RE: 24-26445-01

Subject: Release of Building 14 for Unrestricted Use

Ms Simmons:

We repeated the survey of Building 14 on September 22 & 23, 2011. We did a survey and wipe in each room of the building. In the primary radioactive material use areas we survey each square meter. We only use I-125 RIA kits in this area. Radioactive material has not been used in the building in more than two years. The stored radioactive material was removed prior to the original survey. The area is free of fixed and removable contamination. The diagram of the building is attached. The survey data and equipment is also provided.

Sorry about the delay getting this information to you. I believe this information demonstrates that the area is radiation free and can be released for unrestricted use. If you need further information or have question please contact me at 815-370-6538.

Sincerely,



James C. Hatten  
Radiation Safety Officer

jch

Att: Survey Documents

**EMD Millipore is a division of Merck KGaA, Darmstadt, Germany**

EMD Millipore  
Millipore Corporation  
15 Research Park Drive  
St. Charles, MO 63304  
U.S.A.



EMD Millipore  
24-26445

| Building 14 Close Out Survey |          |            | 22-Sep-11 |                     |
|------------------------------|----------|------------|-----------|---------------------|
| GAMMA COUNTER                |          |            |           | Survey Meter        |
| Micomedic 28019 SN: 534      |          | Eff= 0.774 |           | Ludlum 3 SN: 188807 |
| Count time=5 minute          |          | MDA= 21    |           | Cal: Dec 29, 2011   |
| Location                     | Sample # | CPM        | DPM       | uR/hr               |
| Surveyor: James Hatten       | bkg=     | 41         |           | 0.4                 |
| Office                       | 1        | 0          | 0         | Bkg                 |
| Office                       | 2        | 0          | 0         | Bkg                 |
| Office                       | 3        | 0          | 0         | Bkg                 |
| Office                       | 4        | 0          | 0         | Bkg                 |
| Office                       | 5        | 0          | 0         | Bkg                 |
| Office                       | 6        | 4          | 5         | Bkg                 |
| Office                       | 7        | 0          | 0         | Bkg                 |
| Office                       | 8        | 0          | 0         | Bkg                 |
| Office                       | 9        | 0          | 0         | Bkg                 |
| Office                       | 10       | 0          | 0         | Bkg                 |
| Office                       | 11       | 5          | 6         | Bkg                 |
| Office                       | 12       | 0          | 0         | Bkg                 |
| Office                       | 13       | 0          | 0         | Bkg                 |
| Office                       | 14       | 3          | 4         | Bkg                 |
| Office                       | 15       | 7          | 9         | Bkg                 |
| Office                       | 16       | 4          | 5         | Bkg                 |
| Office                       | 17       | 5          | 6         | Bkg                 |
| Office                       | 18       | 7          | 9         | Bkg                 |
| Office                       | 19       | 0          | 0         | Bkg                 |
| Office                       | 20       | 0          | 0         | Bkg                 |
| Office                       | 21       | 0          | 0         | Bkg                 |
| Office                       | 22       | 0          | 0         | Bkg                 |
| Office                       | 23       | 0          | 0         | Bkg                 |
| Office                       | 24       | 0          | 0         | Bkg                 |
| Office                       | 25       | 0          | 0         | Bkg                 |
| Office                       | 26       | 0          | 0         | Bkg                 |
| Office                       | 27       | 0          | 0         | Bkg                 |
| Office                       | 28       | 0          | 0         | Bkg                 |
| Office                       | 29       | 0          | 0         | Bkg                 |
| Office                       | 30       | 0          | 0         | Bkg                 |
| Wharehouse                   | 31       | 0          | 0         | Bkg                 |
| Wharehouse                   | 32       | 0          | 0         | Bkg                 |
| Storage                      | 33       | 0          | 0         | Bkg                 |
| Storage                      | 34       | 0          | 0         | Bkg                 |
| Storage                      | 35       | 0          | 0         | Bkg                 |
| Main Lab                     | 36       | 0          | 0         | Bkg                 |
| Main Lab                     | 37       | 0          | 0         | Bkg                 |
| Main Lab                     | 38       | 0          | 0         | Bkg                 |
| Main Lab                     | 39       | 0          | 0         | Bkg                 |

EMD Millipore  
24-26445

| Building 14 Close Out Survey |          |            | 22-Sep-11 |                     |
|------------------------------|----------|------------|-----------|---------------------|
| GAMMA COUNTER                |          |            |           | Survey Meter        |
| Micomedic 28019 SN: 534      |          | Eff= 0.774 |           | Ludlum 3 SN: 188807 |
| Count time=5 minute          |          | MDA= 21    |           | Cal: Dec 29, 2011   |
| Location                     | Sample # | CPM        | DPM       | uR/hr               |
| Surveyor: James Hatten       | bkg=     | 41         |           | 0.4                 |
| Main Lab                     | 40       | 6          | 8         | Bkg                 |
| Main Lab                     | 41       | 0          | 0         | Bkg                 |
| Main Lab                     | 42       | 0          | 0         | Bkg                 |
| Main Lab                     | 43       | 0          | 0         | Bkg                 |
| Main Lab                     | 44       | 1          | 1         | Bkg                 |
| Old Lab                      | 45       | 2          | 3         | Bkg                 |
| Old Lab                      | 46       | 7          | 9         | Bkg                 |
| Old Lab                      | 47       | 5          | 6         | Bkg                 |
| Loading Dock                 | 48       | 0          | 0         | Bkg                 |
| Loading Dock                 | 49       | 0          | 0         | Bkg                 |
| Loading Dock                 | 50       | 0          | 0         | Bkg                 |
| Loading Dock                 | 51       | 0          | 0         | Bkg                 |
| Hall Way                     | 52       | 0          | 0         | Bkg                 |
| Hall Way                     | 53       | 0          | 0         | Bkg                 |
| Storage                      | 54       | 0          | 0         | Bkg                 |
| Storage                      | 55       | 8          | 10        | Bkg                 |
| Storage                      | 56       | 6          | 8         | Bkg                 |
| Hallway                      | 57       | 0          | 0         | Bkg                 |
| Storage                      | 58       | 0          | 0         | Bkg                 |
| Storage                      | 59       | 0          | 0         | Bkg                 |
| Storage                      | 60       | 2          | 3         | Bkg                 |