

**Torres, RobertoJ**

---

**From:** Scott Leavitt [saleavitt@yahoo.com]  
**Sent:** Tuesday, October 04, 2011 7:23 PM  
**To:** Torres, RobertoJ  
**Cc:** Scott A Leavitt  
**Subject:** mail control number 575767  
**Attachments:** 110804 Response.pdf

Roberto,

Let me know if you need any additional information.


Thanks,

Scott A Leavitt  
Cel 208 989-4271  
Office 208 463-0003

**RECEIVED**

OCT -5 2011

**DNMS**



4 October 2011  
Mail Control Number **575767**

Roberto J. Torres  
Senior Health Physicist  
US NRC Region IV  
612 East Lamar Boulevard, Suite 400  
Arlington, TX 76011

Dear Mr. Torres,

1 - Idaho Testing & Inspection, Inc. (ITI) has three Troxler gauges for this license. ITI has no plans at this time for any future gauges. We possess two model 3440 gauges and one model 3411.

2 - The actual storage location is better represented by 35 Fourteenth Avenue South.

3 - Operating, Emergency, and Security procedures will be developed, implemented, and maintained and will meet the criteria described in the errata sheet to Appendix H of NUREG-1556, Volume 1, Revision 1, and will provide copies of these procedures to all gauge users and at each job site.

4 - Idaho Testing & Inspection will offer for transportation non-bulk packages that will be marked with the proper shipping name and identification number. The non-bulk packages (Type A shipping containers) will have labels identifying the material as UN3332, Radioactive Material, Type A Package, Special Form 7, RQ. In addition, the inventory log, bill of lading and any other regulatory required documentation will indicate UN3332, Radioactive Material, Type A package, Special Form 7, RQ.

Should you have any questions or require additional information, please contact our office.

Respectfully submitted;  
**Idaho Testing & Inspection, Inc.**

A handwritten signature in black ink, appearing to read 'Scott A. Leavitt', is written over a horizontal line.

Scott A. Leavitt  
Operations Manager

cc: File

**RECEIVED**

OCT -5 2011

**DNMS**



UNITED STATES  
NUCLEAR REGULATORY COMMISSION  
REGION IV  
612 EAST LAMAR BLVD, SUITE 400  
ARLINGTON, TEXAS 76011-4125

September 8, 2011

Idaho Testing & Inspection, Inc.  
ATTN: Scott A. Leavitt  
Radiation Safety Officer  
25 Fourteenth Avenue South, Suite 1  
Nampa, Idaho 83651

SUBJECT: REQUEST FOR ADDITIONAL INFORMATION

Dear Mr. Leavitt:

The NRC has completed the technical review of Idaho Testing & Inspection renewal application dated June 29, 2011, and additional information is needed to complete the renewal process. Please provide the following information or commitments for NRC review within 20 business days in a signed and dated letter in company letterhead. You can provide your response via reply electronic mail in a pdf file. Make reference to mail control number 575767 when providing your response.

- ✓ 1. Indicate how many portable nuclear gauges need to be authorized in the license. Include any future projections. Provide gauge manufacturer's name and model number.
- ✓ 2. The most current amendment for License No. 11-27542-01 authorizes the following address as a storage location: 25 Fourteenth Avenue South, Suite 1. The renewal application requests that the following address be listed as a storage location: 35 Fourteenth Avenue South. Indicate which of these two addresses is the correct address for the storage location.
- ✓ 3. Revised 10 CFR 30.34 requires that "each portable gauge licensee shall use a minimum of two independent physical controls that form tangible barriers to secure portable gauges from unauthorized removal, whenever portable gauges are not under the control and constant surveillance of the licensee" (i.e., when not in use). Guidance on these security procedures is provided in the errata sheet to Appendix H of NUREG-1556, Volume 1, revision 1, which can be found at: <http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v1/r1/apph-errata.pdf>. Read the following two commitments and commit to only one of them.

"We will implement and maintain the "Operating, Emergency, and Security Procedures" described in the errata sheet to Appendix H of NUREG-1556, Volume 1, Revision 1, and will provide copies of these procedures to all gauge users and at each job site."

or

"Operating, Emergency, and Security procedures will be developed, implemented, and maintained and will meet the criteria described in the errata sheet to Appendix H of NUREG-1556, Volume 1, Revision 1, and will provide copies of these procedures to all gauge users and at each job site."

- ✓ 4. 10 CFR 71.5(a) requires that a licensee who transports licensed material outside of the site of usage, as specified in the NRC license, or where transport is on public highways, or who delivers licensed material to a carrier for transport, comply with the applicable requirements of the regulations appropriate to the mode of transport of the DOT in 49 CFR Parts 107, 171-180, and 390-397.

49 CFR 172.301 (a) requires in part that each person who offers for transportation hazardous material in a non-bulk packaging shall mark the package with the proper shipping name and identification number (preceded by "UN") for the material shown in the 172.101 table. Pursuant to 49 CFR 172.101, radioactive material is classified as hazardous material.

Please note that the U.S. Department of Transportation identification number UN2974 for portable gauges is no longer a valid identification number. Please commit to the following language: "Idaho Testing & Inspection will offer for transportation non-bulk packages that will be marked with the proper shipping name and identification number. The non-bulk packages (Type A shipping containers) will have labels identifying the material as UN3332, Radioactive Material, Type A package, Special Form 7, RQ. In addition, the inventory log, bill of lading and any other regulatory required documentation will indicate UN3332, Radioactive Material, Type A package, Special Form 7, RQ".

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the NRC's document system (ADAMS). ADAMS is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Thank you for your cooperation.

Sincerely,



Roberto J. Torres, Senior Health Physicist  
Nuclear Materials Safety Branch B

Docket: 030-34171  
License: 11-27542-01  
Control: 575767

**Request for additional information from US NRC**

Torres, RobertoJ

**Sent:** Thursday, September 08, 2011 6:42 AM**To:** saleavitt@yahoo.com

---

Dear Mr. Leavitt:

The NRC has completed the technical review of Idaho Testing & Inspection renewal application dated June 29, 2011, and additional information is needed to complete the renewal process. Please provide the following information or commitments for NRC review within 20 business days in a signed and dated letter in company letterhead. You can provide your response via reply electronic mail in a pdf file. Make reference to mail control number 575767 when providing your response.

1. Indicate how many portable nuclear gauges need to be authorized in the license. Include any future projections. Provide gauge manufacturer's name and model number.
2. The most current amendment for License No. 11-27542-01 authorizes the following address as a storage location: 25 Fourteenth Avenue South, Suite 1. The renewal application requests that the following address be listed as a storage location: 35 Fourteenth Avenue South. Indicate which of these two addresses is the correct address for the storage location.
3. Revised 10 CFR 30.34 requires that "each portable gauge licensee shall use a minimum of two independent physical controls that form tangible barriers to secure portable gauges from unauthorized removal, whenever portable gauges are not under the control and constant surveillance of the licensee" (i.e., when not in use). Guidance on these security procedures is provided in the errata sheet to Appendix H of NUREG-1556, Volume 1, revision 1, which can be found at: <http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v1/r1/apph-errata.pdf>. Read the following two commitments and commit to only one of them.

"We will implement and maintain the "Operating, Emergency, and Security Procedures" described in the errata sheet to Appendix H of NUREG-1556, Volume 1, Revision 1, and will provide copies of these procedures to all gauge users and at each job site."

or

"Operating, Emergency, and Security procedures will be developed, implemented, and maintained and will meet the criteria described in the errata sheet to Appendix H of NUREG-1556, Volume 1, Revision 1, and will provide copies of these procedures to all gauge users and at each job site."

4. 10 CFR 71.5(a) requires that a licensee who transports licensed material outside of the site of usage, as specified in the NRC license, or where transport is on public highways, or who delivers licensed material to a carrier for transport, comply with the applicable requirements of the regulations appropriate to the mode of transport of the DOT in 49 CFR Parts 107, 171-180, and 390-397.

49 CFR 172.301 (a) requires in part that each person who offers for transportation hazardous material in a non-bulk packaging shall mark the package with the proper shipping name and identification number (preceded by "UN") for the material shown in the 172.101 table. Pursuant to 49 CFR 172.101, radioactive material is classified as hazardous material.

Please note that the U.S. Department of Transportation identification number UN2974 for portable gauges is no longer a valid identification number. Please commit to the following language: "Idaho Testing & Inspection will offer for transportation non-bulk packages that will be marked with the proper shipping name and identification number. The non-bulk packages (Type A shipping containers) will have labels identifying the material as UN3332, Radioactive Material, Type A package, Special Form 7, RQ. In addition, the inventory log, bill of lading and any other regulatory required documentation will indicate UN3332, Radioactive Material, Type A package, Special Form 7, RQ".

Please confirm receipt of this email. Thank you for your cooperation.

Roberto J. Torres  
Senior Health Physicist  
US NRC Region IV  
612 East Lamar Boulevard, Suite 400  
Arlington, TX 76011  
817-860-8189  
[robertoj.torres@nrc.gov](mailto:robertoj.torres@nrc.gov)