

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED: Cameron Regional Medical Center 1600 E. Evergreen Cameron, MO 64429		2. NRC/REGIONAL OFFICE  Region III: 2443 Warrenville Rd., Ste. 210 Lisle, IL 60532-4352	
REPORT NUMBER(S) 2011-001			
3. DOCKET NUMBER(S) 030-34706	4. LICENSE NUMBER(S) 24-32075-01	5. DATE(S) OF INSPECTION <i>Oct. 6, 2011</i>	

**LICENSEE:**

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

Non-cited violation(s) were discussed involving the following requirement(s):

<p>10/6/2011</p> <p>Region III: 2443 Warrenville Rd., Ste. 210 Lisle, IL 60532-4352</p>	<p>10/6/2011</p> <p>Region III: 2443 Warrenville Rd., Ste. 210 Lisle, IL 60532-4352</p>
---	---

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

(Violations and Corrective Actions)

**Statement of Corrective Actions**

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Deborah A. Piskura	<i>Deborah A. Piskura</i>	10/6/2011
BRANCH CHIEF	Tamara E. Bloomer	<i>Tamara Bloomer</i>	10/17/11

**Docket File Information**  
**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED: Cameron Regional Medical Center  REPORT NUMBER(S) 2011-001		2. NRC/REGIONAL OFFICE  Region III: 2443 Warrenville Rd., Ste. 210 Lisle, IL 60532-4352	
3. DOCKET NUMBER(S) 030-34706		4. LICENSE NUMBER(S) 24-32075-01	5. DATE(S) OF INSPECTION 10/06/2011
6. INSPECTION PROCEDURES USED 87130		7. INSPECTION FOCUS AREAS 03.01 - 03.07	

**SUPPLEMENTAL INSPECTION INFORMATION**

1. PROGRAM CODE(S) 2120	2. PRIORITY 3	3. LICENSEE CONTACT Lawrence Lee, M.D., RSO	4. TELEPHONE NUMBER (816) 276-4449
----------------------------	------------------	--	---------------------------------------

Main Office Inspection      Next Inspection Date: Oct. 2014

Field Office Inspection \_\_\_\_\_

Temporary Job Site Inspection \_\_\_\_\_

**PROGRAM SCOPE**

This licensee was a small community hospital and authorized to use licensed material permitted by Sections 35.100, 35.200, and 35.300. The nuclear medicine department was staffed with two technologists who performed approximately 80+ diagnostic procedures monthly. These included a full spectrum of studies. The licensee received unit doses from a licensed radiopharmacy. Although the licensee was authorized for Section 35.300 material, the hospital had not administered these therapeutic procedures since the previous inspection. The licensee retained the services of a consultant who audited the radiation safety program quarterly.

This inspection consisted of interviews with select licensee personnel; a review of select records; a tour of the nuclear medicine department; and independent measurements. The inspector observed the administration of one procedure for a gastric emptying study. The inspection included observations of dose calibrator QA checks, security of byproduct material, use of personnel monitoring, package receipt/transfer and surveys.