


<b>NRC FORM 314</b> <small>(12-2010)          10 CFR 30.36(j)(1); 40.42(j)(1);          70.36(j)(1); and 72.54(k)(5)(1)(1)</small>	<b>U.S. NUCLEAR REGULATORY COMMISSION</b> <div style="font-size: 2em; margin-top: 10px;">Q-0</div>	<b>APPROVED BY OMB: NO. 3150-0028</b> <b>EXPIRES: 10/31/2013</b> <small>Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the Information Services Branch (T-5 F53), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to <a href="mailto:Infocollects.Resource@nrc.gov">Infocollects.Resource@nrc.gov</a>, and to the Desk Officer, Office of Information and Regulatory Affairs, NE08-10202, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.</small>				
<b>CERTIFICATE OF DISPOSITION OF MATERIALS</b>						
<b>LICENSEE NAME AND ADDRESS</b> JohnsonFoilS 40 Progress Avenue Springfield, MA 01104		<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td style="width:50%;"><b>LICENSE NUMBER</b> 20-20982-02</td> <td style="width:50%;"><b>DOCKET NUMBER</b> 030-34367/20-20982-01</td> </tr> <tr> <td colspan="2"><b>LICENSE EXPIRATION DATE</b> September 30, 2011</td> </tr> </table>	<b>LICENSE NUMBER</b> 20-20982-02	<b>DOCKET NUMBER</b> 030-34367/20-20982-01	<b>LICENSE EXPIRATION DATE</b> September 30, 2011	
<b>LICENSE NUMBER</b> 20-20982-02	<b>DOCKET NUMBER</b> 030-34367/20-20982-01					
<b>LICENSE EXPIRATION DATE</b> September 30, 2011						
<b>A. LICENSE STATUS (Check the appropriate box)</b> <input checked="" type="checkbox"/> This license has expired. <input type="checkbox"/> This license has not yet expired; please terminate it.						
<b>B. DISPOSAL OF RADIOACTIVE MATERIAL</b> <small>(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)</small> The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:						
<input type="checkbox"/> 1. No radioactive materials have ever been procured or possessed by the licensee under this license.						
<input checked="" type="checkbox"/> 2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner.						
<input checked="" type="checkbox"/> a. Transfer of radioactive materials to the licensee listed below: <div style="margin-left: 40px;">Two gauges in possession to retained under Commonwealth of Massachusetts Materials License 20-9821</div>						
<input type="checkbox"/> b. Disposal of radioactive materials:						
<input type="checkbox"/> 1. Directly by the licensee:						
<input type="checkbox"/> 2. By licensed disposal site:						
<input type="checkbox"/> 3. By waste contractor:						
<input type="checkbox"/> c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.						
<b>C. SURVEYS PERFORMED AND REPORTED</b>						
<input type="checkbox"/> 1. A radiation survey was conducted by the licensee. The survey confirms:						
<input type="checkbox"/> a. the absence of licensed radioactive materials						
<input type="checkbox"/> b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.						
<input checked="" type="checkbox"/> 2. A copy of the radiation survey results:						
<input checked="" type="checkbox"/> a. is attached; or <input type="checkbox"/> b. is not attached (Provide explanation); or <input type="checkbox"/> c. was forwarded to NRC on: _____ Date _____						
<input type="checkbox"/> 3. A radiation survey is not required as only sealed sources were ever possessed under this license, and						
<input type="checkbox"/> a. The results of the latest leak test are attached; and/or <input type="checkbox"/> b. No leaking sources have ever been identified.						
The person to be contacted regarding the information provided on this form:						
NAME Dean C Miller	TITLE Global Business Manager	TELEPHONE (Include Area Code) 413-750-4629 E-MAIL ADDRESS <a href="mailto:dean.miller@johnsonfoils.com">dean.miller@johnsonfoils.com</a>				
Mail all future correspondence regarding this license to: Dean C Miller, JohnsonFoilS Inc; 40 Progress Avenue; Springfield, MA; 01104						
<b>C. CERTIFYING OFFICIAL</b> I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT						
PRINTED NAME AND TITLE DEAN C MILLER, GLOBAL BUS. MGR.	SIGNATURE 	DATE 10/20/2011				
<b>WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.</b>						

**LEAK TEST CERTIFICATE**

**Owner:** Johnson Foils  
**Address:** Springfield, MA - USA  
**Model/Serial Number:** 104P S/N 2064

**Source Number** 0539AR  
**Manufacturer** AEA Technology  
**Isotope** Am-241  
**Strength** 0.925 GBq / 25 mCi  
**Model Number** AMC.P6  
**Date Wiped** 23-Jun-11  
**Results\*** <4.0 Bq / <.0001 microcuries  
**Shutter Check** Performed

**Dated:**

30-Jun-11

**Approved By:**

A handwritten signature in black ink, appearing to read 'Frank Aguirre', written over a horizontal line.

Frank Aguirre  
Radiation Safety Officer

\*<185 becquerels / <.006 microcuries indicates non-leaking source

NDC Infrared Engineering Inc., 5314 North Irwindale Avenue, Irwindale, CA 91706 USA  
Telephone : (626) 960 - 3300, Fax: (626) 939-3870, info@ndcinfrared.com

www.ndcinfrared.com

**LEAK TEST CERTIFICATE**

**Owner:** Johnson Foils  
**Address:** Springfield, MA - USA  
**Model/Serial Number:** 104P S/N 1721

**Source Number** 049/02  
**Manufacturer** IPL  
**Isotope** Am-241  
**Strength** 0.925 GBq / 25 mCi  
**Model Number** AM1.P08  
**Date Wiped** 23-Jun-11  
**Results\*** <4.0 Bq / <.0001 microcuries  
**Shutter Check** Performed

**Dated:**

30-Jun-11

**Approved By:**

A handwritten signature in black ink, appearing to read 'Frank Aguirre', written over a horizontal line.

Frank Aguirre  
Radiation Safety Officer

\*<186 becquerels / <.005 microcuries indicates non-leaking source

NDC Infrared Engineering Inc., 5314 North Irwindale Avenue, Irwindale, CA 91706 USA  
Telephone: (626) 960 - 3300, Fax: (626) 939-3870, [info@ndcinfrared.com](mailto:info@ndcinfrared.com)

[www.ndcinfrared.com](http://www.ndcinfrared.com)