

From: Ullrich, Elizabeth
Sent: Friday, November 04, 2011 8:16 AM
To: Edmond DeRosa
Subject: request for distribution, control No. 576259

License No. 19-25203-01
Docket No. 030-38295
Control No. 576259
Neutron Products, Inc.

Mr. DeRosa,

This refers to your letter dated October 25, 2011 in which you request to authorization to distribute and transfer re-furbished teletherapy units, unit components and new and used cobalt-60 sources.

1. The NRC does not have jurisdiction over the distribution of the re-furbished teletherapy units or unit components, but does have jurisdiction over the sealed sources contained in the unit. The NRC's jurisdiction extends to the cobalt-60 sources and the use of these sources only in authorized devices as approved in the Sealed Source and Device (SSD) registry. The sealed sources may only be used in approved devices as listed on the license. Those devices must meet the requirements of the SSD registration for those devices. No response to this item is needed.
2. The license will be written to authorize distribution of new sources and re-distribution of used sources; if you have any questions or concerns about this, please respond.
3. You described the tests and inspections to be a wipe test, a visual inspection, and a helium pressurization bubble test. Please state the standard or procedures that you follow for these tests; if acceptance criteria is not specified in the standard or procedures, describe how you know that sources have passed each test; and describe the actions you will take if a source does not pass the acceptance criteria.
4. Confirm that you will maintain records of the testing with other survey records, as described in 10 CFR 20.1501 and 20.2103.

Do not reply by electronic mail. A hard copy, facsimile, or pdf response is acceptable. Contact me as listed below if you have any questions.

Thanks,
Betsy