

**From:** Torres, RobertoJ  
**Sent:** Tuesday, October 11, 2011 4:11 PM  
**To:** 'Suzanne Fraser'

Mr. McManus:

Please revise the "Pocket Dosimeter Calibration Procedures" submitted to the NRC on August 2011, and resubmit via reply email. As discussed over the phone, I have not been able to determine, from the procedure, the methodology for ensuring that the pocket dosimeter is calibrated.

Also, please review all calibration procedures to explain that inverse square law is used for determining distance, etc., as appropriate.

Thank you.

Roberto J. Torres  
Senior Health Physicist  
U.S. Nuclear Regulatory Commission - Region IV  
Division of Nuclear Materials Safety  
Nuclear Materials Safety Branch B  
612 East Lamar Boulevard, Suite 400  
Arlington, Texas 76011-4125  
Telephone 817-860-8189  
Facsimile 817-860-8263  
[robertoj.torres@nrc.gov](mailto:robertoj.torres@nrc.gov)