U.S.NRC
United States Nuclear Regulatory Commission
Protecting People and the Environment

NUCLEAR REGULATORY COMMISSION
REGION III
2443 WARRENVILLE ROAD
LISLE, ILLINOIS 60532-4351
(630) 829-9892 FAX: (630) 515-1078

TO: file

COMPANY: N/A

Number of Pages: 1

Email: N/A

CONVERSATION RECORD

TIME	DATE
9:30 am	10/28/2011

NAME OF PERSON(S) CONTACTED	TELEPHONE NO.	ORGANIZATION
Lisa Tincher, RSO	(314) 653-4350	Christian Hospital – Dept. of Nuclear Medicine

REPRESENTED PERSON or PERSONS	ORGANIZATION
Lisa Tincher, RSO	Christian Hospital – Dept. of Nuclear Medicine

SUBJECT
License No.: 24-13383-01 Control No.: 576041

SUMMARY

We have conducted a preliminary review your license amendment request dated August 29, 2011, and find that we are unable to continue this action until we have received information regarding the following:

1. The request was silent as to whether PET is being used at the proposed new location of use. If PET is being used, additional shielding calculations may be required.
   RESPONSE: The licensee indicated, via phone conversation of October 28, 2011 that the facility will be limited to the use of isotopes under 10 CFR 35.200, and that no PET isotopes would be used. No additional information is required.

2. The submitted facility diagram for the proposed new location of use lacks detail as to the areas adjacent to the radioactive materials use area. Please resubmit including a description of any activities being conducted in the areas contiguous to the proposed radioactive materials use area(s). In addition, the diagram should show details for the nuclear medicine suite, including locations of any sinks, waste areas, and exhaust hoods. The hot lab should be clearly indicated on the diagram, and be free of superfluous architectural markings, unrelated to radioactive materials use concerns.
   RESPONSE: The contact person indicated, via phone conversation on October 28, 2011, that the requested response would be submitted via facsimile. Additional information was received on October28, 2011. No additional response is required.

Please submit :

- Revised facility diagram

to my attention via either facsimile (630-515-1078), phone (630-829-9892), or as a pdf file attached to an email addressed to sara.forster@nrc.gov. **Reference Control No. 576041**, listed at the top of this memo.

Note: Facsimile submitted October 28, 2011 is complete. No additional response is needed.

NAME OF PERSON DOCUMENTING CONVERSATION	SIGNATURE	DATE
Sara A.B. Forster	Sara A.B. Forster	11/1/2011