



**Docket File Information**  
**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED:  Mirview Incorporated 120 Enterprise Drive Ann Arbor, MI 48103  REPORT NUMBER(S) 11-01	2. NRC/REGIONAL OFFICE  Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352
--	---

3. DOCKET NUMBER(S)  030-34321	4. LICENSE NUMBER(S)  21-26775-01	5. DATE(S) OF INSPECTION  November 7, 2011
--------------------------------------	---	--

6. INSPECTION PROCEDURES USED  87130	7. INSPECTION FOCUS AREAS  03.01-03.07
--	--

**SUPPLEMENTAL INSPECTION INFORMATION**

1. PROGRAM CODE(S)  2220	2. PRIORITY  3	3. LICENSEE CONTACT  Ray Carlson	4. TELEPHONE NUMBER  (734) 455-4730
--------------------------------	----------------------	--	---

Main Office Inspection                      Next Inspection Date: 11/07/2014

Field Office Inspection

Temporary Job Site Inspection Eaton Rapids Med. Ctr., Eaton Rapids, MI 48827

**PROGRAM SCOPE**

Mirview Incorporated was authorized by the license for 10 CFR 35.100 and 35.200, as a mobile medical service and currently has two mobile medical vans. The mobile van at Eaton Rapids Medical Center conducts nuclear medicine studies on Mondays, Fridays, and every other Wednesday from 7:00am to 4:30pm. The licensee has one nuclear medicine technologist (NMT) and 2-3 patients are scheduled for each day. The majority of studies are cardiac studies, with bone scans and HIDA scans scheduled as needed. Unit dosages are used and received from a local nuclear pharmacy delivered to the NMT directly. No Mo-99/Tc-99m generators, Tc-99m bulk dosages, or PET materials are used. The licensee has one gamma camera and one treadmill. The licensee's radiation safety program is reviewed by a consulting physicist.

The inspector interviewed the technologist who demonstrated knowledge of operating procedures, emergency procedures, and NRC radiation safety requirements. A review of dose calibrator tests, leak test results, calibration of survey meters, area surveys and wipe tests, radiation safety training, internal audits, inventory records, and waste disposal records were reviewed with no concerns. A review of dosimetry records did not indicate any whole body or ring badge overexposures or abnormal readings. Security of radioactive material was adequate meeting 10 CFR 20.1801 and 1802 requirements. Area postings were adequate.

No violations or safety concerns were identified during this inspection.

The second mobile unit is located at Spectrum Health Zeeland Community Hospital, 8333 Felch Street, Zeeland, MI 49464. The mobile unit operates Monday, Tuesday and Friday from 7:00am – 4:00pm.