

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

<p>1. LICENSEE/LOCATION INSPECTED: Carson City Hospital 406 E. Elm Street Carson City, Michigan REPORT NUMBER(S): 11-01</p>	<p>2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission, Region III 2443 Warrenville Road, Suite 210 Lisle, Illinois 60532</p>
---	---

<p>3. DOCKET NUMBER(S) 030-10847</p>	<p>4. LICENSEE NUMBER(S) 21-16339-01</p>	<p>5. DATE(S) OF INSPECTION November 15, 2011</p>
--	--	---

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied

_____ Non-cited violation(s) were discussed involving the following requirement(s):
- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Andrew M. Bramnik	<i>Andrew M. Bramnik</i>	11/15/2011
Branch Chief	Tamara E. Bloomer	<i>Tamara E. Bloomer</i>	11/25/11

Docket File Information
SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE Carson City Hospital REPORT NUMBER(S) 11-01	2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission, Region III 2443 Warrenville Road, Suite 210 Lisle, Illinois 60532
---	---

3. DOCKET NUMBER(S) 030-10847	4. LICENSEE NUMBER(S) 21-16339-01	5. DATE(S) OF INSPECTION November 15, 2011
----------------------------------	--------------------------------------	---

6. INSPECTION PROCEDURES 87131	7. INSPECTION FOCUS AREAS 03.01 – 03.07
-----------------------------------	--

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM 2120	2. PRIORITY 3	3. LICENSEE CONTACT Donna R. Moyer, D.O. – RSO	4. TELEPHONE NUMBER 989-584-6165
--------------------	------------------	---	-------------------------------------

Main Office Inspection

Field Office Inspection

Temporary Job Site Inspection _____

Next Inspection Date: November 2014

PROGRAM SCOPE

This was a routine inspection of an 80 bed hospital that performed approximately 85 diagnostic nuclear medicine procedures per month. Two full time nuclear medicine technologists performed all patient procedures. The licensee obtained unit doses and occasional bulk doses of licensed material from an area nuclear pharmacy, and did not use xenon-133 or molybdenum/technetium generators. The licensee performed primarily cardiac, bone, HIDA, gastric emptying, and thyroid uptake scans. The previous inspection report, dated September 5, 2008, erroneously stated that the licensee performed therapy treatments using iodine-131. Although the licensee was authorized for activities under Title 10 of the Code of Federal Regulations (10 CFR) Section 35.300, the licensee had never performed any therapeutic administrations of byproduct material.

PERFORMANCE OBSERVATIONS

No administrations of byproduct materials were available for observation during the inspection. Interviews of available staff revealed an adequate level of understanding of emergency and material handling procedures and techniques. Dose calibrator constancy checks, package receipt, daily surveys, and waste handling and disposal procedures were successfully demonstrated. An outside consultant performed quarterly program audits that were adequate to oversee the program.

Licensed material was adequately secured and not readily accessible to members of the general public. The licensee possessed a radiation survey meter that was calibrated, operational, and performed well in side-by-side comparison with an NRC instrument. Independent measurements did not indicate readings in excess of 10 CFR Part 20 limits in restricted or unrestricted areas. Personal whole body and extremity dosimetry were observed being worn by the staff during the inspection, and records did not indicate doses in excess of 10 CFR Part 20 limits. Dosimetry records indicated that the highest annual whole body and extremity readings since the previous inspection were 139 millirem (mrem) and 490 mrem, respectively.

No violations were identified during this inspection.