

**From:** Elliott, Robin  
**Sent:** Thursday, December 01, 2011 1:54 PM  
**To:** nuclearmedhearts@yahoo.com  
**Cc:** Gabriel, Sandra  
**Subject:** USNRC license renewal

Laboratorio Vascular y Cardiovascular  
License no: 52-25574-01  
Docket no: 030-35809  
Mail control no: 576001

Dear Carmen,

Additional information is needed to process the renewal of your license. As we discussed on the phone November 28 the information needed and a review of our discussion is described below:

1. To facilitate future communications, you provided the business email addresses for Dr. Novoa and Dr. Pascual Ferrer as [nuclearmedhearts@yahoo.com](mailto:nuclearmedhearts@yahoo.com) . You also provided Dr. Pascual Ferrer's phone number as (787) 621-3700 ext. 1276.
2. We discussed your continued authorization for 35.300 (e.g. I131). It was noted that you have not used this material since your last renewal and notes in our inspection files indicate that you do not have any plans to use it in the near future. I indicated that your license would be viewed as a lower priority and would receive a longer term for regular scheduled inspections if you discontinued this authorization. Please state if you wish to drop the 35.300 authorization.
3. We discussed the facility diagram you submitted and you stated that the area it represents is the same as the previously submitted diagram. You indicated that no renovations have taken place since the last renewal. You provided the following additional information for the diagram:
  - a. To the left of the camera room is the garage
  - b. To the right of the stress test room is the exterior of the building
  - c. The Hot lab, injection area, and computer area also sit along an exterior wall of the building.
  - d. The facility is one story so the roof is above and earth below.
  - e. You indicated there is lead shielding between the stress test room and the computer area as well as shielding for the contaminated sharps.

Please confirm the above information regarding the facility is correct by re-stating it in your response to this communication.

4. You stated that you continue to use decay correction as the method for determining your doses prior to administration as required by 10 CFR 35.63(b)(2). Please submit a statement that you have developed and will implement and maintain written procedures for decay correction of unit doses based on the activity or activity concentration determined by the manufacturer in accordance with 10 CFR 20.63(b)(2).

Please let me know if I have misrepresented any of the above information or if you have any additional information to add. Once we receive your responses to items 2, 3, and 4 above, we should be able to finalize the processing of your license renewal. You may respond to my

attention in writing by letter or fax (610) 337-5269 referencing mail control number 576001. All communication must be signed by Dr. Novoa as the management representative. In order to continue a prompt review of your application, we request that you submit your response to this communication within 30 calendar days from the date of this email.

Please send an email confirming receipt of this communication.

Regards,

*Robin L. Elliott*

Health Physicist

Region I

USNRC

King of Prussia, PA 19406

(610) 337-5076 voice

(610) 337-5269 fax

[robin.elliott@nrc.gov](mailto:robin.elliott@nrc.gov)



Think Green - Not every email needs to be printed.