

Goshen, John

From: Greene, Ken [Ken.Greene@cengllc.com]
Sent: Tuesday, December 06, 2011 8:51 AM
To: Goshen, John
Subject: Calvert Cliffs ISFSI License Renewal

John,

During our recent (11/30/11) conference call in regards to Calvert Cliffs ISFSI License Renewal a couple of questions were raised by the NRC. Below are the questions and their associated responses:

1. Where do we have our air particulate filter samples analyzed? – The air filters to be sampled are sent to Constellation Energy's lab located at Constellation's Fort Smallwood facility in Anne Arundel county, Maryland.
2. What is the instrument we use to measure the weekly beta activity collected from the air particulate filters? - Samples are analyzed by an LB5100 Tennelec.
3. What is the uncertainty associated with measuring the beta activity? - The uncertainty of the beta analysis is whichever is greater; 2 sigma of the calculated result or 3 sigma of the MDA. The LIMS (the software used to calculate the results) calculates both values automatically and returns the greater of the two in the reports.

If you have further questions please contact me.

Ken Greene
Licensing Engineer
Calvert Cliffs Nuclear Power Plant
(410) 495-4385

>>> This e-mail and any attachments are confidential, may contain legal, professional or other privileged information, and are intended solely for the addressee. If you are not the intended recipient, do not use the information in this e-mail in any way, delete this e-mail and notify the sender. CEG-IP1