

Note to file regarding mail control 576027
Rutland Regional Medical Center
44-14121-01
030-36360

Termination of 1 Commons and 6 Commons

The licensee submitted close out surveys of the two locations listed above along with leak tests for the sealed sources that were transferred to 12 Commons. I called to verify that the leak test report submitted labeled "Heart Center Leak Test" for 6 Common Street was actually reported in uCi units rather than nCi as stated in the report. Paul Waite, with whom I spoke, and who conducted the test stated that was the case and it was simply an oversight on his part to list nCi rather than uCi on the report.

I also discussed how I would be unable to fully remove old tie-down documents on the license related to the facilities that were being removed due to the equipment listed on them. I informed Paul that Rutland could submit a new commitment relative to equipment that would remove these old documents from their license. If they committed to more generic types of equipment Rutland would have more flexibility rather than listing the exact equipment. He said they would consider doing so in the future.

Robin Elliott /**RA**/
Health Physicist
December 14, 2011