

**U.S. NUCLEAR REGULATORY COMMISSION**

Date: 12/14/11

**TELEPHONE CONVERSATION RECORD**

Time: 1:45pm

Mail Control  
or Report No(s).

575973

License No(s). 06-30657-01

Docket No(s). 03035779

Name of Licensee: Heart Specialists, PC of Southern Connecticut

Name of Participant(s): Roseann Shore, Chief Nuclear Medicine Technologist / Technical Director

Telephone No. (203) 929-2062

Subject: Additional information regarding application for renewal of license

(NOTE: This will be used as the  
Documents Title in ADAMS)

NRC staff contacted the licensee to discuss and confirm the following information provided in a deficiency response letter dated December 1, 2011:

1. The letter indicated that Suite 100 is located on the first floor of a multiple floor building. I asked Roseann to provide a description of the areas and rooms above. I also emphasized to her that the dose to individual members of the public from licensed operation is not to exceed 100 mrem in a year in accordance with 10 CFR 20.1301. Roseanne stated that the areas above Suite 100 are medical offices that are currently vacant.
2. The e-mail addresses listed for Roseann and RSO, Dr. Anthony D'Souza, appear to be personal e-mail addresses. I asked Roseann to indicate whether both e-mail addresses publicly available, otherwise to provide a public contact (business) e-mail address for Dr. D'Souza. Roseann stated that both e-mail addresses are personal and provided a business e-mail address for Dr D'Souza, [awdmd@cttel.net](mailto:awdmd@cttel.net).

**Action Required:** Issue Renewal**Document Availability:** Publicly Available Non-Publicly Available
 Non-Sensitive  
 Sensitive – Privacy Act/ PII  
 Immediate Release

 Non-Sensitive Copyright  
 Sensitive – Internal  
Normal Release Date: 12/22/2011

 Sensitive- Proprietary  
 Sensitive – Security-Related  
Delayed Release Date:

SUNSI Review Completed By: M. Abogunde / RA /

Document Accession #: ML113490208