

**U.S. NUCLEAR REGULATORY COMMISSION**

Date: 1/30/2012

**TELEPHONE CONVERSATION RECORD**Mail Control  
or Report No(s).

576675

License No(s). 06-08349-04

Docket No(s). 03009293

**Name of Licensee:** Charlotte Hungerford Hospital

**Name of Participant(s):** Susan Musser, Ph.D., Physicist  
George Pavlonnis, M.S., Physicist

**Telephone No.** SM: 860-489-6718, mussersu@msu.edu  
GP: 860-331-9379, gp219@columbia.edu

**Subject:** Amendment request to add HDR AMPs and AU  
(NOTE: This will be used as the Documents Title in ADAMS)

This record summarizes a series of telephone and email communications on the date referenced above.

Susan called to ask if it is acceptable for her and/or George to perform an HDR source exchange in the next few days. I informed her that she meets the exception in 10 CFR 35.13(b)(3) and George meets the exception in 35.13(b)(4)(1), so both may work prior to issuance of the amendment. The amendment request included the required notification information per 35.14(a). Therefore, the amendment request may be viewed as a notification. Susan and/or George may perform the source exchange and associated full calibration.

I contacted George to ask if the proposed AU, Dr. Kratzer, has received training on GammaMed HDR use. Dr. Kratzer meets the exception in 35.13(b)(4)(1), however NRC records show only Nucletron training and experience. George said that he will provide to Dr. Kratzer the training required by 35.690(c) for use of the licensee's GammaMed unit within the next few days, then submit documentation.

**Action Required:** Await documentation of GammaMed training for Dr. Kratzer, then issue amendment

Document Availability:

Publicly Available

Non-Publicly Available

Non-Sensitive

Non-Sensitive Copyright

Sensitive- Proprietary

Sensitive - Privacy Act/ PII

Sensitive - Internal

Sensitive - Security-Related

Immediate Release

Normal Release Date: 02/07/2012

Delayed Release Date:

SUNSI Review Completed By: SGabriel / RA /Document Accession #: ML120300438