

U.S. NUCLEAR REGULATORY COMMISSION

Date: 2/6/2012

TELEPHONE CONVERSATION RECORDMail Control
or Report No(s).

576428

License No(s). 06-31416-01

Docket No(s). 03038335

Name of Licensee: The Harold Leever Regional Cancer Center (THLRCC)

Name of Participant(s): Ian Crooks, Director of Physics

Telephone No. 203-575-5533, e-mail icrooks@leevecancercenter.org

Subject: Additional information for amendment request

(NOTE: This will be used as the
Documents Title in ADAMS)

This record summarizes a series of communications on the date referenced above.

- 1) Item 1c of the letter dated February 2, 2012 included the following regarding Dr. Housman: "However he participated in two HDR treatments using the Nucletron Microselectron at Danbury Hospital (see attached letter of attestation) on January 25th of this year." It appears that the referenced letter was not attached. Ian e-mailed a scanned copy of the letter.
- 2) Nucletron Model 106.990 is approved in the Sealed Source and Device Registry for both high dose rate and pulsed dose use. I asked Ian to confirm that THLRCC intends to use it only in high dose rate mode. He confirmed this.

Action Required: Issue amendment

Document Availability:

 Publicly Available Non-Publicly Available Non-Sensitive
 Sensitive – Privacy Act/ PII
 Immediate Release Non-Sensitive Copyright
 Sensitive – Internal
Normal Release Date: 02/14/2012 Sensitive- Proprietary
 Sensitive – Security-Related
Delayed Release Date:SUNSI Review Completed By: SGabriel / RA /Document Accession #: ML120380184