

817 339-6581

PAU FORNES



TroXler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
 Research Triangle Park, NC 27709
 Tel: (877) 876-9537 Fax: (919) 485-2250

License: NC 032-0182-1

LEAK TEST CERTIFICATE

DEVICE:

Model: 3440 Serial No: 61758

SEALED SOURCES:

SERIAL NO.	MEASURE DATE	NUCLIDE	ACTIVITY	
			(GBq)	(mCi)
78-5234	5/28/2008	Am-241:Be	1.48	40
77-8544	10/1/2008	Cs-137	0.296	8

LEAK TEST ANALYSIS:

Sample collected on: 06/07/2011

Sample analyzed on: 06/17/2011 at 9:52:00 AM

Analyzed by: D. Rose

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.22E+01	1.95E+01
Background measurement (cpm)	2	27
Sample measurement (cpm)	0	30
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	7.2E-01	1.4E+00

This certifies that the above leak test results are:

 Less than 185 Bq (0.005 uCi) Greater than 185 Bq (0.005 uCi)

If greater than 185 Bq (0.005 uCi):

Person Notified _____ Date _____

Phone _____ and/or Fax _____

RECEIVED FEB 29 2012