



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION IV
1600 EAST LAMAR BOULEVARD
ARLINGTON, TEXAS 76011-4511

March 9, 2012

Team Industrial Services, Inc.
ATTN: David P. Tebo
Corporate Radiation Safety Officer
200 Hermann Drive
Alvin, Texas 77511

SUBJECT: VOIDANCE OF AMENDMENT REQUEST

The Nuclear Regulatory Commission (NRC) has received your letter dated December 28, 2011, requesting authorization to use the instrument Sentinel Radeye G in different modalities with the intent to comply with Title 10 of the Code of Federal Regulations, Part 34, "Licenses for Industrial Radiography and Radiation Safety Requirements for Industrial Radiographic Operations."

The NRC has reviewed your request and has determined that the use of one device alone, such as the Radeye G, even if it serves the function of two or three devices during normal radiographic operations and during emergency situations as described in your letter dated December 28, 2011, will not meet the intent of 10 CFR 34.47, "Personnel monitoring," and 10 CFR 34.25, "Radiation survey instruments." Therefore, your amendment request dated December 28, 2011, is being voided without prejudice.

The NRC has a process for developing and modifying NRC regulations and is called rulemaking. Any member of the public may petition the NRC to develop, change, or rescind one of its regulations. If you believe your request has merits, you can go to the following website to obtain information on how to submit a petition for rulemaking:

<http://www.nrc.gov/about-nrc/regulatory/rulemaking/petition-rule.html>.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter will be available electronically for public inspection in the NRC Public Document Room or from the NRC's document system (ADAMS). ADAMS is accessible from the NRC Web site at

<http://www.nrc.gov/reading-rm/adams.html>.

Thank you for your cooperation.

Sincerely,

/RA/

Roberto J. Torres, Senior Health Physicist
Nuclear Materials Safety Branch B

Docket: 030-35252
License: 42-32219-01
Control: 576614