



TO: file

COMPANY: N/A

NUCLEAR REGULATORY COMMISSION
REGION III
2443 WARRENVILLE ROAD
LISLE, ILLINOIS 60532-4351
(630) 829-9892 FAX: (630) 515-1078

PAGES: 1 TEL: N/A
FAX #: N/A
EMAIL: N/A

CONVERSATION RECORD

| | | TIME | DATE |
|--|----------------|--|-------------------|
| | | 11:15 am | February 24, 2012 |
| NAME OF PERSON(S) CONTACTED | TELEPHONE NO. | ORGANIZATION | |
| Kerry Krugh, Ph.D. | (419) 291-4310 | Toledo Hosp Cardiac Testing Ctr Monroe | |
| REPRESENTED PERSON or PERSONS | | ORGANIZATION | |
| Kerry Krugh, Ph.D., Radiation Safety Officer | | Toledo Hosp Cardiac Testing Ctr Monroe | |
| SUBJECT | | | |
| License No.: 21-26562-01 | | Control No.: 576871 | |

SUMMARY

We have reviewed your license amendment request and find that we are unable to continue this action until we have received information regarding the following:

1. The submitted facility diagram for the proposed new location of use lacks detail as to the areas adjacent to the radioactive materials use area. Please resubmit including a description of any activities being conducted in the areas contiguous to the proposed radioactive materials use area(s). In addition, the diagram should show details for the nuclear medicine suite, including locations of any sinks, waste areas, and exhaust hoods. The hot lab should be clearly indicated on the diagram, and be free of superfluous architectural markings, unrelated to radioactive materials use concerns.
RESPONSE: The contact person resubmitted the facility diagram via facsimile; it was received on February 29, 2012. No additional response is required.
2. The request was silent as to whether PET is being used at the proposed new location of use. If PET is being used, additional shielding calculations may be required.
RESPONSE: The licensee indicated, via phone conversation of February 24, 2012 that no PET isotopes would be used. No additional information is required.

We have requested that you submit the referenced item:

- Revised facility diagram

- via facsimile, to (630) 515-1078. Please reference the Control No. 576871, as listed at the top of this memo. We expected to – and did – hear from you on or before March 2, 2012. No additional information is required.

Please submit the requested information within 7 days of this record. **Include reference control number 576871, Please FAX your response to my attention at (630) 515-1078.** You may also scan your response and send to me via email, as a pdf file.

Please direct any questions you have to me at (630) 829-9892 or sara.forster@nrc.gov.

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

Sara A.B. Forster

Sara A.B. Forster 3/9/2012