

Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the Information Services Branch (T-5 F53), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to Infocollects.Resource@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

CERTIFICATE OF DISPOSITION OF MATERIALS

LICENSEE NAME AND ADDRESS

URS Corporation
1501 4th Avenue, Suite 1400
Seattle, Washington 98101

LICENSE NUMBER

46-29312-01

DOCKET NUMBER

030-37856

LICENSE EXPIRATION DATE

02/28/2019

A. LICENSE STATUS (Check the appropriate box)☐ This license has expired.☒ This license has not yet expired; please terminate it.**B. DISPOSAL OF RADIOACTIVE MATERIAL**

(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

1. No radioactive materials have ever been procured or possessed by the licensee under this license.
2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner:
 - a. Transfer of radioactive materials to the licensee listed below:

- b. Disposal of radioactive materials:

1. Directly by the licensee:

2. By licensed disposal site:

3. By waste contractor:

- c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

RECEIVED

MAR 12 2012

DNMS

C. SURVEYS PERFORMED AND REPORTED

1. A radiation survey was conducted by the licensee. The survey confirms:
 - a. the absence of licensed radioactive materials
 - b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.
2. A copy of the radiation survey results:
 - a. is attached; or
 - b. is not attached (Provide explanation); or
 - c. was forwarded to NRC on: _____ Date _____
- ☒ 3. A radiation survey is not required as only sealed sources were ever possessed under this license, and
 - ☒ a. The results of the latest leak test are attached; and/or
 - b. No leaking sources have ever been identified.

The person to be contacted regarding the information provided on this form:

NAME

David I Johnson

TITLE

Graduate Geotechnical Engineer

TELEPHONE (Include Area Code)

(206) 438-2700

E-MAIL ADDRESS

david.i.johnson@urs.com

Mail all future correspondence regarding this license to:

1501 4th Avenue, Suite 1400, Seattle, WA 98101

C. CERTIFYING OFFICIAL**I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT**

PRINTED NAME AND TITLE

DAVID I. JOHNSON

SIGNATURE



DATE

2-9-2012

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

CERTIFICATE OF DISPOSITION OF MATERIALS

PLEASE READ THESE INSTRUCTIONS BEFORE COMPLETING NRC FORM 314.

Subpart E of 10 CFR Part 20 establishes the radiological criteria for license terminations/decommissioning of facilities licensed under 10 CFR Parts 30, 40, 50, 60, 61, 70, and 72, as well as other facilities subject to the Commission's jurisdiction under the Atomic Energy Act of 1954, as amended, and the Energy Reorganization Act of 1974, as amended.

INSTRUCTIONS

Section B, Item 2.

Licensees should describe the specific radioactive material transfer actions. If radioactive wastes were generated in terminating this license, the licensee should describe the disposal actions taken, including the disposition of low-level radioactive waste, mixed waste, greater-than-Class-C waste, and sealed sources.

Section B, Item 2.a.

The information provided concerning the transfer of radioactive material to another licensee should specify the date of the transfer, the name of the licensee recipient, an individual contact name and telephone number for the licensee recipient, and the recipient's NRC or Agreement State license number.

Section B, Item 2.b.

For disposal of radioactive materials, licensees should describe the specific disposal method or procedure (e.g., decay-in-storage). For those cases when radioactive materials are disposed of by a licensed disposal site or by a waste contractor, the licensee should specify the name, address, and telephone number of the licensed disposal site operator or waste contractor.

Section B, Item 2.c.

"Residual radioactivity," as defined in 10 CFR 20.1003, means radioactivity in 'areas' (structures, materials, soils, etc.) remaining as a result of activities (licensed and unlicensed) under the licensee's control from sources used by the licensee, excluding background radiation. ALARA is defined in 10 CFR 20.1003.

FILE CERTIFICATES AS FOLLOWS:

IF YOU ARE LOCATED IN:

ALABAMA, CONNECTICUT, DELAWARE, DISTRICT OF COLUMBIA, FLORIDA, GEORGIA, KENTUCKY, MAINE, MARYLAND, MASSACHUSETTS, NEW HAMPSHIRE, NEW JERSEY, NEW YORK, NORTH CAROLINA, PENNSYLVANIA, PUERTO RICO, RHODE ISLAND, SOUTH CAROLINA, TENNESSEE, VERMONT, VIRGINIA, VIRGIN ISLANDS, OR WEST VIRGINIA, SEND CERTIFICATES TO:

LICENSING ASSISTANT SECTION
NUCLEAR MATERIALS SAFETY BRANCH
U.S. NUCLEAR REGULATORY COMMISSION, REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PA 19406-1415


IF YOU ARE LOCATED IN:

ALASKA, ARIZONA, ARKANSAS, CALIFORNIA, COLORADO, HAWAII, IDAHO, KANSAS, LOUISIANA, MISSISSIPPI, MONTANA, NEBRASKA, NEVADA, NEW MEXICO, NORTH DAKOTA, OKLAHOMA, OREGON, PACIFIC TRUST TERRITORIES, SOUTH DAKOTA, TEXAS, UTAH, WASHINGTON, OR WYOMING, SEND CERTIFICATES TO:

MATERIAL RADIATION PROTECTION SECTION
U. S. NUCLEAR REGULATORY COMMISSION, REGION IV
612 E. LAMAR BOULEVARD, SUITE 400
ARLINGTON, TX 76011-4125

ILLINOIS, INDIANA, IOWA, MICHIGAN, MINNESOTA, MISSOURI, OHIO, OR WISCONSIN, SEND CERTIFICATES TO:

MATERIALS LICENSING SECTION
U.S. NUCLEAR REGULATORY COMMISSION, REGION III
2443 WARRENVILLE ROAD, SUITE 210
LISLE, IL 60532-4352



"Your One Stop for Calibration & Service"

2513 Weaver Street Ste A
Fort Worth, TX 76117
(817) 834-5411
(817) 834-5435 Fax

702 West 48th Avenue Unit G
Denver Colorado 80216
(303) 292-2412
(303) 292-2409 Fax

Attention: RSO

URS Corporation
1501 4th Avenue, Suite 1400
Seattle, WA 98101-1616

Certificate Number: OSCS32811

Leak Test Certificate

Gauge Serial Number: MD00805858

Gauge Model: MC1 DRP

Sample Procedure by: S Jacks

Sample Date: 3/28/2011

Sealed Source Isotope:

Activity:

Removable Contamination:

Cs-137 (gamma)

10 mCi

< .005 μ Ci

Am-241 (alpha)

50 mCi

< .005 μ Ci

The analysis procedure was performed on 4/4/2011 by OSCS.

This portable nuclear gauge displays no removable contamination and may remain in active service.

The .001 μ Ci Cs-137 and Am-241 calibration check sources are NIST traceable.

The survey meter used to perform the analysis was calibrated on 28 April 2010.

Federal and Agreement State limits for removable contamination is 0.005 μ Ci (5.0×10^{-3} μ Ci).

OSCS is authorized to analyze portable and fixed gauge leak test samples; State of Texas

License Number: LO-5813.

Notice: The RSO or owner of the gauge that fails the leak test analysis will be alerted immediately by telephone, e-mail and or fax to remove the gauge from active service.

Certified by:

James Stiles



"Your One Stop for Calibration & Service"

2513 Weaver Street Ste A
Fort Worth, TX 76117
(817) 834-5411
(817) 834-5435 Fax

702 West 48th Avenue Unit G
Denver Colorado 80216
(303) 292-2412
(303) 292-2409 Fax

Attention: RSO

URS Corporation
1501 4th Avenue, Suite 1400
Seattle, WA 98101-1616

Certificate Number: OSCS41311

Leak Test Certificate

Gauge Serial Number: MD00805861

Gauge Model: MC1 DRP

Sample Procedure by: S Jacks

Sample Date: 4/13/2011

Sealed Source Isotope:

Activity:

Removable Contamination:

Cs-137 (gamma)

10 mCi

< .005 μ Ci

Am-241 (alpha)

50 mCi

< .005 μ Ci

The analysis procedure was performed on 4/20/2011 by OSCS.

This portable nuclear gauge displays no removable contamination and may remain in active service.

The .001 μ Ci Cs-137 and Am-241 calibration check sources are NIST traceable.

The survey meter used to perform the analysis was calibrated on 28 April 2010.

Federal and Agreement State limits for removable contamination is 0.005 μ Ci (5.0×10^{-3} μ Ci).

OSCS is authorized to analyze portable and fixed gauge leak test samples; State of Texas

License Number: LO-5813.

Notice: The RSO or owner of the gauge that fails the leak test analysis will be alerted immediately by telephone, e-mail and or fax to remove the gauge from active service.

Certified by:

James Stiles

577119



"Your One Stop for Calibration & Service"

2513 Weaver Street Ste A
Fort Worth, TX 76117
(817) 834-5411
(817) 834-5435 Fax

702 West 48th Avenue Unit G
Denver Colorado 80216
(303) 292-2412
(303) 292-2409 Fax

Attention: RSO

URS Corporation
1501 4th Avenue, Suite 1400
Seattle, WA 98101-1616

Certificate Number: OSCS32811

Leak Test Certificate

Gauge Serial Number: MD01005902

Gauge Model: MC1 DRP

Sample Procedure by: S Jacks

Sample Date: 3/28/2011

Sealed Source Isotope:

Activity:

Removable Contamination:

Cs-137 (gamma)

10 mCi

< .005 μ Ci

Am-241 (alpha)

50 mCi

< .005 μ Ci

The analysis procedure was performed on 4/4/2011 by OSCS.

This portable nuclear gauge displays no removable contamination and may remain in active service.

The .001 μ Ci Cs-137 and Am-241 calibration check sources are NIST traceable.

The survey meter used to perform the analysis was calibrated on 28 April 2010.

Federal and Agreement State limits for removable contamination is 0.005 μ Ci (5.0×10^{-3} μ Ci).

OSCS is authorized to analyze portable and fixed gauge leak test samples; State of Texas

License Number: LO-5813.

Notice: The RSO or owner of the gauge that fails the leak test analysis will be alerted immediately by telephone, e-mail and or fax to remove the gauge from active service.

Certified by:

A handwritten signature in black ink, appearing to read 'James Stiles'.

James Stiles



"Your One Stop for Calibration & Service"

2513 Weaver Street Ste A
Fort Worth, TX 76117
(817) 834-5411
(817) 834-5435 Fax

702 West 48th Avenue Unit G
Denver Colorado 80216
(303) 292-2412
(303) 292-2409 Fax

Attention: RSO

URS Corporation
1501 4th Avenue, Suite 1400
Seattle, WA 98101-1616

Certificate Number: OSCS41311

Leak Test Certificate

Gauge Serial Number: MD01005903

Gauge Model: MC1 DRP

Sample Procedure by: S Jacks

Sample Date: 4/13/2011

Sealed Source Isotope:

Activity:

Removable Contamination:

Cs-137 (gamma)

10 mCi

< .005 μ Ci

Am-241 (alpha)

50 mCi

< .005 μ Ci

The analysis procedure was performed on 4/20/2011 by OSCS.

This portable nuclear gauge displays no removable contamination and may remain in active service.

The .001 μ Ci Cs-137 and Am-241 calibration check sources are NIST traceable.

The survey meter used to perform the analysis was calibrated on 28 April 2010.

Federal and Agreement State limits for removable contamination is 0.005 μ Ci (5.0×10^{-3} μ Ci).

OSCS is authorized to analyze portable and fixed gauge leak test samples; State of Texas

License Number: LO-5813.

Notice: The RSO or owner of the gauge that fails the leak test analysis will be alerted immediately by telephone, e-mail and or fax to remove the gauge from active service.

Certified by:

James Stiles


URS

DAVID L. JOHNSON
URS Corporation
Century Square
1501 4th Avenue, Suite 1400
Seattle, WA 98101-3685

RECEIVED
MAR 12 2012
DNMS

SEATTLE WASH 9801

07 MAR 2012 PM 5:1



MATERIAL RADIATION REGIONAL SECTION
U.S. Nuclear Regulatory Commission
612 F. LAMAR BOULEVARD, SUITE 400
ARLINGTON, TX 76011-4125

7601184125



16 577119



DATE
03/12/2012

NAME AND ADDRESS OF APPLICANT AND/OR LICENSEE

URS Corporation
ATTN: Katy Chancy
Vice President
1501 4th Avenue, Suite 1400
Seattle, Washington 98101

LICENSE NUMBER

46-29312-01

MAIL CONTROL NUMBER

577119

LICENSING AND/OR TECHNICAL REVIEWER

ch

This is to acknowledge the receipt of your:

☐ LETTER and/or ☒ APPLICATION DATED: 02/09/2012

The initial processing, which included an administrative review, has been performed.

☐ AMENDMENT ☒ TERMINATION ☐ NEW LICENSE ☐ RENEWAL

- ☒ There were no administrative omissions identified during our initial review.
- ☐ This is to acknowledge receipt of your application for renewal of the material(s) license identified above. Your application is deemed timely filed, and accordingly, the license will not expire until final action has been taken by this office.
- ☐ Your application for a new NRC license did not include your taxpayer identification number. Please fill out NRC Form 531, located at the following link:

<http://www.nrc.gov/reading-rm/doc-collections/forms/nrc531.pdf>

Send the completed NRC Form 531, by facsimile, to the following number: (301) 415-5387

A copy of your action has been emailed to our License Fee and Accounts Receivable Branch, in our Headquarters office in Rockville, MD. You will be contacted separately if there is a fee issue involved.

Your application has been assigned the above listed **MAIL CONTROL NUMBER**. When calling to inquire about this action, please refer to this control number. Your application has been forwarded to a technical reviewer. Please note that the technical review, which is normally completed within 180 days for a renewal application (90 days for all other requests), may identify additional omissions or require additional information. If you have any questions concerning the processing of your application, our contact information is listed below:

Region IV
U. S. Nuclear Regulatory Commission
DNMS/NMSB - B
1600 E. Lamar Boulevard
Arlington, TX 76011-4511
(817) 200-1103 or (817) 200-1140

BETWEEN:

Accounts Receivable/Payable
and
Regional Licensing Branches

[FOR ARPB USE]
INFORMATION FROM LTS

Program Code: 03121
Status Code: Pending Termination
Fee Category: 3P
Exp. Date:
Fee Comments:
Decom Fin Assur Req'd: N

License Fee Worksheet - License Fee Transmittal

A. REGION

1. APPLICATION ATTACHED

Applicant/Licensee: URS CORPORATION
Received Date: 03/12/2012
Docket Number: 3037856
Mail Control Number: 577119
License Number: 46-29312-01
Action Type: Termination

2. FEE ATTACHED

Amount: _____

Check No.: _____

3. COMMENTS

Signed: _____

Date: _____

B. LICENSE FEE MANAGEMENT BRANCH (Check when milestone 03 is entered / /)

1. Fee Category and Amount: _____

2. Correct Fee Paid. Application may be processed for:

Amendment: _____

Renewal: _____

License: _____

3. OTHER _____

Signed: _____

Date: _____