

U.S. NUCLEAR REGULATORY COMMISSION**Date:** 4/11/2012**TELEPHONE CONVERSATION RECORD**

Mail Control or Report No(s).	577196 577197	License No(s).	06-30764-01 06-30764-02MD	Docket No(s).	03036099 03036179
--	------------------	-----------------------	------------------------------	----------------------	----------------------

Name of Licensee: Biocompatibles, Inc.**Name of Participant(s):** Dennis Lawyer, R1, DNMS
Wayne Richardson, RSO**Telephone No.** 203-262-4194**Subject:** RAI, Sealed Source Material Amount(NOTE: This will be used as the
Documents Title in ADAMS)

Mr. Richardson, you requested the additional sealed source to contain 100 millicuries. The sealed source and device registry only allows for 20 millicuries maximum. Will that be sufficient to be licensed for 20 millicuries.?

Mr. Richardson responded that 20 millicuries will be sufficient.

Action Required: None**Document Availability:**☒**Publicly Available**☐**Non-Publicly Available**☒**Non-Sensitive**☐**Non-Sensitive Copyright**☐**Sensitive- Proprietary**☐**Sensitive – Privacy Act/ PII**☐**Sensitive – Internal**☐**Sensitive – Security-Related**☐**Immediate Release****Normal Release Date:** 4/19/12**Delayed Release Date:****SUNSI Review Completed By:** Dennis Lawyer / RA /**Document Accession #:** ML121020474