

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Bob's HI-TECH Corporation
177 Canfield Road
Chardon, OH 44024

REPORT NUMBER(S) 12-01

2. NRC/REGIONAL OFFICE

Region III
U. S. Nuclear Regulatory Commission
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

3. DOCKET NUMBER(S)

030-35051

4. LICENSE NUMBER(S)

34-26544-02

5. DATE(S) OF INSPECTION

April 5, 2012

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- ☒ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

- ☐ 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
- (Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

| TITLE | PRINTED NAME | SIGNATURE | DATE |
|---------------------------|-------------------|--|---------|
| LICENSEE'S REPRESENTATIVE | | | |
| NRC INSPECTOR | Robert P. Hays |  | 4/5/12 |
| BRANCH CHIEF | Tamara E. Bloomer |  | 4/12/12 |

Docket File Information**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED:

Bob's HI-TECH Corporation
177 Canfield Road
Chardon, OH 44024

REPORT NUMBER(S) 12-01

2. NRC/REGIONAL OFFICE

Region III
U. S. Nuclear Regulatory Commission
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

3. DOCKET NUMBER(S)

030-35051

4. LICENSE NUMBER(S)

34-26544-02

5. DATE(S) OF INSPECTION

April 5, 2012

6. INSPECTION PROCEDURES USED

87126

7. INSPECTION FOCUS AREAS

03.01-03.07

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S)

03225

2. PRIORITY

5

3. LICENSEE CONTACT

Robert Janda, RSO

4. TELEPHONE NUMBER

(440) 285-2385

☒ Main Office Inspection

Next Inspection Date: 04/05/2017

☐ Field Office Inspection☐ Temporary Job Site Inspection**PROGRAM SCOPE**

The licensee is a small gauge service business authorized by their NRC license to perform servicing of fixed gauges as specified by the license at customer locations as temporary job sites of the licensee anywhere in the United States where the U.S. Nuclear Regulatory Commission maintains jurisdiction for regulating the use of licensed material. The licensee consists of one individual who performs leak testing, maintenance and repair, shutter checks, packaging of gauges for return to the manufacturer, and training of gauge operator staff. Currently, the licensee provides services to one client in Michigan. The licensee maintains detailed records for each client serviced by the licensee.

Performance Observations

During the inspection, the licensee's RSO/owner discussed radiation safety procedures which included: (1) radiation surveys and survey meter calibrations; (2) leak tests; (3) package receipt and handling procedures; (4) gauge transfers and disposals; (5) dosimetry; (6) safe use practices; (7) survey meter availability and calibrations; and (10) DOT Training and refresher training.