

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

<p>1. LICENSEE/LOCATION INSPECTED:</p> <p>Solar Testing Laboratories, Inc. 1125 Valley Belt Road Brooklyn, Heights, OH 44131</p> <p>REPORT NUMBER(S) 12-01</p>	<p>2. NRC/REGIONAL OFFICE</p> <p>Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352</p>
--	---

<p>3. DOCKET NUMBER(S)</p> <p>030-35130</p>	<p>4. LICENSE NUMBER(S)</p> <p>34-16119-04</p>	<p>5. DATE(S) OF INSPECTION</p> <p>April 4, 2012</p>
---	--	--

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Robert P. Hays		4/4/12
BRANCH CHIEF	Tamara E. Bloomer		4/12/12

Docket File Information
SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Solar Testing Laboratories, Inc. 1125 Valley Belt Road Brooklyn, Heights, OH 44131 REPORT NUMBER(S) 12-01	2. NRC/REGIONAL OFFICE Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352
---	---

3. DOCKET NUMBER(S) 030-35130	4. LICENSE NUMBER(S) 34-16119-04	5. DATE(S) OF INSPECTION April 4, 2012
--------------------------------------	---	---

6. INSPECTION PROCEDURES USED 87124	7. INSPECTION FOCUS AREAS 03.01-03.07
--	--

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 03121	2. PRIORITY 5	3. LICENSEE CONTACT Dave Snetkamp, RSO	4. TELEPHONE NUMBER (216) 741-7007
---------------------------------	----------------------	---	---

Main Office Inspection Next Inspection Date: 04/04/2017

Field Office Inspection

Temporary Job Site Inspection

PROGRAM SCOPE

The licensee is authorized to possess Troxler Model 3400 series, Humbolt Scientific Model 5001, or Troxler asphalt content gauges at temporary job sites anywhere in the United States where the NRC maintains regulatory jurisdiction. At the time of the inspection, the licensee did not have any temporary job site work scheduled in NRC jurisdiction. The gauges are currently used in the Cleveland, OH area under the licensee's Ohio license. The licensee maintains their NRC license for authorization to use a gauge at National Aeronautics and Space Administration sites located in the State of Ohio.

Performance Observations

During this inspection, the inspector reviewed and the RSO demonstrated/discussed: (1) Two tangible barriers for gauge security and padlocks while unattended at a temporary job site; (2) DOT requirements and shipping papers; (3) dosimetry (RSO only individual monitored); (4) sealed source inventories and leak tests (February 2012); (5) survey meter availability and use; (6) emergency procedures and reporting requirements; (7) annual program reviews.

The inspector performed independent radiation measurements which indicated no dose concerns and consistent with expected readings.